

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100622\
 Data File : BN022017.D
 Acq On : 06 Oct 2022 12:12
 Operator : CG/JU
 Sample : N4785-05
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PT-BN-WP

Quant Time: Oct 07 00:39:45 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 01:32:21 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.978	152	4598	0.400	ng	0.00
7) Naphthalene-d8	10.805	136	13325	0.400	ng	#-0.01
13) Acenaphthene-d10	14.653	164	6792	0.400	ng	0.00
19) Phenanthrene-d10	17.400	188	14647	0.400	ng	#-0.01
29) Chrysene-d12	21.618	240	11307	0.400	ng	0.00
35) Perylene-d12	24.093	264	10043	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.508	112	1839836	172.734	ng	0.00
5) Phenol-d6	7.141	99	2486356	179.903	ng	0.00
8) Nitrobenzene-d5	9.161	82	1276053	119.109	ng	0.00
11) 2-Methylnaphthalene-d10	12.368	152	66	0.003	ng	-0.04
14) 2,4,6-Tribromophenol	16.146	330	636618	245.758	ng	0.00
15) 2-Fluorobiphenyl	13.274	172	2531088	85.291	ng	-0.01
27) Fluoranthene-d10	19.441	212	18	0.000	ng	0.00
31) Terphenyl-d14	20.042	244	2004387	88.175	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.320	88	126	0.021	ng	# 22
3) n-Nitrosodimethylamine	3.638	42	989	0.153	ng	# 1
6) bis(2-Chloroethyl)ether	7.401	93	2050832	140.421	ng	97
9) Naphthalene	10.859	128	4801618	119.329	ng	100
10) Hexachlorobutadiene	11.147	225	1546	0.222	ng	# 99
16) Acenaphthylene	14.375	152	3949249	122.205	ng	98
17) Acenaphthene	14.707	154	1055172	46.442	ng	95
18) Fluorene	15.701	166	1001272	36.550	ng	98
20) 4,6-Dinitro-2-methylph...	15.786	198	4	0.193	ng	# 1
21) 4-Bromophenyl-phenylether	16.593	248	593097	66.470	ng	# 87
22) Hexachlorobenzene	16.705	284	553691	49.896	ng	98
24) Pentachlorophenol	17.052	266	113	0.033	ng	# 69
25) Phenanthrene	17.449	178	2795095	55.515	ng	99
26) Anthracene	17.549	178	4979268	123.178	ng	97
28) Fluoranthene	19.470	202	2184538	40.495	ng	98
30) Pyrene	19.833	202	94283	1.899	ng	# 95
32) Benzo(a)anthracene	21.600	228	3765836	88.571	ng	98
33) Chrysene	21.654	228	4387068	92.977	ng	97
34) Bis(2-ethylhexyl)phtha...	21.510	149	4080457	200.016	ng	95
36) Indeno(1,2,3-cd)pyrene	26.730	276	4677222	92.171	ng	96
37) Benzo(b)fluoranthene	23.397	252	5304913	111.437	ng	# 91
38) Benzo(k)fluoranthene	23.397	252	5262592	111.042	ng	# 91
39) Benzo(a)pyrene	23.994	252	3962895	109.137	ng	# 85
40) Dibenzo(a,h)anthracene	26.745	278	1423017	35.163	ng	95
41) Benzo(g,h,i)perylene	27.508	276	1250648	28.626	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100622\
Data File : BN022017.D
Acq On : 06 Oct 2022 12:12
Operator : CG/JU
Sample : N4785-05
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Quant Time: Oct 07 00:39:45 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 04 01:32:21 2022
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.978 min Scan# 693
Delta R.T. -0.007 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Instrument :
BNA_N
ClientSampleId :
PT-BN-WP







Tgt Ion:152 Resp: 4598
Ion Ratio Lower Upper
152 100
150 163.1 127.2 190.8
115 65.1 51.4 77.0



#2
1,4-Dioxane
Concen: 0.021 ng
RT: 3.320 min Scan# 47
Delta R.T. -0.007 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion: 88 Resp: 126
Ion Ratio Lower Upper
88 100
43 0.0 22.6 34.0#
58 0.0 58.6 88.0#









#3
 n-Nitrosodimethylamine
 Concen: 0.153 ng
 RT: 3.638 min Scan# 91
 Delta R.T. -0.029 min
 Lab File: BN022017.D
 Acq: 06 Oct 2022 12:12




Instrument :
 BNA_N
 ClientSampleId :
 PT-BN-WP

Tgt Ion: 42 Resp: 989
 Ion Ratio Lower Upper
 42 100
 74 0.0 101.6 152.4#
 44 1.8 10.4 15.6#



#4
 2-Fluorophenol
 Concen: 172.734 ng
 RT: 5.508 min Scan# 350
 Delta R.T. -0.007 min
 Lab File: BN022017.D
 Acq: 06 Oct 2022 12:12

Tgt Ion: 112 Resp: 1839836
 Ion Ratio Lower Upper
 112 100
 64 59.1 46.9 70.3
 63 30.7 24.6 37.0









#5
Phenol-d6
Concen: 179.903 ng
RT: 7.141 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Instrument :
BNA_N
ClientSampleId :
PT-BN-WP

Tgt Ion: 99 Resp: 2486356




Ion	Ratio	Lower	Upper
99	100		
42	19.1	15.9	23.9
71	33.0	25.6	38.4




#6
bis(2-Chloroethyl)ether
Concen: 140.421 ng
RT: 7.401 min Scan# 612
Delta R.T. -0.007 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion: 93 Resp: 2050832

Ion	Ratio	Lower	Upper
93	100		
63	77.9	64.8	97.2
95	33.1	26.6	40.0









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.805 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Tgt Ion:136 Resp: 13325
Ion Ratio Lower Upper
136 100
137 11.7 9.2 13.8
54 11.0 7.1 10.7#
68 9.0 6.2 9.2



#8
Nitrobenzene-d5
Concen: 119.109 ng
RT: 9.161 min Scan# 818
Delta R.T. 0.000 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion: 82 Resp: 1276053
Ion Ratio Lower Upper
82 100
128 46.8 29.9 44.9#
54 46.0 43.5 65.3









#9
Naphthalene
Concen: 119.329 ng
RT: 10.859 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP

Tgt Ion:128 Resp: 4801618
Ion Ratio Lower Upper
128 100
129 11.9 9.3 13.9
127 13.4 10.6 16.0



#10
Hexachlorobutadiene
Concen: 0.222 ng
RT: 11.147 min Scan# 1004
Delta R.T. 0.000 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion:225 Resp: 1546
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 51.3 76.9









#11
2-Methylnaphthalene-d10
Concen: 0.003 ng
RT: 12.368 min Scan# 11
Delta R.T. -0.042 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Tgt Ion:152 Resp: 66
Ion Ratio Lower Upper
152 100
151 48.5 13.8 20.6#



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.653 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12


Tgt Ion:164 Resp: 6792
Ion Ratio Lower Upper
164 100
162 102.8 84.3 126.5
160 48.7 41.7 62.5








#14
2,4,6-Tribromophenol
Concen: 245.758 ng
RT: 16.146 min Scan# 1
Delta R.T. 0.000 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12


Instrument :
BNA_N
ClientSampleId :
PT-BN-WP





Tgt Ion:330 Resp: 636618
Ion Ratio Lower Upper
330 100
332 97.4 77.0 115.4
141 47.0 35.3 52.9




#15
2-Fluorobiphenyl
Concen: 85.291 ng
RT: 13.274 min Scan# 1373
Delta R.T. -0.011 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12



Tgt Ion:172 Resp: 2531088
Ion Ratio Lower Upper
172 100
171 34.6 27.2 40.8
170 23.7 18.3 27.5









#16
Acenaphthylene
Concen: 122.205 ng
RT: 14.375 min Scan# 1476
Delta R.T. 0.000 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Tgt Ion:152 Resp: 3949249
Ion Ratio Lower Upper
152 100
151 19.9 15.5 23.3
153 13.9 10.4 15.6



#17
Acenaphthene
Concen: 46.442 ng
RT: 14.707 min Scan# 1507
Delta R.T. -0.011 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion:154 Resp: 1055172
Ion Ratio Lower Upper
154 100
153 111.3 92.1 138.1
152 53.9 47.8 71.6





#18

Fluorene

Concen: 36.550 ng

RT: 15.701 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12

Instrument :

BNA_N

ClientSampleId :

PT-BN-WP





Tgt Ion:166 Resp: 1001272

Ion Ratio Lower Upper

166 100

165 97.9 80.0 120.0

167 13.7 10.9 16.3



#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.400 min Scan# 1737

Delta R.T. -0.012 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12




Tgt Ion:188 Resp: 14647


Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 16.3 7.5 11.3#









#20
4,6-Dinitro-2-methylphenol
Concen: 0.193 ng
RT: 15.786 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Tgt Ion:198 Resp: 4
Ion Ratio Lower Upper
198 100
51 805.9 96.6 144.8#
105 427.5 64.4 96.6#



#21
4-Bromophenyl-phenylether
Concen: 66.470 ng
RT: 16.593 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion:248 Resp: 593097
Ion Ratio Lower Upper
248 100
250 98.2 76.6 114.8
141 50.5 57.5 86.3#





#22

Hexachlorobenzene

Concen: 49.896 ng

RT: 16.705 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12

Instrument :

BNA_N

ClientSampleId :

PT-BN-WP





Tgt Ion:284 Resp: 553691

Ion Ratio Lower Upper

284 100

142 43.4 33.4 50.0

249 31.3 25.1 37.7



#24

Pentachlorophenol

Concen: 0.033 ng

RT: 17.052 min Scan# 1709

Delta R.T. -0.012 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12




Tgt Ion:266 Resp: 113


Ion Ratio Lower Upper

266 100

264 101.8 49.8 74.6#

268 70.8 49.6 74.4





#25

Phenanthrene

Concen: 55.515 ng

RT: 17.449 min Scan# 1741

Delta R.T. 0.000 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12

Instrument :

BNA_N

ClientSampleId :

PT-BN-WP





Tgt Ion:178 Resp: 2795095

Ion Ratio Lower Upper

178 100

176 19.7 15.4 23.0

179 15.9 12.2 18.2



#26

Anthracene

Concen: 123.178 ng

RT: 17.549 min Scan# 1749

Delta R.T. 0.013 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12




Tgt Ion:178 Resp: 4979268


Ion Ratio Lower Upper

178 100

176 19.6 14.8 22.2

179 16.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.000 ng
RT: 19.441 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Tgt Ion:212 Resp: 18
Ion Ratio Lower Upper
212 100
106 383.3 15.2 22.8#
104 0.0 8.6 13.0#



#28
Fluoranthene
Concen: 40.495 ng
RT: 19.470 min Scan# 2131
Delta R.T. -0.004 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion:202 Resp: 2184538
Ion Ratio Lower Upper
202 100
101 15.1 11.4 17.0
203 17.9 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.618 min Scan# 2438
Delta R.T. 0.009 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Tgt Ion:240 Resp: 11307
Ion Ratio Lower Upper
240 100
120 13.6 11.8 17.6
236 28.0 22.4 33.6



#30
Pyrene
Concen: 1.899 ng
RT: 19.833 min Scan# 2217
Delta R.T. -0.004 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion:202 Resp: 94283
Ion Ratio Lower Upper
202 100
200 21.2 16.6 25.0
203 22.4 14.1 21.1#









#31
 Terphenyl-d14
 Concen: 88.175 ng
 RT: 20.042 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN022017.D
 Acq: 06 Oct 2022 12:12




Instrument :
 BNA_N
 ClientSampleId :
 PT-BN-WP


Tgt Ion:244 Resp: 2004387
 Ion Ratio Lower Upper
 244 100
 212 9.6 6.5 9.7
 122 15.9 9.1 13.7#



#32
 Benzo(a)anthracene
 Concen: 88.571 ng
 RT: 21.600 min Scan# 2436
 Delta R.T. 0.009 min
 Lab File: BN022017.D
 Acq: 06 Oct 2022 12:12

Tgt Ion:228 Resp: 3765836
 Ion Ratio Lower Upper
 228 100
 226 29.0 22.2 33.4
 229 20.8 16.2 24.2





#33

Chrysene

Concen: 92.977 ng

RT: 21.654 min Scan# 2442

Delta R.T. 0.009 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12

Instrument :

BNA_N

ClientSampleId :

PT-BN-WP





Tgt Ion:228 Resp: 4387068

Ion Ratio Lower Upper

228 100

226 32.0 24.2 36.2

229 20.7 15.8 23.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 200.016 ng

RT: 21.510 min Scan# 2426

Delta R.T. 0.009 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12




Tgt Ion:149 Resp: 4080457


Ion Ratio Lower Upper

149 100

167 28.1 20.2 30.4

279 3.0 2.3 3.5





#35

Perylene-d12

Concen: 0.400 ng

RT: 24.093 min Scan# 31

Delta R.T. 0.006 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12

Instrument :

BNA_N

ClientSampleId :

PT-BN-WP





Tgt Ion:264 Resp: 10043

Ion Ratio Lower Upper

264 100

260 26.3 21.3 31.9

265 62.8 63.0 94.4#



#36

Indeno(1,2,3-cd)pyrene

Concen: 92.171 ng

RT: 26.730 min Scan# 4005

Delta R.T. 0.044 min

Lab File: BN022017.D

Acq: 06 Oct 2022 12:12




Tgt Ion:276 Resp: 4677222


Ion Ratio Lower Upper

276 100

138 31.3 28.0 42.0

277 24.7 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 111.437 ng
RT: 23.397 min Scan# 2865
Delta R.T. 0.073 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Tgt Ion:252 Resp: 5304913
Ion Ratio Lower Upper
252 100
253 22.9 20.6 31.0
125 13.6 15.6 23.4#



#38
Benzo(k)fluoranthene
Concen: 111.042 ng
RT: 23.397 min Scan# 2865
Delta R.T. 0.023 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion:252 Resp: 5262592
Ion Ratio Lower Upper
252 100
253 22.9 20.4 30.6
125 13.6 15.8 23.6#









#39
Benzo(a)pyrene
Concen: 109.137 ng
RT: 23.994 min Scan# 3069
Delta R.T. 0.018 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12




Instrument :
BNA_N
ClientSampleId :
PT-BN-WP


Tgt Ion:252 Resp: 3962895
Ion Ratio Lower Upper
252 100
253 22.8 23.2 34.8#
125 16.6 21.1 31.7#



#40
Dibenzo(a,h)anthracene
Concen: 35.163 ng
RT: 26.745 min Scan# 4010
Delta R.T. 0.032 min
Lab File: BN022017.D
Acq: 06 Oct 2022 12:12

Tgt Ion:278 Resp: 1423017
Ion Ratio Lower Upper
278 100
139 25.8 19.9 29.9
279 23.3 21.9 32.9





#41

Benzo(g,h,i)perylene

Concen: 28.626 ng

RT: 27.508 min Scan# 41

Delta R.T. 0.021 min

Lab File: BN022017.D


Acq: 06 Oct 2022 12:12

Instrument :

BNA_N

ClientSampleId :

PT-BN-WP



Tgt Ion:276 Resp: 1250648

Ion Ratio Lower Upper

276 100

277 23.7 20.0 30.0

138 29.9 25.2 37.8

