

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100622\
 Data File : BN022015.D
 Acq On : 06 Oct 2022 10:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 00:39:18 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 01:32:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.356	11.0	114	0.00
3	n-Nitrosodimethylamine	0.400	0.365	8.8	123	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	123	0.00
5 S	Phenol-d6	0.400	0.372	7.0	127	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	128	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	132	0.00
8 S	Nitrobenzene-d5	0.400	0.363	9.3	131	0.00
9	Naphthalene	0.400	0.371	7.3	132	-0.01
10	Hexachlorobutadiene	0.400	0.370	7.5	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	135	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	151	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.400	0.350	12.5	141	0.00
15 S	2-Fluorobiphenyl	0.400	0.373	6.8	134	0.00
16	Acenaphthylene	0.400	0.345	13.8	134	0.00
17	Acenaphthene	0.400	0.369	7.8	136	0.00
18	Fluorene	0.400	0.398	0.5	150	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	134	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.420	-5.0	139	0.00
21	4-Bromophenyl-phenylether	0.400	0.371	7.3	135	0.00
22	Hexachlorobenzene	0.400	0.369	7.8	133	0.00
23	Atrazine	0.400	0.348	13.0	136	0.00
24	Pentachlorophenol	0.400	0.353	11.8	147	0.00
25	Phenanthrene	0.400	0.379	5.3	137	0.00
26	Anthracene	0.400	0.354	11.5	135	0.00
27 SURR	Fluoranthene-d10	0.400	0.364	9.0	136	0.00
28	Fluoranthene	0.400	0.361	9.8	134	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	127	0.00
30	Pyrene	0.400	0.397	0.8	134	0.00
31 S	Terphenyl-d14	0.400	0.393	1.8	132	0.00
32	Benzo(a)anthracene	0.400	0.358	10.5	126	0.00
33	Chrysene	0.400	0.372	7.0	123	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.362	9.5	130	0.00
35 I	Perylene-d12	0.400	0.400	0.0	115	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.345	13.8	110	0.00
37	Benzo(b)fluoranthene	0.400	0.386	3.5	122	0.00
38	Benzo(k)fluoranthene	0.400	0.379	5.3	117	0.00
39 C	Benzo(a)pyrene	0.400	0.362	9.5	115	0.00
40	Dibenzo(a,h)anthracene	0.400	0.345	13.8	109	0.00
41	Benzo(g,h,i)perylene	0.400	0.339	15.3	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4