

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101020\
 Data File : BN012135.D
 Acq On : 09 Oct 2020 22:43
 Operator : CG/JU
 Sample : L4342-07
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BACKFILL OFFSITE-2-SO-2

Quant Time: Oct 10 01:58:36 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 09 18:10:29 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	2152	0.40	ng	0.00
7) Naphthalene-d8	10.415	136	8729	0.40	ng	# 0.00
13) Acenaphthene-d10	14.268	164	4943	0.40	ng	-0.01
19) Phenanthrene-d10	17.029	188	10425	0.40	ng	# 0.00
29) Chrysene-d12	21.227	240	12017	0.40	ng	# 0.00
36) Perylene-d12	23.425	264	12652	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.243	112	2815	0.37	ng	0.00
5) Phenol-d6	6.813	99	3474	0.37	ng	0.00
8) Nitrobenzene-d5	8.780	82	3662	0.41	ng	0.00
11) 2-Methylnaphthalene-d10	12.006	152	5426	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.778	330	1150	0.48	ng	0.00
15) 2-Fluorobiphenyl	12.898	172	7333	0.39	ng	0.00
27) Fluoranthene-d10	19.062	212	12568	0.41	ng	0.00
31) Terphenyl-d14	19.668	244	12318	0.45	ng	0.00
Target Compounds						Qvalue
28) Fluoranthene	19.092	202	760	0.02	ng	# 35
30) Pyrene	19.459	202	960	0.02	ng	97
32) Benzo(a)anthracene	21.216	228	1026	0.03	ng	# 80
33) Chrysene	21.258	228	1240	0.03	ng	# 87
34) Bis(2-ethylhexyl)phtha...	21.152	149	18083	0.66	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101020\
Data File : BN012135.D
Acq On : 09 Oct 2020 22:43
Operator : CG/JU
Sample : L4342-07
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BACKFILL OFFSITE-2-SO-2

Quant Time: Oct 10 01:58:36 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 09 18:10:29 2020
Response via : Initial Calibration

