

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101020\
 Data File : BN012126.D
 Acq On : 09 Oct 2020 17:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 09 18:18:05 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 09 18:10:29 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	58	0.00
2	1,4-Dioxane	0.696	0.568	18.4	52	0.00
3	n-Nitrosodimethylamine	1.001	0.850	15.1	57	0.00
4 S	2-Fluorophenol	1.401	1.276	8.9	59	0.00
5 S	Phenol-d6	1.767	1.532	13.3	56	0.00
6	bis(2-Chloroethyl)ether	1.511	1.203	20.4	52	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	59	0.01
8 S	Nitrobenzene-d5	0.413	0.363	12.1	59	0.00
9	Naphthalene	1.146	0.985	14.0	56	0.01
10	Hexachlorobutadiene	0.187	0.195	-4.3	66	0.00
11 SURR	2-Methylnaphthalene-d10	0.633	0.569	10.1	59	0.00
12	2-Methylnaphthalene	0.729	0.651	10.7	58	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	65	0.01
14 S	2,4,6-Tribromophenol	0.194	0.204	-5.2	78	0.01
15 S	2-Fluorobiphenyl	1.513	1.311	13.4	61	0.01
16	Acenaphthylene	1.928	1.678	13.0	64	0.00
17	Acenaphthene	1.320	1.101	16.6	61	0.00
18	Fluorene	1.612	1.381	14.3	62	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	66	0.01
20	4,6-Dinitro-2-methylphenol	0.076	0.063	17.1	64	0.01
21	4-Bromophenyl-phenylether	0.212	0.198	6.6	68	0.00
22	Hexachlorobenzene	0.247	0.243	1.6	71	0.00
23	Atrazine	0.203	0.198	2.5	74	0.00
24	Pentachlorophenol	0.121	0.101	16.5	63	0.00
25	Phenanthrene	1.227	1.079	12.1	65	0.00
26	Anthracene	1.075	0.936	12.9	65	0.01
27 SURR	Fluoranthene-d10	1.174	1.091	7.1	69	0.00
28	Fluoranthene	1.371	1.364	0.5	74	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
30	Pyrene	1.491	1.253	16.0	75	0.00
31 S	Terphenyl-d14	0.914	0.738	19.3	72	0.00
32	Benzo(a)anthracene	1.285	1.094	14.9	78	0.01
33	Chrysene	1.383	1.224	11.5	82	0.01
34	Bis(2-ethylhexyl)phthalate	0.907	0.864	4.7	90	0.00
35	Indeno(1,2,3-cd)pyrene	1.570	1.467	6.6	85	0.01
36 I	Perylene-d12	1.000	1.000	0.0	89	0.00
37	Benzo(b)fluoranthene	1.368	1.101	19.5	81	0.00
38	Benzo(k)fluoranthene	1.487	1.209	18.7	77	0.00
39 C	Benzo(a)pyrene	1.261	1.049	16.8	81	0.00
40	Dibenzo(a,h)anthracene	1.305	1.128	13.6	85	0.00
41	Benzo(g,h,i)perylene	1.410	1.190	15.6	82	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4