

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101020\
 Data File : BN012126.D
 Acq On : 09 Oct 2020 17:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 09 18:18:05 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 09 18:10:29 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	58	0.00
2	1,4-Dioxane	0.400	0.326	18.5	52	0.00
3	n-Nitrosodimethylamine	0.400	0.339	15.3	57	0.00
4 S	2-Fluorophenol	0.400	0.364	9.0	59	0.00
5 S	Phenol-d6	0.400	0.347	13.3	56	0.00
6	bis(2-Chloroethyl)ether	0.400	0.319	20.3	52	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	59	0.01
8 S	Nitrobenzene-d5	0.400	0.352	12.0	59	0.00
9	Naphthalene	0.400	0.344	14.0	56	0.01
10	Hexachlorobutadiene	0.400	0.417	-4.2	66	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	59	0.00
12	2-Methylnaphthalene	0.400	0.357	10.8	58	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	65	0.01
14 S	2,4,6-Tribromophenol	0.400	0.419	-4.7	78	0.01
15 S	2-Fluorobiphenyl	0.400	0.347	13.3	61	0.01
16	Acenaphthylene	0.400	0.348	13.0	64	0.00
17	Acenaphthene	0.400	0.334	16.5	61	0.00
18	Fluorene	0.400	0.343	14.2	62	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	66	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.329	17.8	64	0.01
21	4-Bromophenyl-phenylether	0.400	0.374	6.5	68	0.00
22	Hexachlorobenzene	0.400	0.394	1.5	71	0.00
23	Atrazine	0.400	0.390	2.5	74	0.00
24	Pentachlorophenol	0.400	0.334	16.5	63	0.00
25	Phenanthrene	0.400	0.352	12.0	65	0.00
26	Anthracene	0.400	0.348	13.0	65	0.01
27 SURR	Fluoranthene-d10	0.400	0.372	7.0	69	0.00
28	Fluoranthene	0.400	0.398	0.5	74	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	83	0.00
30	Pyrene	0.400	0.336	16.0	75	0.00
31 S	Terphenyl-d14	0.400	0.323	19.3	72	0.00
32	Benzo(a)anthracene	0.400	0.340	15.0	78	0.01
33	Chrysene	0.400	0.354	11.5	82	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.381	4.8	90	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.374	6.5	85	0.01
36 I	Perylene-d12	0.400	0.400	0.0	89	0.00
37	Benzo(b)fluoranthene	0.400	0.322	19.5	81	0.00
38	Benzo(k)fluoranthene	0.400	0.325	18.8	77	0.00
39 C	Benzo(a)pyrene	0.400	0.333	16.8	81	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	85	0.00
41	Benzo(g,h,i)perylene	0.400	0.338	15.5	82	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4