

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN101020\  
 Data File : BN012141.D  
 Acq On : 10 Oct 2020 02:20  
 Operator : CG/JU  
 Sample : L4324-01  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 UST-CONTENT-LIQUID

Manual Integrations  
 APPROVED

mohammad  
 10/12/2020 11:35:58 AM

Quant Time: Oct 10 05:14:04 2020  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN092320.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Oct 09 18:10:29 2020  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.635  | 152  | 2148     | 0.40   | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.415 | 136  | 11957    | 0.40   | ng    | # 0.00   |
| 13) Acenaphthene-d10          | 14.268 | 164  | 9585     | 0.40   | ng    | -0.01    |
| 19) Phenanthrene-d10          | 17.017 | 188  | 17001    | 0.40   | ng    | #-0.01   |
| 29) Chrysene-d12              | 21.227 | 240  | 16347    | 0.40   | ng    | # 0.00   |
| 36) Perylene-d12              | 23.422 | 264  | 15836    | 0.40   | ng    | # 0.00   |
|                               |        |      |          |        |       |          |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 4) 2-Fluorophenol             | 5.243  | 112  | 978      | 0.13   | ng    | 0.00     |
| 5) Phenol-d6                  | 6.813  | 99   | 978      | 0.10   | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 8.780  | 82   | 3721     | 0.30   | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 12.002 | 152  | 8360     | 0.44   | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 15.765 | 330  | 1587     | 0.34   | ng    | -0.01    |
| 15) 2-Fluorobiphenyl          | 12.885 | 172  | 9698     | 0.27   | ng    | -0.01    |
| 27) Fluoranthene-d10          | 19.062 | 212  | 16702    | 0.33   | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.668 | 244  | 16342    | 0.44   | ng    | 0.00     |
|                               |        |      |          |        |       |          |
| Target Compounds              |        |      |          |        |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.229  | 88   | 9517m    | 2.55   | ng    |          |
| 9) Naphthalene                | 10.466 | 128  | 15629657 | 456.26 | ng    | 98       |
| 12) 2-Methylnaphthalene       | 12.082 | 142  | 2646334  | 121.49 | ng    | 99       |
| 16) Acenaphthylene            | 14.002 | 152  | 4231     | 0.09   | ng    | # 1      |
| 17) Acenaphthene              | 14.268 | 154  | 5398     | 0.17   | ng    | # 39     |
| 18) Fluorene                  | 15.321 | 166  | 3179     | 0.08   | ng    | # 85     |
| 24) Pentachlorophenol         | 16.676 | 266  | 409      | 0.08   | ng    | 89       |
| 25) Phenanthrene              | 17.066 | 178  | 33162    | 0.64   | ng    | # 90     |
| 26) Anthracene                | 17.151 | 178  | 2350     | 0.05   | ng    | # 80     |
| 28) Fluoranthene              | 19.092 | 202  | 26739    | 0.46   | ng    | # 93     |
| 30) Pyrene                    | 19.454 | 202  | 24182    | 0.40   | ng    | 95       |
| 32) Benzo(a)anthracene        | 21.206 | 228  | 5918     | 0.11   | ng    | # 46     |
| 33) Chrysene                  | 21.259 | 228  | 7696     | 0.14   | ng    | # 56     |
| 34) Bis(2-ethylhexyl)phtha... | 21.142 | 149  | 99383    | 2.68   | ng    | # 92     |
| 35) Indeno(1,2,3-cd)pyrene    | 25.613 | 276  | 1367     | 0.02   | ng    | # 77     |
| 37) Benzo(b)fluoranthene      | 22.765 | 252  | 5131     | 0.09   | ng    | # 1      |
| -----                         |        |      |          |        |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN101020\  
Data File : BN012141.D  
Acq On : 10 Oct 2020 02:20  
Operator : CG/JU  
Sample : L4324-01  
Misc :  
ALS Vial : 17 Sample Multiplier: 1


Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID

Manual Integrations  
APPROVED


mohammad  
10/12/2020 11:35:58 AM

Quant Time: Oct 10 05:14:04 2020  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN092320.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Oct 09 18:10:29 2020  
Response via : Initial Calibration









#1  
 1,4-Dichlorobenzene-d4  
 Concen: 0.40 ng  
 RT: 7.635 min Scan# 578  
 Delta R.T. -0.007 min  
 Lab File: BN012141.D  
 Acq: 10 Oct 2020 02:20





Tgt Ion:152 Resp: 2148  
 Ion Ratio Lower Upper  
 152 100  
 150 157.6 118.3 177.5  
 115 75.8 51.3 76.9



#2  
 1,4-Dioxane  
 Concen: 2.55 ng m  
 RT: 3.229 min Scan# 31  
 Delta R.T. 0.016 min  
 Lab File: BN012141.D  
 Acq: 10 Oct 2020 02:20




Tgt Ion: 88 Resp: 9517  
 Ion Ratio Lower Upper  
 88 100  
 58 53.5 63.3 94.9#  
 43 0.0 33.0 49.4#



Instrument :  
 BNA\_N  
 ClientSampleId :  
 UST-CONTENT-LIQUID

Manual Integrations  
 APPROVED

mohammad  
 10/12/2020 11:35:58 AM







#4  
2-Fluorophenol  
Concen: 0.13 ng  
RT: 5.243 min Scan# 281  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion: 112 Resp: 978  
Ion Ratio Lower Upper  
112 100  
64 66.9 49.0 73.6  
63 41.5 31.8 47.6

Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID




Manual Integrations  
APPROVED


mohammad  
10/12/2020 11:35:58 AM



#5  
Phenol-d6  
Concen: 0.10 ng  
RT: 6.813 min Scan# 476  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20


Tgt Ion: 99 Resp: 978  
Ion Ratio Lower Upper  
99 100  
42 80.3 23.4 35.0#  
71 203.5 38.1 57.1#





#7  
Naphthalene-d8  
Concen: 0.40 ng  
RT: 10.415 min Scan# 828  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20




Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID




|             |       |           |
|-------------|-------|-----------|
| Tgt Ion:136 | Resp: | 11957     |
| Ion Ratio   | Lower | Upper     |
| 136         | 100   |           |
| 137         | 20.1  | 9.5 14.3# |
| 54          | 42.6  | 5.6 8.4#  |
| 68          | 41.0  | 3.2 4.8#  |

Manual Integrations  
APPROVED



mohammad  
10/12/2020 11:35:58 AM




#8  
Nitrobenzene-d5  
Concen: 0.30 ng  
RT: 8.780 min Scan# 699  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20



|             |       |           |
|-------------|-------|-----------|
| Tgt Ion: 82 | Resp: | 3721      |
| Ion Ratio   | Lower | Upper     |
| 82          | 100   |           |
| 128         | 47.7  | 34.2 51.2 |
| 54          | 61.2  | 43.1 64.7 |









#9  
Naphthalene  
Concen: 456.26 ng  
RT: 10.466 min Scan# 832  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion:128 Resp:15629657  
Ion Ratio Lower Upper  
128 100  
129 13.1 9.7 14.5  
127 14.6 11.0 16.4

Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID




Manual Integrations  
APPROVED


mohammad  
10/12/2020 11:35:58 AM



#11  
2-Methylnaphthalene-d10  
Concen: 0.44 ng  
RT: 12.002 min Scan# 1024  
Delta R.T. -0.004 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20


Tgt Ion:152 Resp: 8360  
Ion Ratio Lower Upper  
152 100  
151 124.9 16.7 25.1#





#12  
2-Methylnaphthalene  
Concen: 121.49 ng  
RT: 12.082 min Scan# 1042  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20




Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID




Tgt Ion:142 Resp: 2646334  
Ion Ratio Lower Upper  
142 100  
141 87.7 70.0 105.0  
115 43.0 35.8 53.6

Manual Integrations  
APPROVED



mohammad  
10/12/2020 11:35:58 AM




#13  
Acenaphthene-d10  
Concen: 0.40 ng  
RT: 14.268 min Scan# 1276  
Delta R.T. -0.012 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20




Tgt Ion:164 Resp: 9585  
Ion Ratio Lower Upper  
164 100  
162 87.9 83.6 125.4  
160 45.8 41.2 61.8









#14  
2,4,6-Tribromophenol  
Concen: 0.34 ng  
RT: 15.765 min Scan# 1394  
Delta R.T. -0.012 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20





Tgt Ion: 330 Resp: 1587  
Ion Ratio Lower Upper  
330 100  
332 0.0 0.0 0.0  
141 197.0 33.7 50.5#



#15  
2-Fluorobiphenyl  
Concen: 0.27 ng  
RT: 12.885 min Scan# 1167  
Delta R.T. -0.012 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20




Tgt Ion: 172 Resp: 9698  
Ion Ratio Lower Upper  
172 100  
171 36.5 27.4 41.2  
170 25.9 17.8 26.6



Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID

Manual Integrations  
APPROVED

mohammad  
10/12/2020 11:35:58 AM







#16  
Acenaphthylene  
Concen: 0.09 ng  
RT: 14.002 min Scan# 1255  
Delta R.T. 0.013 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion:152 Resp: 4231  
Ion Ratio Lower Upper  
152 100  
151 89.5 15.9 23.9#  
153 76.3 10.5 15.7#

Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID




Manual Integrations  
APPROVED


mohammad  
10/12/2020 11:35:58 AM




#17  
Acenaphthene  
Concen: 0.17 ng  
RT: 14.268 min Scan# 1276  
Delta R.T. -0.063 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion:154 Resp: 5398  
Ion Ratio Lower Upper  
154 100  
153 25.7 83.4 125.2#  
152 31.7 42.8 64.2#









#18  
Fluorene  
Concen: 0.08 ng  
RT: 15.321 min Scan# 1359  
Delta R.T. -0.012 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20





Tgt Ion:166 Resp: 3179  
Ion Ratio Lower Upper  
166 100  
165 110.5 78.8 118.2  
167 27.9 11.2 16.8#



#19  
Phenanthrene-d10  
Concen: 0.40 ng  
RT: 17.017 min Scan# 1496  
Delta R.T. -0.012 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20




Tgt Ion:188 Resp: 17001  
Ion Ratio Lower Upper  
188 100  
94 0.0 0.0 0.0  
80 33.9 7.7 11.5#



Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID

Manual Integrations  
APPROVED

mohammad  
10/12/2020 11:35:58 AM







#24  
Pentachlorophenol  
Concen: 0.08 ng  
RT: 16.676 min Scan# 1468  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion: 266 Resp: 409  
Ion Ratio Lower Upper  
266 100  
264 76.3 51.2 76.8  
268 61.1 52.4 78.6

Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID




Manual Integrations  
APPROVED


mohammad  
10/12/2020 11:35:58 AM



#25  
Phenanthrene  
Concen: 0.64 ng  
RT: 17.066 min Scan# 1500  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion: 178 Resp: 33162  
Ion Ratio Lower Upper  
178 100  
176 20.2 15.1 22.7  
179 23.3 12.3 18.5#









#26  
 Anthracene  
 Concen: 0.05 ng  
 RT: 17.151 min Scan# 1507  
 Delta R.T. -0.012 min  
 Lab File: BN012141.D  
 Acq: 10 Oct 2020 02:20

Tgt Ion:178 Resp: 2350  
 Ion Ratio Lower Upper  
 178 100  
 176 20.4 14.6 21.8  
 179 0.0 12.4 18.6#

Instrument :  
 BNA\_N  
 ClientSampleId :  
 UST-CONTENT-LIQUID




Manual Integrations  
 APPROVED


mohammad  
 10/12/2020 11:35:58 AM




#27  
 Fluoranthene-d10  
 Concen: 0.33 ng  
 RT: 19.062 min Scan# 1789  
 Delta R.T. 0.000 min  
 Lab File: BN012141.D  
 Acq: 10 Oct 2020 02:20

Tgt Ion:212 Resp: 16702  
 Ion Ratio Lower Upper  
 212 100  
 106 15.6 6.4 9.6#  
 104 8.4 3.9 5.9#









#28  
Fluoranthene  
Concen: 0.46 ng  
RT: 19.092 min Scan# 1795  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20





Tgt Ion:202 Resp: 26739  
Ion Ratio Lower Upper  
202 100  
101 14.7 5.4 8.2#  
203 17.3 13.8 20.6



#29  
Chrysene-d12  
Concen: 0.40 ng  
RT: 21.227 min Scan# 2094  
Delta R.T. 0.000 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20




Tgt Ion:240 Resp: 16347  
Ion Ratio Lower Upper  
240 100  
120 39.2 3.5 5.3#  
236 34.6 20.5 30.7#



Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID

Manual Integrations  
APPROVED

mohammad  
10/12/2020 11:35:58 AM



#30

Pyrene

Concen: 0.40 ng

RT: 19.454 min Scan# 1868

Delta R.T. -0.005 min

Lab File: BN012141.D

Acq: 10 Oct 2020 02:20

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 202   | Resp: | 24182 |
| Ion Ratio | Lower | Upper |       |
| 202       | 100   |       |       |
| 200       | 22.2  | 16.5  | 24.7  |
| 203       | 20.4  | 13.9  | 20.9  |

Instrument :





BNA\_N

ClientSampleId :

UST-CONTENT-LIQUID

Manual Integrations  
APPROVED

mohammad  
10/12/2020 11:35:58 AM



#31

Terphenyl-d14

Concen: 0.44 ng




RT: 19.668 min Scan# 1911


Delta R.T. -0.005 min

Lab File: BN012141.D

Acq: 10 Oct 2020 02:20

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 244   | Resp: | 16342 |
| Ion Ratio | Lower | Upper |       |
| 244       | 100   |       |       |
| 212       | 17.6  | 7.3   | 10.9# |
| 122       | 40.2  | 7.8   | 11.6# |









#32  
Benzo(a)anthracene  
Concen: 0.11 ng  
RT: 21.206 min Scan# 2092  
Delta R.T. -0.010 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion:228 Resp: 5918  
Ion Ratio Lower Upper  
228 100  
226 46.1 21.0 31.4#  
229 54.5 16.0 24.0#

Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID




Manual Integrations  
APPROVED


mohammad  
10/12/2020 11:35:58 AM



#33  
Chrysene  
Concen: 0.14 ng  
RT: 21.259 min Scan# 2097  
Delta R.T. -0.010 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion:228 Resp: 7696  
Ion Ratio Lower Upper  
228 100  
226 44.4 23.3 34.9#  
229 50.2 15.8 23.8#









#34  
Bis(2-ethylhexyl)phthalate  
Concen: 2.68 ng  
RT: 21.142 min Scan# 2086  
Delta R.T. -0.010 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion:149 Resp: 99383  
Ion Ratio Lower Upper  
149 100  
167 24.6 23.6 35.4  
279 3.4 4.0 6.0#

Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID




Manual Integrations  
APPROVED


mohammad  
10/12/2020 11:35:58 AM




#35  
Indeno(1,2,3-cd)pyrene  
Concen: 0.02 ng  
RT: 25.613 min Scan# 3142  
Delta R.T. -0.017 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20

Tgt Ion:276 Resp: 1367  
Ion Ratio Lower Upper  
276 100  
138 21.7 13.0 19.4#  
277 9.8 20.0 30.0#









#36  
Perylene-d12  
Concen: 0.40 ng  
RT: 23.422 min Scan# 2502  
Delta R.T. -0.010 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20





Tgt Ion:264 Resp: 15836  
Ion Ratio Lower Upper  
264 100  
260 28.2 19.6 29.4  
265 61.6 38.6 58.0#



#37  
Benzo(b)fluoranthene  
Concen: 0.09 ng  
RT: 22.765 min Scan# 2310  
Delta R.T. -0.007 min  
Lab File: BN012141.D  
Acq: 10 Oct 2020 02:20



Tgt Ion:252 Resp: 5131  
Ion Ratio Lower Upper  
252 100  
253 86.3 19.1 28.7#  
125 241.6 6.2 9.2#



Instrument :  
BNA\_N  
ClientSampleId :  
UST-CONTENT-LIQUID

Manual Integrations  
APPROVED

mohammad  
10/12/2020 11:35:58 AM