


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102221\
 Data File : BN017036.D
 Acq On : 21 Oct 2021 08:01
 Operator : CG/JU
 Sample : DFTPP209
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP209

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN101921.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 22 15:39:11 2021



AutoFind: Scans 2453, 2454, 2455; Background Corrected with Scan 2447

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.5	50525	PASS
68	69	0.00	2	1.7	897	PASS
69	198	0.00	100	35.4	51840	PASS
70	69	0.00	2	0.3	144	PASS
127	198	10	80	46.6	68211	PASS
197	198	0.00	2	0.2	306	PASS
198	198	100	100	100.0	146528	PASS
199	198	5	9	6.4	9353	PASS
275	198	10	60	27.4	40112	PASS
365	198	1	100	3.8	5633	PASS
441	198	0.01	100	14.1	20680	PASS
442	198	50	100	93.5	136965	PASS
443	442	15	24	18.6	25419	PASS

m/z	Abundance
39.90	544.0
44.00	1296.0
207.00	632.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

m/z	Abundance
38.00	696.0
39.00	4263.0
40.00	653.0
44.00	1120.0
49.00	735.0
50.00	15330.0
51.00	61656.0
52.00	3420.0
54.90	439.0
56.00	1803.0
57.00	4053.0
61.00	900.0
62.00	951.0
63.00	2495.0
65.00	1169.0
68.00	1070.0
69.00	60928.0
70.00	432.0
72.90	612.0
74.00	6580.0
75.00	10370.0
76.00	3649.0
77.00	74624.0
78.10	5093.0
79.00	4778.0
80.00	3581.0
81.00	5223.0
82.00	1310.0
83.00	1251.0
83.90	368.0
85.00	816.0
86.00	1411.0
87.00	759.0
91.00	1037.0
92.00	1233.0
93.00	7761.0
93.90	621.0
96.00	381.0
98.00	6233.0
99.00	6184.0
99.90	560.0
100.90	2861.0
103.00	915.0
104.00	1688.0
105.00	1528.0
106.00	528.0
107.00	21320.0
108.00	3436.0
109.00	505.0
110.00	39560.0
111.00	6185.0
111.90	737.0
116.00	1212.0
117.00	18152.0
118.00	1588.0
122.00	1340.0
123.00	2376.0
123.90	996.0
125.00	951.0
127.00	76152.0
128.00	5985.0
129.00	28080.0
130.00	2426.0
131.00	528.0
133.90	888.0
135.00	2173.0
136.00	1038.0
137.00	1718.0
138.00	356.0
141.00	3621.0
142.00	1118.0
142.90	791.0
146.00	591.0
147.00	1949.0
148.00	4456.0
149.00	1265.0
151.00	449.0
152.90	1192.0
154.00	940.0
155.00	2216.0
156.00	3201.0
157.00	598.0
158.00	594.0
158.90	497.0
160.00	1040.0
161.00	1673.0
161.90	451.0
164.90	1355.0
166.00	1045.0
167.00	7663.0
168.00	7034.0
169.00	752.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

171.00	349.0
171.90	600.0
173.00	858.0
174.00	1517.0
175.00	2946.0
176.10	883.0
177.00	1228.0
177.90	431.0
179.00	5229.0
180.00	4363.0
181.00	1909.0
183.90	493.0
185.00	2851.0
186.00	20976.0
187.00	5932.0
188.00	702.0
189.00	1436.0
191.00	661.0
192.00	1760.0
193.00	1984.0
194.00	439.0
196.00	4384.0
196.80	919.0
198.00	151040.0
199.00	10258.0
200.00	916.0
201.40	768.0
203.00	1050.0
204.00	5730.0
205.00	9151.0
206.10	38352.0
207.10	5809.0
208.00	1457.0
209.00	530.0
210.20	641.0
211.00	1478.0
214.90	455.0
216.00	713.0
217.00	10179.0
218.00	1333.0
221.10	7138.0
221.80	1680.0
223.00	2178.0
224.10	21192.0
225.00	5267.0
226.00	763.0
227.00	8291.0
228.00	1140.0
229.00	1939.0
231.00	735.0
234.00	550.0
235.00	579.0
235.90	347.0
237.00	1010.0
239.00	381.0
241.00	542.0
242.00	1171.0
243.10	1247.0
244.00	17760.0
245.00	2150.0
246.00	3116.0
247.00	606.0
249.00	600.0
253.10	342.0
255.00	90328.0
256.00	12907.0
257.00	1239.0
258.00	4444.0
259.00	805.0
265.00	1856.0
273.00	2458.0
274.00	6668.0
275.00	38888.0
276.10	5843.0
277.00	3101.0
278.00	635.0
283.00	424.0
285.10	564.0
293.00	803.0
296.00	10158.0
297.00	1569.0
303.00	1275.0
304.00	308.0
314.10	508.0
315.00	1197.0
316.10	713.0
321.00	379.0
323.00	3927.0
324.00	674.0
327.00	725.0
328.00	334.0
333.10	317.0
334.00	2144.0
335.00	557.0
341.00	485.0
346.00	779.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

352.00	1176.0
353.00	766.0
354.10	1294.0
355.00	301.0
365.00	4654.0
365.90	746.0
371.00	354.0
372.00	1834.0
373.00	535.0
383.00	564.0
401.90	875.0
403.00	1312.0
404.00	495.0
421.00	1074.0
422.00	1046.0
423.10	6879.0
424.10	2471.0
425.00	397.0
441.00	16424.0
442.00	113552.0
443.10	22328.0
444.10	2060.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

m/z	Abundance
38.00	688.0
39.10	4408.0
40.00	654.0
44.00	1252.0
49.10	520.0
50.00	15266.0
51.00	58888.0
52.00	2966.0
54.90	399.0
56.00	1659.0
57.00	3824.0
61.00	808.0
62.00	787.0
63.00	2285.0
64.00	361.0
65.00	1254.0
68.00	1054.0
69.00	61440.0
73.00	649.0
74.00	6249.0
75.00	9332.0
76.00	3477.0
77.10	73200.0
78.00	5086.0
79.00	4778.0
80.00	3382.0
81.00	4806.0
82.00	1258.0
83.00	1189.0
83.90	321.0
85.00	691.0
86.00	1591.0
87.00	596.0
91.00	1034.0
92.00	1466.0
93.00	8263.0
94.00	612.0
96.00	396.0
98.00	5977.0
99.00	5844.0
100.00	463.0
101.00	2798.0
103.00	961.0
104.00	1707.0
105.00	1861.0
106.00	521.0
107.00	21632.0
108.00	3488.0
109.00	648.0
110.00	41696.0
111.00	6149.0
112.00	731.0
116.00	1177.0
117.00	18928.0
118.00	1449.0
122.00	1479.0
123.00	2266.0
123.90	1080.0
125.00	936.0
127.00	82200.0
128.00	6184.0
129.00	29632.0
130.00	2695.0
131.00	532.0
134.00	818.0
135.00	2457.0
136.00	1182.0
137.00	1518.0
137.90	364.0
140.00	389.0
141.00	4136.0
142.00	1325.0
142.90	901.0
145.90	750.0
147.00	2223.0
148.00	4782.0
149.00	1102.0
151.10	496.0
153.00	1061.0
154.00	890.0
155.00	2146.0
156.00	3323.0
157.00	646.0
157.90	885.0
159.00	643.0
160.00	1228.0
161.00	1916.0
162.00	503.0
165.00	1429.0
166.00	1266.0
167.00	8797.0
168.00	6690.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

169.00	903.0
170.90	386.0
172.00	699.0
173.00	1016.0
174.00	1678.0
175.10	3005.0
176.10	916.0
177.00	1473.0
178.00	530.0
179.00	6125.0
180.00	4562.0
181.00	2162.0
181.90	308.0
183.90	566.0
185.00	2875.0
186.00	23632.0
187.00	6825.0
188.00	711.0
189.00	1466.0
191.00	767.0
192.00	2022.0
193.00	2251.0
194.00	546.0
196.00	4846.0
198.00	176320.0
199.00	11225.0
200.00	870.0
201.50	914.0
203.00	1070.0
204.00	6211.0
205.00	10508.0
206.00	45384.0
207.10	6787.0
208.00	1481.0
209.00	543.0
210.10	653.0
211.00	1681.0
215.00	465.0
216.00	1084.0
217.00	10977.0
218.00	1466.0
221.00	7390.0
223.00	2593.0
224.00	24248.0
225.00	6446.0
226.10	719.0
227.00	9651.0
228.00	1465.0
229.00	2061.0
231.00	756.0
234.00	580.0
235.00	756.0
236.00	562.0
237.00	1076.0
238.90	353.0
240.00	384.0
241.00	514.0
242.00	1506.0
243.10	1400.0
244.00	21960.0
245.00	2639.0
246.00	3871.0
247.00	875.0
249.00	691.0
253.10	421.0
255.00	104056.0
256.00	13821.0
257.00	1364.0
258.00	5790.0
259.00	899.0
265.00	2575.0
266.00	320.0
273.00	3505.0
274.00	8861.0
275.00	49528.0
276.00	6906.0
277.00	4094.0
278.00	726.0
283.00	427.0
285.10	806.0
293.00	1010.0
294.00	316.0
296.00	13651.0
297.00	1968.0
303.00	1516.0
304.00	448.0
314.10	628.0
315.10	1598.0
316.00	978.0
321.00	427.0
323.00	4590.0
324.00	854.0
327.00	1002.0
327.90	356.0
332.00	339.0
333.00	489.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

334.00	2819.0
335.00	741.0
341.00	609.0
346.00	1120.0
352.00	1584.0
353.00	1173.0
354.00	1774.0
365.00	6909.0
366.00	1048.0
371.00	482.0
372.00	2422.0
373.10	673.0
383.00	920.0
389.90	434.0
402.00	1096.0
402.90	1664.0
404.00	671.0
421.00	1488.0
422.00	1492.0
423.00	10794.0
424.00	3144.0
425.00	401.0
441.00	24800.0
442.00	162304.0
443.10	29352.0
444.10	2953.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

m/z	Abundance
38.10	448.0
39.00	2230.0
39.90	657.0
44.00	1329.0
48.90	481.0
50.00	8152.0
51.00	31032.0
52.00	1613.0
56.00	945.0
57.00	2199.0
61.00	433.0
62.00	412.0
63.00	1217.0
65.00	644.0
68.00	566.0
69.00	33152.0
73.00	340.0
74.00	3334.0
75.00	5709.0
76.00	1707.0
77.00	39296.0
78.00	2819.0
79.00	2251.0
79.90	1846.0
81.00	2780.0
82.00	676.0
82.90	690.0
83.90	342.0
84.90	427.0
86.00	812.0
86.90	364.0
91.00	610.0
91.90	696.0
93.00	4641.0
93.90	330.0
98.00	3420.0
99.00	3065.0
101.00	1637.0
103.00	477.0
104.00	972.0
105.00	895.0
107.00	12990.0
108.00	1904.0
108.90	360.0
110.00	22592.0
111.00	3653.0
112.00	311.0
116.00	836.0
117.00	10192.0
118.00	881.0
122.00	895.0
123.00	1191.0
123.90	572.0
124.90	526.0
127.00	46280.0
128.00	3430.0
129.00	17216.0
130.00	1512.0
133.90	554.0
135.00	1271.0
135.90	443.0
137.00	1011.0
141.00	2289.0
142.00	690.0
143.00	574.0
146.00	385.0
147.00	1206.0
148.00	2745.0
149.00	693.0
152.90	673.0
153.90	643.0
155.00	1212.0
156.00	1951.0
158.00	346.0
159.00	318.0
160.00	740.0
160.90	1051.0
162.10	332.0
165.00	958.0
166.00	717.0
167.00	5062.0
168.00	3742.0
169.00	455.0
171.90	370.0
173.00	570.0
174.00	938.0
175.00	1826.0
176.10	566.0
176.90	844.0
179.00	3584.0
180.00	2797.0
181.00	1136.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

185.00	1734.0
186.00	14457.0
187.00	3968.0
188.00	396.0
189.00	876.0
191.00	543.0
191.90	1134.0
193.00	1437.0
193.90	323.0
196.00	2964.0
197.10	664.0
198.00	112224.0
199.00	6575.0
200.00	567.0
201.40	438.0
203.00	771.0
204.00	3648.0
205.00	6206.0
206.00	27896.0
207.10	4354.0
208.00	972.0
208.90	407.0
211.10	1120.0
216.00	556.0
217.00	6879.0
218.00	874.0
221.00	4434.0
223.00	1475.0
224.00	15392.0
225.00	3955.0
226.00	413.0
227.00	6766.0
228.00	1027.0
229.00	1420.0
231.00	552.0
234.00	442.0
235.00	488.0
237.00	561.0
241.00	358.0
242.00	892.0
243.10	1003.0
244.00	13771.0
245.00	1809.0
246.00	2638.0
247.00	519.0
249.00	597.0
253.00	364.0
254.10	392.0
255.00	66336.0
256.00	9044.0
257.00	757.0
258.00	3926.0
259.00	595.0
265.00	1698.0
273.00	2128.0
274.00	5859.0
275.00	31920.0
276.00	4135.0
277.00	2479.0
278.00	460.0
285.10	535.0
293.00	736.0
296.00	9559.0
297.00	1402.0
303.00	1238.0
313.90	386.0
315.00	1185.0
316.00	597.0
321.00	317.0
323.10	3176.0
324.10	560.0
327.00	653.0
328.00	354.0
333.00	326.0
334.00	2227.0
335.00	502.0
341.10	417.0
345.90	610.0
352.10	1056.0
353.00	712.0
354.00	1356.0
365.00	5335.0
366.00	672.0
370.90	353.0
372.10	2124.0
373.00	660.0
383.00	564.0
390.00	444.0
402.00	883.0
403.00	1387.0
403.90	602.0
421.00	1149.0
422.00	1186.0
423.10	8232.0
424.00	1863.0
441.00	20816.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209

442.00	135040.0
443.00	24576.0
444.10	2134.0

Instrument :
BNA_N
ClientSampleId :
DFTPP209