


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102221\
 Data File : BN017071.D
 Acq On : 22 Oct 2021 14:07
 Operator : CG/JU
 Sample : DFTPP210
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP210

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN101921.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 22 15:39:11 2021



AutoFind: Scans 2452, 2453, 2454; Background Corrected with Scan 2446

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.1	51488	PASS
68	69	0.00	2	1.7	891	PASS
69	198	0.00	100	38.9	51275	PASS
70	69	0.00	2	0.2	111	PASS
127	198	10	80	49.6	65373	PASS
197	198	0.00	2	0.3	432	PASS
198	198	100	100	100.0	131771	PASS
199	198	5	9	6.7	8812	PASS
275	198	10	60	25.8	33965	PASS
365	198	1	100	3.5	4601	PASS
441	198	0.01	100	11.8	15499	PASS
442	198	50	100	75.9	99981	PASS
443	442	15	24	19.7	19704	PASS

m/z	Abundance
39.90	502.0
44.00	963.0
207.00	501.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

m/z	Abundance
38.00	512.0
39.00	3254.0
40.00	589.0
44.00	1164.0
48.90	397.0
50.00	10700.0
51.00	43176.0
52.10	2223.0
56.00	1235.0
57.00	3074.0
61.00	729.0
62.00	546.0
63.00	1654.0
65.00	869.0
68.00	776.0
69.00	41504.0
73.00	330.0
74.00	4083.0
75.00	6604.0
76.00	2300.0
77.00	48512.0
78.00	3476.0
79.00	3221.0
80.00	2434.0
81.00	3092.0
81.90	914.0
83.00	814.0
83.90	385.0
85.00	419.0
85.90	1066.0
86.90	411.0
91.00	733.0
92.00	972.0
93.00	5524.0
96.00	354.0
98.00	4155.0
99.00	4711.0
100.00	353.0
101.00	1970.0
102.90	602.0
104.00	1016.0
105.00	1006.0
106.00	366.0
107.00	13869.0
108.00	2162.0
109.00	347.0
110.00	25872.0
111.00	3887.0
111.90	490.0
116.00	689.0
117.00	12173.0
117.90	1137.0
122.00	899.0
123.00	1424.0
124.00	729.0
124.90	634.0
127.00	48432.0
128.00	3959.0
129.00	18048.0
130.00	1653.0
134.00	505.0
135.00	1439.0
136.00	620.0
137.00	1326.0
141.00	2433.0
142.00	727.0
143.00	680.0
145.90	437.0
147.00	1224.0
148.00	2916.0
149.00	876.0
151.10	328.0
153.00	754.0
154.00	605.0
155.00	1167.0
156.00	1850.0
158.00	428.0
159.00	422.0
160.00	716.0
161.00	981.0
161.90	324.0
165.00	908.0
166.00	662.0
167.00	4733.0
168.00	5264.0
169.00	699.0
171.90	325.0
173.00	487.0
174.00	898.0
175.00	1619.0
176.00	480.0
176.90	793.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

179.00	3269.0
180.00	2541.0
181.00	924.0
185.00	1466.0
186.00	12278.0
187.00	3578.0
188.00	448.0
189.00	684.0
190.90	376.0
192.00	1023.0
193.00	1358.0
196.00	2428.0
196.80	459.0
197.00	465.0
198.00	87792.0
199.00	6299.0
199.90	532.0
201.50	455.0
203.00	588.0
204.00	3244.0
205.00	5700.0
206.10	23328.0
207.00	3522.0
208.00	921.0
208.90	336.0
210.20	317.0
210.90	902.0
216.00	553.0
217.00	5869.0
218.00	953.0
221.10	3739.0
221.90	726.0
223.00	1337.0
224.10	11508.0
225.00	3103.0
226.00	438.0
227.00	5113.0
228.00	790.0
229.00	951.0
231.00	509.0
234.00	353.0
235.00	382.0
237.00	609.0
242.00	664.0
243.00	700.0
244.10	9108.0
245.00	1431.0
246.00	1967.0
247.00	350.0
249.00	380.0
254.00	326.0
255.00	47824.0
256.00	6871.0
257.00	746.0
258.00	2701.0
258.90	418.0
264.90	1163.0
273.00	1302.0
274.00	3916.0
275.00	20880.0
276.00	3250.0
277.00	2067.0
278.00	305.0
293.00	439.0
296.00	5864.0
296.90	816.0
303.00	763.0
315.00	604.0
316.10	325.0
323.00	1683.0
324.00	423.0
326.80	304.0
334.00	1278.0
334.90	321.0
346.00	345.0
352.00	674.0
353.00	356.0
353.90	559.0
365.00	2567.0
366.00	346.0
372.00	1004.0
383.00	350.0
401.90	411.0
403.00	495.0
420.90	468.0
422.00	458.0
423.00	3299.0
424.00	1006.0
441.00	6842.0
442.10	50800.0
443.10	9623.0
444.00	850.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

m/z	Abundance
38.00	703.0
39.00	4335.0
40.00	614.0
44.00	1084.0
49.00	554.0
50.00	15189.0
51.00	61600.0
52.00	3069.0
56.00	1718.0
57.00	4061.0
61.00	791.0
62.00	858.0
63.00	2353.0
63.90	353.0
65.10	1051.0
68.00	1061.0
69.00	61824.0
70.00	334.0
73.00	602.0
74.00	6444.0
75.00	9823.0
76.00	3570.0
77.00	72480.0
78.00	5390.0
79.00	4817.0
80.00	3697.0
81.00	4992.0
82.00	1204.0
83.00	1180.0
85.00	757.0
86.00	1589.0
87.00	614.0
88.00	327.0
91.00	1046.0
92.00	1272.0
93.00	8036.0
94.00	536.0
95.90	302.0
98.00	6284.0
99.00	5585.0
100.00	519.0
101.00	2839.0
103.00	931.0
104.00	1768.0
105.00	1689.0
106.00	646.0
107.00	21624.0
108.00	3272.0
109.00	672.0
110.00	40792.0
111.00	6342.0
112.00	717.0
116.00	1118.0
117.00	18648.0
118.00	1311.0
120.00	309.0
122.00	1402.0
123.00	2270.0
123.90	969.0
124.90	1034.0
126.10	318.0
127.00	78032.0
128.00	6146.0
129.00	28400.0
130.00	2494.0
131.00	498.0
134.00	723.0
135.00	2376.0
136.00	970.0
137.00	1389.0
139.90	325.0
141.00	3997.0
142.00	1469.0
142.90	774.0
146.00	666.0
147.00	2052.0
148.00	4555.0
149.00	1131.0
151.00	489.0
151.80	322.0
152.90	1169.0
154.00	917.0
155.00	2134.0
156.00	3079.0
157.00	732.0
157.90	743.0
159.00	547.0
160.00	1224.0
161.00	1733.0
161.90	642.0
165.00	1433.0
166.00	1204.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

167.00	7710.0
168.00	6202.0
169.00	826.0
170.90	350.0
171.90	800.0
173.00	741.0
174.00	1765.0
175.00	2846.0
176.00	951.0
176.90	1470.0
178.00	435.0
179.00	5391.0
180.00	4321.0
181.00	1991.0
183.90	461.0
185.00	2684.0
186.00	21336.0
187.00	6392.0
188.00	618.0
189.00	1217.0
191.00	731.0
192.00	1849.0
193.00	2144.0
193.90	407.0
196.00	4374.0
196.70	838.0
198.00	158592.0
199.00	10672.0
200.00	869.0
201.50	754.0
203.00	1043.0
204.00	5671.0
205.00	9643.0
206.10	39488.0
207.00	6032.0
208.00	1378.0
209.00	542.0
210.00	594.0
211.00	1576.0
214.90	435.0
216.00	828.0
217.00	10367.0
218.00	1470.0
221.00	6432.0
223.00	2297.0
224.00	21320.0
225.00	5967.0
226.00	689.0
227.00	8980.0
228.00	1388.0
229.00	2036.0
231.00	789.0
234.00	556.0
234.90	574.0
235.90	449.0
237.00	855.0
239.00	378.0
240.90	570.0
242.10	1111.0
243.00	1112.0
244.00	17328.0
245.00	2532.0
246.00	3194.0
246.90	711.0
249.00	601.0
253.00	364.0
254.10	609.0
255.00	93664.0
256.00	13154.0
257.00	1059.0
258.00	5033.0
259.00	871.0
265.00	2106.0
273.00	2851.0
274.00	6911.0
275.00	42496.0
276.00	5956.0
277.00	3412.0
278.00	639.0
283.00	397.0
283.90	348.0
285.00	645.0
293.00	770.0
296.00	11159.0
297.00	1526.0
303.00	1455.0
304.00	326.0
314.00	525.0
315.00	1427.0
316.00	798.0
321.00	367.0
323.00	3768.0
324.10	791.0
326.90	687.0
328.00	429.0
332.00	355.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

333.00	393.0
334.00	2759.0
335.00	658.0
341.00	505.0
346.00	788.0
352.10	1017.0
353.00	811.0
354.00	1209.0
365.00	5394.0
366.00	821.0
372.00	2202.0
373.00	505.0
383.00	621.0
390.00	342.0
402.00	906.0
403.00	1274.0
403.90	389.0
421.00	868.0
422.10	1201.0
423.10	7291.0
424.00	2197.0
425.00	346.0
441.10	18064.0
442.10	117368.0
443.10	20928.0
444.00	2091.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

m/z	Abundance
38.00	530.0
39.00	3557.0
39.90	616.0
44.00	1058.0
48.90	402.0
50.00	12578.0
51.00	49688.0
52.00	2313.0
55.90	1337.0
57.00	3205.0
61.00	657.0
62.00	736.0
63.00	2027.0
65.00	1080.0
68.00	837.0
69.00	50496.0
73.00	439.0
74.00	5425.0
75.00	8688.0
76.00	2854.0
77.00	61416.0
78.00	3965.0
79.00	3976.0
80.00	2973.0
81.00	4190.0
82.00	892.0
83.00	919.0
83.90	322.0
85.00	702.0
86.00	1255.0
86.90	623.0
90.90	883.0
92.00	1129.0
93.00	6934.0
93.90	449.0
96.00	360.0
98.00	5218.0
99.00	4750.0
100.00	393.0
101.00	2426.0
103.00	769.0
104.00	1443.0
105.00	1482.0
106.00	442.0
107.00	19112.0
108.00	2835.0
109.00	492.0
110.00	35064.0
111.00	5558.0
111.90	665.0
116.00	1159.0
117.00	16400.0
118.00	1383.0
122.00	1117.0
123.00	1975.0
123.90	870.0
125.00	882.0
127.00	69656.0
128.00	5029.0
129.00	25656.0
129.90	2046.0
131.00	419.0
134.00	585.0
135.00	2253.0
135.90	816.0
137.00	1488.0
140.00	366.0
141.00	3471.0
142.00	1123.0
143.00	706.0
146.00	636.0
147.00	1751.0
148.00	4317.0
149.00	941.0
151.00	415.0
153.00	1144.0
154.00	823.0
155.00	1964.0
156.00	3040.0
156.90	598.0
158.00	680.0
159.00	516.0
160.00	1050.0
161.00	1649.0
161.90	438.0
164.90	1442.0
166.00	1119.0
167.00	7375.0
168.00	5062.0
168.90	788.0
170.90	371.0
171.90	629.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

173.00	974.0
174.00	1613.0
175.00	2755.0
176.00	717.0
176.90	1182.0
177.90	423.0
179.00	4811.0
180.00	3621.0
181.00	1546.0
184.00	512.0
185.00	2492.0
186.00	20704.0
187.00	5854.0
188.10	674.0
189.00	1239.0
191.00	566.0
192.00	1805.0
193.00	1801.0
194.00	357.0
196.00	4429.0
197.00	693.0
198.00	148928.0
199.00	9465.0
199.90	685.0
201.50	583.0
203.00	910.0
204.00	4788.0
205.00	8509.0
206.00	38904.0
207.10	5642.0
208.00	1438.0
209.00	486.0
210.20	641.0
211.00	1517.0
215.00	337.0
216.00	727.0
217.00	10469.0
218.00	1234.0
221.00	6017.0
223.00	2190.0
224.00	21744.0
225.00	5496.0
226.00	611.0
227.00	8420.0
228.00	1210.0
229.00	1867.0
231.00	896.0
233.90	606.0
235.00	666.0
235.90	443.0
237.00	812.0
238.90	311.0
241.00	499.0
242.00	1175.0
243.00	1258.0
244.00	18136.0
245.00	2289.0
246.00	3303.0
247.00	682.0
249.00	564.0
253.10	348.0
255.00	86376.0
256.00	12173.0
257.00	1037.0
258.00	5051.0
259.00	892.0
265.00	2250.0
265.90	399.0
270.90	335.0
273.00	2737.0
274.00	7466.0
275.00	38520.0
276.00	5207.0
277.00	3331.0
278.00	585.0
282.90	432.0
285.00	630.0
293.00	865.0
296.00	11407.0
297.00	1540.0
303.00	1415.0
304.00	326.0
314.00	484.0
315.00	1444.0
316.00	681.0
321.10	354.0
323.00	4189.0
324.00	921.0
327.00	741.0
328.00	491.0
332.00	414.0
333.00	339.0
334.00	2516.0
335.00	826.0
341.00	496.0
346.00	1008.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210

352.00	1383.0
353.00	949.0
354.10	1186.0
365.00	5842.0
366.00	792.0
371.00	497.0
372.10	2466.0
373.00	658.0
383.00	667.0
389.90	385.0
391.00	327.0
402.00	943.0
403.00	1282.0
404.00	571.0
421.00	1442.0
422.00	1172.0
423.10	8398.0
424.00	1979.0
441.00	21592.0
442.00	131776.0
443.10	28560.0
444.10	2606.0

Instrument :
BNA_N
ClientSampleId :
DFTPP210