

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102620\
 Data File : BN012401.D
 Acq On : 26 Oct 2020 17:08
 Operator : CG/JU
 Sample : PB132270BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB132270BL

Quant Time: Oct 26 17:40:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 26 15:06:07 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.578	152	831	0.40	ng	0.00
7) Naphthalene-d8	10.352	136	3190	0.40	ng	# 0.00
13) Acenaphthene-d10	14.217	164	1921	0.40	ng	0.00
19) Phenanthrene-d10	16.980	188	4043	0.40	ng	# 0.01
29) Chrysene-d12	21.184	240	4146	0.40	ng	# 0.00
36) Perylene-d12	23.361	264	4926	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.186	112	1014	0.35	ng	0.00
5) Phenol-d6	6.757	99	1061	0.31	ng	0.00
8) Nitrobenzene-d5	8.729	82	1382	0.39	ng	0.00
11) 2-Methylnaphthalene-d10	11.958	152	1809	0.35	ng	0.00
14) 2,4,6-Tribromophenol	15.740	330	426	0.31	ng	0.01
15) 2-Fluorobiphenyl	12.847	172	2728	0.38	ng	0.00
27) Fluoranthene-d10	19.018	212	4280	0.34	ng	0.00
31) Terphenyl-d14	19.623	244	4688	0.48	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102620\
Data File : BN012401.D
Acq On : 26 Oct 2020 17:08
Operator : CG/JU
Sample : PB132270BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB132270BL

Quant Time: Oct 26 17:40:55 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 26 15:06:07 2020
Response via : Initial Calibration

