

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102620\
 Data File : BN012400.D
 Acq On : 26 Oct 2020 16:03
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102620

Quant Time: Oct 26 17:02:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 26 15:06:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.651	0.570	12.4	98	0.00
3	n-Nitrosodimethylamine	1.581	1.384	12.5	92	0.00
4 S	2-Fluorophenol	1.398	1.235	11.7	93	0.00
5 S	Phenol-d6	1.655	1.432	13.5	92	0.00
6	bis(2-Chloroethyl)ether	1.127	1.015	9.9	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.441	0.373	15.4	95	0.00
9	Naphthalene	1.120	1.000	10.7	97	0.00
10	Hexachlorobutadiene	0.237	0.211	11.0	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.643	0.574	10.7	98	0.00
12	2-Methylnaphthalene	0.727	0.646	11.1	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.284	0.243	14.4	97	0.00
15 S	2-Fluorobiphenyl	1.488	1.299	12.7	97	0.00
16	Acenaphthylene	1.975	1.687	14.6	97	0.00
17	Acenaphthene	1.262	1.088	13.8	95	0.00
18	Fluorene	1.608	1.407	12.5	98	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.093	0.084	9.7	100	0.01
21	4-Bromophenyl-phenylether	0.243	0.207	14.8	96	0.00
22	Hexachlorobenzene	0.294	0.258	12.2	95	0.01
23	Atrazine	0.232	0.199	14.2	95	0.00
24	Pentachlorophenol	0.137	0.110	19.7	93	0.01
25	Phenanthrene	1.167	0.996	14.7	97	0.00
26	Anthracene	1.081	0.901	16.7	95	0.01
27 SURR	Fluoranthene-d10	1.235	1.059	14.3	96	0.00
28	Fluoranthene	1.388	1.177	15.2	95	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	94	0.01
30	Pyrene	1.353	1.195	11.7	96	0.00
31 S	Terphenyl-d14	0.939	0.825	12.1	96	0.00
32	Benzo(a)anthracene	1.217	1.055	13.3	94	0.00
33	Chrysene	1.365	1.161	14.9	95	0.00
34	Bis(2-ethylhexyl)phthalate	0.797	0.673	15.6	92	0.00
35	Indeno(1,2,3-cd)pyrene	1.961	1.727	11.9	93	0.00
36 I	Perylene-d12	1.000	1.000	0.0	93	0.00
37	Benzo(b)fluoranthene	1.262	1.071	15.1	93	0.00
38	Benzo(k)fluoranthene	1.375	1.213	11.8	96	0.00
39 C	Benzo(a)pyrene	1.225	1.039	15.2	91	0.00
40	Dibenzo(a,h)anthracene	1.383	1.190	14.0	92	0.00
41	Benzo(g,h,i)perylene	1.453	1.266	12.9	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN102620