

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102620\
 Data File : BN012400.D
 Acq On : 26 Oct 2020 16:03
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102620

Quant Time: Oct 26 17:02:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 26 15:06:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	94	0.00
2	1,4-Dioxane	0.400	0.350	12.5	98	0.00
3	n-Nitrosodimethylamine	0.400	0.350	12.5	92	0.00
4 S	2-Fluorophenol	0.400	0.353	11.8	93	0.00
5 S	Phenol-d6	0.400	0.346	13.5	92	0.00
6	bis(2-Chloroethyl)ether	0.400	0.360	10.0	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.338	15.5	95	0.00
9	Naphthalene	0.400	0.357	10.8	97	0.00
10	Hexachlorobutadiene	0.400	0.357	10.8	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.357	10.8	98	0.00
12	2-Methylnaphthalene	0.400	0.356	11.0	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.400	0.342	14.5	97	0.00
15 S	2-Fluorobiphenyl	0.400	0.349	12.8	97	0.00
16	Acenaphthylene	0.400	0.342	14.5	97	0.00
17	Acenaphthene	0.400	0.345	13.8	95	0.00
18	Fluorene	0.400	0.350	12.5	98	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.360	10.0	100	0.01
21	4-Bromophenyl-phenylether	0.400	0.341	14.8	96	0.00
22	Hexachlorobenzene	0.400	0.351	12.3	95	0.01
23	Atrazine	0.400	0.343	14.2	95	0.00
24	Pentachlorophenol	0.400	0.321	19.8	93	0.01
25	Phenanthrene	0.400	0.341	14.8	97	0.00
26	Anthracene	0.400	0.333	16.8	95	0.01
27 SURR	Fluoranthene-d10	0.400	0.343	14.2	96	0.00
28	Fluoranthene	0.400	0.339	15.3	95	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	94	0.01
30	Pyrene	0.400	0.353	11.8	96	0.00
31 S	Terphenyl-d14	0.400	0.351	12.3	96	0.00
32	Benzo(a)anthracene	0.400	0.347	13.3	94	0.00
33	Chrysene	0.400	0.340	15.0	95	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	92	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.352	12.0	93	0.00
36 I	Perylene-d12	0.400	0.400	0.0	93	0.00
37	Benzo(b)fluoranthene	0.400	0.340	15.0	93	0.00
38	Benzo(k)fluoranthene	0.400	0.353	11.8	96	0.00
39 C	Benzo(a)pyrene	0.400	0.339	15.3	91	0.00
40	Dibenzo(a,h)anthracene	0.400	0.344	14.0	92	0.00
41	Benzo(g,h,i)perylene	0.400	0.348	13.0	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN102620