

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN111020\
 Data File : BN012526.D
 Acq On : 10 Nov 2020 10:44
 Operator : CG/JU
 Sample : SST00.105
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST00.1105

Quant Time: Nov 10 12:35:35 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN111020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 12:31:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.46	152	884	0.40	ng/ul	-0.01
4) Naphthalene-d8	10.23	136	3140	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.12	164	1682	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.87	188	3471	0.40	ng/ul	0.00
19) Chrysene-d12	21.08	240	2678	0.40	ng/ul	0.00
23) Perylene-d12	23.20	264	2853	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.01	96	115	0.13	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.83	152	445	0.09	ng/ul	0.00
17) Fluoranthene-d10	18.91	212	893	0.10	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.04	88	119	0.122	ng/ul#	77
5) Naphthalene	10.28	128	878	0.099	ng/ul#	83
7) 2-Methylnaphthalene	11.91	142	585	0.095	ng/ul	99
8) 1-Methylnaphthalene	12.14	142	539	0.093	ng/ul	99
10) Acenaphthylene	13.84	152	727	0.097	ng/ul#	85
11) Acenaphthene	14.18	153	639	0.112	ng/ul	94
12) Fluorene	15.18	166	705	0.104	ng/ul#	94
14) Pentachlorophenol	16.53	266	90	0.079	ng/ul#	89
15) Phenanthrene	16.91	178	1219	0.120	ng/ul#	91
16) Anthracene	17.01	178	909	0.091	ng/ul#	86
18) Fluoranthene	18.94	202	1297	0.110	ng/ul#	89
20) Pyrene	19.31	202	1392	0.133	ng/ul#	91
21) Benzo(a)anthracene	21.12	228	1280	0.131	ng/ul	93
22) Chrysene	21.12	228	1208	0.120	ng/ul	93
24) Benzo(b)fluoranthene	22.58	252	1190	0.104	ng/ul	63
25) Benzo(k)fluoranthene	22.62	252	1424	0.125	ng/ul#	62
26) Benzo(a)pyrene	23.11	252	1167	0.112	ng/ul#	48
27) Indeno(1,2,3-cd)pyrene	25.29	276	1484	0.122	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.31	278	1190	0.117	ng/ul#	61
29) Benzo(a,h,i)perylene	25.93	276	1446	0.141	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN111020\
Data File : BN012526.D
Acq On : 10 Nov 2020 10:44
Operator : CG/JU
Sample : SST0.105
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.1105

Quant Time: Nov 10 12:35:35 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN111020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 12:31:02 2020
Response via : Initial Calibration

