

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN111020\
 Data File : BN012531.D
 Acq On : 10 Nov 2020 13:44
 Operator : CG/JU
 Sample : SSTD3.204
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2104

Quant Time: Nov 10 14:24:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN111020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 12:31:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	862	0.40	ng/ul	0.00
4) Naphthalene-d8	10.24	136	3349	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.12	164	1851	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.87	188	3804	0.40	ng/ul	0.00
19) Chrysene-d12	21.08	240	3544	0.40	ng/ul	0.00
23) Perylene-d12	23.20	264	3360	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.00	96	3526	4.00	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.83	152	15859	2.85	ng/ul	0.00
17) Fluoranthene-d10	18.90	212	32556	3.20	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.03	88	3704	3.892	ng/ul#	90
5) Naphthalene	10.28	128	29556	3.114	ng/ul	94
7) 2-Methylnaphthalene	11.91	142	20164	3.083	ng/ul	99
8) 1-Methylnaphthalene	12.13	142	19161	3.102	ng/ul	98
10) Acenaphthylene	13.83	152	27467	3.321	ng/ul	96
11) Acenaphthene	14.18	153	22159	3.526	ng/ul	99
12) Fluorene	15.18	166	25778	3.471	ng/ul	99
14) Pentachlorophenol	16.52	266	3805	3.034	ng/ul	96
15) Phenanthrene	16.91	178	40728	3.669	ng/ul	98
16) Anthracene	17.00	178	37844	3.474	ng/ul	96
18) Fluoranthene	18.93	202	48442	3.741	ng/ul	96
20) Pyrene	19.30	202	47768	3.454	ng/ul	97
21) Benzo(a)anthracene	21.06	228	44106	3.403	ng/ul	96
22) Chrysene	21.11	228	46320	3.482	ng/ul	98
24) Benzo(b)fluoranthene	22.57	252	49001	3.628	ng/ul	86
25) Benzo(k)fluoranthene	22.61	252	50793	3.776	ng/ul#	88
26) Benzo(a)pyrene	23.10	252	43571	3.562	ng/ul#	78
27) Indeno(1,2,3-cd)pyrene	25.28	276	59041	4.106	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.29	278	48659	4.068	ng/ul#	89
29) Benzo(a,h,i)perylene	25.91	276	50949	4.208	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN111020\
Data File : BN012531.D
Acq On : 10 Nov 2020 13:44
Operator : CG/JU
Sample : SSTD3.204
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2104

Quant Time: Nov 10 14:24:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN111020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 12:31:02 2020
Response via : Initial Calibration

