

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111020\
 Data File : BN012545.D
 Acq On : 10 Nov 2020 23:16
 Operator : CG/JU
 Sample : L4485-10
 Misc : MDL-W-03
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MDL-WATER-03

Quant Time: Nov 11 00:52:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN111020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 14:27:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	652	0.40	ng/ul	0.00
4) Naphthalene-d8	10.25	136	2561	0.40	ng/ul	0.01
9) Acenaphthene-d10	14.12	164	1424	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.87	188	2673	0.40	ng/ul	0.00
19) Chrysene-d12	21.08	240	2373	0.40	ng/ul	0.00
23) Perylene-d12	23.19	264	2519	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.01	96	515	0.58	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.86	152	933	0.26	ng/ul	0.02
17) Fluoranthene-d10	18.91	212	1956	0.29	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.05	88	60	0.067	ng/ul#	79
5) Naphthalene	10.30	128	407	0.057	ng/ul#	48
7) 2-Methylnaphthalene	11.94	142	238	0.051	ng/ul	94
8) 1-Methylnaphthalene	12.14	142	249	0.056	ng/ul	95
10) Acenaphthylene	13.84	152	359	0.057	ng/ul#	64
11) Acenaphthene	14.19	153	308	0.057	ng/ul#	89
12) Fluorene	15.19	166	309	0.051	ng/ul#	92
14) Pentachlorophenol	16.54	266	40	0.055	ng/ul	95
15) Phenanthrene	16.92	178	508	0.057	ng/ul#	70
16) Anthracene	17.03	178	379	0.052	ng/ul#	53
18) Fluoranthene	18.94	202	653	0.067	ng/ul#	72
20) Pyrene	19.31	202	719	0.064	ng/ul#	80
21) Benzo(a)anthracene	21.06	228	574	0.066	ng/ul#	88
22) Chrysene	21.11	228	731	0.070	ng/ul#	86
24) Benzo(b)fluoranthene	22.57	252	674	0.063	ng/ul#	41
25) Benzo(k)fluoranthene	22.61	252	862	0.070	ng/ul#	39
26) Benzo(a)pyrene	23.10	252	652	0.064	ng/ul#	12
27) Indeno(1,2,3-cd)pyrene	25.29	276	793	0.060	ng/ul#	56
28) Dibenzo(a,h)anthracene	25.31	278	639	0.060	ng/ul#	18
29) Benzo(a,h,i)perylene	25.92	276	768	0.062	ng/ul#	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111020\
Data File : BN012545.D
Acq On : 10 Nov 2020 23:16
Operator : CG/JU
Sample : L4485-10
Misc : MDL-W-03
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MDL-WATER-03

Quant Time: Nov 11 00:52:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN111020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 14:27:13 2020
Response via : Initial Calibration

