


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN111020\
 Data File : BN012525.D
 Acq On : 10 Nov 2020 09:52
 Operator : CG/JU
 Sample : DFTPP012
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP012

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN111020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 10 14:27:13 2020



AutoFind: Scans 2437, 2438, 2439; Background Corrected with Scan 2431

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.1	122077	PASS
68	69	0.00	2	1.9	2374	PASS
69	198	0.00	100	52.1	124544	PASS
70	69	0.00	2	0.7	841	PASS
127	198	10	80	51.8	123752	PASS
197	198	0.00	2	0.2	479	PASS
198	198	100	100	100.0	238826	PASS
199	198	5	9	6.9	16446	PASS
275	198	10	60	29.5	70445	PASS
365	198	1	100	4.2	10108	PASS
441	198	0.01	100	14.6	34970	PASS
442	198	50	100	91.5	218488	PASS
443	442	15	24	19.3	42266	PASS

Scan 2431 (17.286 min): BN012525.D
DFTPP012

m/z	Abundance
40.00	529.0
44.00	2875.0
49.10	343.0
73.10	1397.0
193.00	328.0
207.10	1184.0
207.90	461.0
221.00	356.0
280.90	367.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

m/z	Abundance
37.10	856.0
38.00	1914.0
39.00	9486.0
40.00	1326.0
41.10	318.0
43.00	490.0
44.00	3339.0
45.10	392.0
49.00	1734.0
50.00	35616.0
51.00	119560.0
52.10	7196.0
54.90	1018.0
56.00	3651.0
57.00	6916.0
61.00	2038.0
62.00	1773.0
63.00	4900.0
64.10	522.0
65.10	1913.0
68.00	2104.0
69.00	119456.0
70.10	753.0
73.10	2121.0
74.00	14073.0
75.00	19784.0
76.00	6522.0
77.10	120360.0
78.10	8619.0
79.00	10545.0
80.00	7922.0
81.00	10063.0
82.00	2354.0
83.00	1687.0
84.00	463.0
85.10	1857.0
86.00	2911.0
87.00	1389.0
88.10	319.0
90.90	2466.0
92.00	3042.0
93.00	18352.0
94.00	1557.0
94.90	355.0
95.90	957.0
96.90	350.0
98.00	15574.0
99.00	10614.0
100.00	961.0
101.00	4587.0
101.90	421.0
103.00	1417.0
104.00	3231.0
105.00	3774.0
106.00	723.0
107.00	38960.0
108.00	4727.0
108.90	893.0
110.00	60512.0
111.00	10747.0
111.90	1108.0
112.90	565.0
115.00	523.0
115.90	2661.0
117.00	39360.0
118.00	2957.0
119.10	477.0
120.10	405.0
122.00	2725.0
123.00	4229.0
123.90	2474.0
125.00	2010.0
126.00	624.0
127.00	108752.0
128.00	8846.0
129.00	48432.0
130.00	3969.0
130.90	824.0
133.10	507.0
134.00	1108.0
134.90	3701.0
136.00	1221.0
137.10	3045.0
138.00	547.0
140.00	889.0
141.00	7518.0
142.00	2070.0
143.00	1588.0
144.00	889.0
145.00	362.0
146.00	1210.0
147.00	3712.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

148.00	8228.0
149.00	2274.0
149.90	421.0
151.20	784.0
151.80	420.0
153.00	1996.0
154.00	1039.0
155.00	3770.0
156.00	5181.0
157.20	805.0
157.90	1156.0
159.10	728.0
160.00	2013.0
161.00	2925.0
162.00	723.0
162.90	363.0
163.90	458.0
164.90	2549.0
166.00	1581.0
167.00	13380.0
168.00	10232.0
168.90	1480.0
170.00	686.0
171.00	526.0
171.90	1149.0
173.00	1307.0
174.00	2630.0
175.00	4638.0
175.90	1154.0
176.90	1967.0
178.00	644.0
179.00	8343.0
180.00	7204.0
181.00	3067.0
181.90	770.0
183.90	798.0
185.00	4942.0
186.00	30264.0
187.10	8702.0
188.00	1496.0
189.00	2417.0
189.90	408.0
191.00	1167.0
191.90	3338.0
193.00	3808.0
194.00	845.0
196.00	5436.0
198.00	200640.0
199.00	14018.0
199.90	1201.0
201.50	1043.0
203.00	1685.0
204.00	7301.0
205.00	13949.0
206.00	53696.0
207.00	8121.0
208.00	2826.0
209.10	1242.0
209.90	998.0
211.10	2204.0
211.90	387.0
215.00	864.0
216.00	1079.0
217.00	13708.0
218.10	1859.0
221.00	6137.0
221.90	2244.0
223.00	3352.0
224.00	27120.0
225.00	8045.0
226.10	902.0
227.00	13276.0
228.00	1740.0
229.00	2414.0
231.00	914.0
234.10	680.0
234.90	936.0
235.90	981.0
236.90	2078.0
238.10	416.0
238.90	714.0
240.10	334.0
240.90	589.0
242.00	1623.0
243.00	1353.0
244.00	22544.0
245.00	3440.0
246.00	5232.0
246.90	981.0
249.10	584.0
251.10	422.0
253.20	412.0
255.00	108640.0
256.00	16824.0
257.00	1791.0
258.00	8290.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

259.00	1493.0
265.00	3095.0
265.90	506.0
270.80	363.0
273.00	2718.0
274.00	7900.0
275.00	49440.0
275.90	7552.0
277.00	5315.0
277.90	1227.0
279.00	315.0
281.00	615.0
283.00	837.0
284.10	562.0
285.00	1033.0
293.00	1289.0
294.00	315.0
296.00	16968.0
297.00	2445.0
302.00	346.0
303.00	1734.0
303.90	835.0
313.90	887.0
315.00	1805.0
316.10	948.0
320.90	380.0
322.00	424.0
323.00	5017.0
323.90	850.0
327.00	1328.0
328.00	488.0
332.00	446.0
332.80	414.0
334.00	2066.0
335.00	821.0
341.10	461.0
345.80	1159.0
352.00	1156.0
353.10	533.0
354.10	1189.0
354.90	354.0
364.90	6453.0
366.00	1212.0
371.00	323.0
372.00	1996.0
373.10	1115.0
382.90	438.0
401.90	545.0
403.00	1284.0
404.00	406.0
420.80	1045.0
421.30	574.0
422.00	918.0
423.00	6166.0
424.10	2525.0
425.00	317.0
441.00	20496.0
442.00	120424.0
443.00	26336.0
444.00	2379.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

m/z	Abundance
37.00	819.0
38.10	2561.0
39.10	12083.0
40.00	1448.0
40.90	500.0
42.90	543.0
44.00	3425.0
45.00	331.0
49.00	1967.0
50.00	43720.0
51.10	153792.0
52.00	7589.0
53.10	497.0
55.00	1203.0
56.00	4889.0
57.00	10399.0
58.00	583.0
61.00	2351.0
62.00	2191.0
63.00	6091.0
64.00	646.0
65.00	2278.0
68.00	3062.0
69.00	152832.0
70.10	1070.0
73.00	2766.0
74.00	18840.0
75.00	27568.0
76.00	8700.0
77.10	160064.0
78.00	11243.0
79.00	13382.0
80.00	9360.0
81.00	13444.0
82.00	2811.0
83.10	2264.0
83.90	483.0
85.00	2561.0
86.00	4201.0
86.90	1841.0
88.10	569.0
89.00	364.0
91.00	3006.0
92.00	3319.0
93.00	22752.0
94.00	1424.0
95.00	354.0
96.00	784.0
98.00	20592.0
99.00	13339.0
100.00	1127.0
101.10	6369.0
101.90	332.0
103.00	2670.0
104.00	4824.0
104.90	4239.0
106.20	1451.0
107.00	50072.0
108.00	7102.0
109.10	1405.0
110.00	80104.0
111.00	14644.0
111.90	2501.0
112.90	383.0
113.10	378.0
116.10	3816.0
117.00	55712.0
118.00	4800.0
118.90	664.0
119.90	519.0
120.90	463.0
122.00	3816.0
123.00	5809.0
124.00	2521.0
125.00	2817.0
127.00	150144.0
128.00	11742.0
129.00	65600.0
130.00	5433.0
130.90	1520.0
132.00	611.0
132.90	328.0
133.20	364.0
134.00	2559.0
135.00	6082.0
136.00	2748.0
137.00	3029.0
138.10	604.0
139.10	522.0
139.90	799.0
141.00	10134.0
142.00	3413.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

143.00	1723.0
143.90	842.0
145.10	370.0
146.00	1866.0
147.00	5073.0
148.00	13593.0
149.00	2786.0
149.90	786.0
151.10	1130.0
152.00	716.0
153.00	2645.0
154.00	2258.0
155.00	5430.0
156.00	6364.0
157.10	997.0
158.00	2078.0
159.00	1263.0
160.10	3035.0
161.00	4117.0
162.10	1354.0
163.00	357.0
164.00	747.0
165.00	4007.0
166.00	2256.0
167.00	20752.0
168.00	11688.0
169.00	1497.0
169.90	691.0
170.90	694.0
171.90	1993.0
173.00	2142.0
174.00	4046.0
175.10	5587.0
175.90	2167.0
177.00	3916.0
178.00	1929.0
179.00	13118.0
180.00	10848.0
181.00	3750.0
182.10	431.0
183.90	1297.0
185.00	5878.0
186.00	47896.0
187.00	15270.0
188.10	1542.0
189.00	3766.0
190.00	831.0
191.00	1682.0
192.00	4966.0
193.00	5676.0
194.00	1290.0
195.00	463.0
196.00	9933.0
198.00	294272.0
199.00	19944.0
200.00	1759.0
201.50	1615.0
203.00	2376.0
204.00	14332.0
205.00	21728.0
206.00	88760.0
207.00	10771.0
208.00	3282.0
208.90	1408.0
210.00	1583.0
211.00	2552.0
215.10	1032.0
216.00	1876.0
217.00	24632.0
218.00	2936.0
220.20	603.0
221.00	10895.0
221.90	3361.0
223.00	4551.0
224.00	46696.0
225.10	11123.0
226.10	1480.0
227.00	23320.0
227.90	3460.0
229.00	5583.0
230.00	602.0
231.00	1341.0
232.00	313.0
233.00	430.0
233.90	1922.0
235.00	1888.0
236.00	1015.0
237.00	2919.0
238.10	415.0
238.90	707.0
240.10	512.0
241.00	1124.0
242.00	2273.0
243.00	2598.0
244.00	36144.0
245.00	4623.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

246.00	7515.0
246.90	1788.0
247.90	404.0
248.90	1449.0
251.00	368.0
251.80	356.0
252.90	805.0
255.00	187520.0
256.00	26928.0
257.00	2650.0
258.00	14433.0
259.00	2030.0
263.90	524.0
265.00	5200.0
266.00	747.0
269.90	353.0
271.00	375.0
272.00	452.0
273.00	5988.0
274.00	15763.0
275.00	88808.0
276.00	11803.0
277.00	8704.0
278.00	2304.0
279.00	433.0
280.90	705.0
282.10	586.0
283.10	1101.0
284.00	938.0
284.90	1267.0
285.90	335.0
291.00	342.0
292.00	368.0
293.00	1485.0
294.00	699.0
295.10	758.0
296.00	26392.0
297.00	3745.0
302.00	397.0
303.00	2366.0
304.00	857.0
308.00	362.0
309.90	332.0
312.90	381.0
314.00	1462.0
314.90	3463.0
316.00	2112.0
317.00	358.0
321.10	649.0
322.10	508.0
323.10	9218.0
324.00	1787.0
326.90	1700.0
328.00	858.0
331.90	653.0
333.00	736.0
333.90	5029.0
335.10	1256.0
341.10	1190.0
346.00	1263.0
346.90	336.0
351.90	2225.0
353.00	1405.0
353.90	3026.0
355.10	447.0
365.00	12286.0
365.90	2093.0
371.10	811.0
372.00	4232.0
373.00	1058.0
382.90	937.0
383.70	429.0
389.90	549.0
392.00	403.0
402.00	1777.0
403.00	2645.0
404.00	800.0
420.90	1963.0
422.00	1965.0
423.00	13573.0
424.00	4989.0
425.00	447.0
441.00	40728.0
442.10	264640.0
443.00	48072.0
444.00	4396.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

m/z	Abundance
37.00	305.0
38.00	1893.0
39.10	6556.0
39.90	807.0
41.00	458.0
44.00	3634.0
45.00	456.0
49.10	1292.0
50.00	26344.0
51.10	92880.0
52.00	5042.0
55.00	996.0
56.00	2825.0
57.00	6858.0
58.00	314.0
61.00	1625.0
62.00	1534.0
63.00	3625.0
64.00	812.0
65.00	1656.0
68.10	1958.0
69.00	101344.0
69.90	700.0
72.90	2209.0
74.00	11789.0
75.00	17128.0
76.00	5131.0
77.00	105672.0
78.10	6227.0
79.00	8203.0
80.00	5844.0
81.00	8675.0
82.10	1886.0
83.00	1903.0
83.90	395.0
85.00	1410.0
85.90	2881.0
87.00	1244.0
87.90	472.0
91.10	2426.0
92.00	2621.0
93.00	17208.0
94.00	1188.0
94.90	381.0
95.90	333.0
96.80	311.0
98.00	12197.0
99.00	9384.0
99.90	797.0
101.00	3921.0
102.00	439.0
102.90	1507.0
104.00	3121.0
105.00	3411.0
106.10	845.0
107.00	36864.0
108.00	5825.0
109.00	1188.0
110.00	56136.0
111.00	8222.0
112.00	1519.0
115.90	2236.0
117.00	40592.0
117.90	2971.0
118.90	317.0
119.90	821.0
121.90	2572.0
122.90	3711.0
124.00	1937.0
124.90	1552.0
127.00	112360.0
128.00	7785.0
129.00	47968.0
130.00	4332.0
130.90	682.0
131.80	374.0
132.20	394.0
132.90	368.0
134.00	1769.0
135.00	4271.0
136.00	1814.0
137.00	2595.0
138.00	460.0
139.90	676.0
141.00	7206.0
142.00	2765.0
142.90	1614.0
144.00	573.0
145.10	530.0
146.10	1351.0
147.00	3933.0
148.00	9495.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

149.00	2306.0
150.10	957.0
151.10	1144.0
152.10	509.0
153.00	2924.0
154.00	1406.0
155.10	2916.0
156.00	5462.0
157.00	676.0
158.00	1127.0
159.00	783.0
160.00	2385.0
161.00	3474.0
162.00	955.0
162.90	349.0
163.90	477.0
165.00	3237.0
166.00	2089.0
167.00	14893.0
168.00	8786.0
169.00	1468.0
169.90	460.0
171.00	905.0
172.00	1094.0
173.00	1583.0
174.10	3450.0
175.00	5390.0
176.10	1209.0
176.90	2846.0
178.00	1355.0
178.90	12401.0
180.00	7146.0
181.00	3016.0
182.00	642.0
182.90	372.0
183.90	583.0
185.10	5267.0
186.00	36408.0
187.00	9397.0
188.00	1460.0
189.00	2982.0
190.00	332.0
191.00	1192.0
192.00	3609.0
193.00	4202.0
193.90	1305.0
195.10	484.0
196.00	6607.0
196.90	1439.0
198.00	221568.0
198.90	15378.0
199.90	971.0
201.50	1279.0
203.00	2345.0
204.10	10535.0
205.00	16984.0
206.00	67728.0
207.00	9387.0
208.00	2696.0
209.00	1377.0
210.10	1078.0
211.00	3179.0
211.80	801.0
215.00	1743.0
216.00	1628.0
217.00	19072.0
218.00	2252.0
220.10	349.0
221.10	8975.0
222.00	2153.0
223.00	4149.0
224.00	35896.0
225.00	8650.0
226.10	1355.0
227.00	16142.0
228.00	2324.0
229.00	3411.0
229.90	638.0
231.00	1217.0
233.10	716.0
233.90	1289.0
235.00	1117.0
235.90	721.0
237.00	1856.0
239.10	691.0
240.00	691.0
241.00	916.0
242.00	1700.0
243.10	2603.0
244.10	28352.0
245.10	3016.0
246.00	6561.0
246.90	2001.0
248.00	576.0
249.00	1283.0
250.90	437.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012

251.80	457.0
253.00	722.0
254.10	1293.0
255.00	154112.0
256.00	25280.0
257.10	2643.0
258.00	10054.0
259.10	1756.0
259.90	306.0
265.00	4804.0
266.00	698.0
271.00	539.0
272.00	401.0
272.90	4250.0
274.00	11932.0
275.00	73088.0
276.00	9702.0
277.00	7238.0
278.00	1800.0
281.20	660.0
282.10	389.0
282.90	726.0
283.90	549.0
284.20	520.0
285.00	1410.0
285.90	365.0
293.00	1718.0
294.10	345.0
295.10	561.0
296.00	24808.0
296.90	3245.0
301.10	330.0
302.00	389.0
303.00	3002.0
303.90	938.0
309.00	369.0
310.00	447.0
313.90	1106.0
314.90	2954.0
316.00	1537.0
316.90	324.0
321.00	634.0
322.10	340.0
322.90	6376.0
324.00	1892.0
327.00	1626.0
327.90	699.0
332.00	770.0
333.00	715.0
334.00	3947.0
334.90	970.0
341.10	1056.0
345.90	1803.0
347.00	546.0
351.90	2297.0
353.00	1671.0
354.00	2193.0
355.00	424.0
365.00	11585.0
365.90	1765.0
371.00	704.0
372.00	3389.0
373.00	1181.0
382.90	756.0
390.00	609.0
392.00	352.0
400.80	471.0
402.00	1508.0
403.00	1692.0
404.00	694.0
420.90	1915.0
422.00	2262.0
423.00	13141.0
424.00	3352.0
425.10	476.0
441.00	43688.0
442.00	270400.0
443.00	52392.0
444.00	4499.0

Instrument :
BNA_N
ClientSampleId :
DFTPP012