

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111621\
 Data File : BN017473.D
 Acq On : 16 Nov 2021 13:34
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111621

Quant Time: Nov 16 14:36:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 16 14:06:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.176	0.161	8.5	106	0.00
3	n-Nitrosodimethylamine	0.335	0.332	0.9	108	-0.03
4 S	2-Fluorophenol	0.952	1.038	-9.0	105	0.00
5 S	Phenol-d6	1.024	1.109	-8.3	105	0.00
6	bis(2-Chloroethyl)ether	1.098	1.131	-3.0	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.248	0.243	2.0	107	0.00
9	Naphthalene	1.005	1.013	-0.8	104	0.00
10	Hexachlorobutadiene	0.210	0.212	-1.0	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.657	0.664	-1.1	104	0.00
12	2-Methylnaphthalene	0.636	0.653	-2.7	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.102	0.101	1.0	114	0.00
15 S	2-Fluorobiphenyl	1.509	1.558	-3.2	103	0.00
16	Acenaphthylene	1.490	1.395	6.4	107	0.00
17	Acenaphthene	1.081	1.074	0.6	105	0.00
18	Fluorene	1.242	1.243	-0.1	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.016	0.014	12.5	175#	0.00
21	4-Bromophenyl-phenylether	0.227	0.222	2.2	105	0.00
22	Hexachlorobenzene	0.256	0.257	-0.4	106	0.00
23	Atrazine	0.129	0.129	0.0	113	0.00
24	Pentachlorophenol	0.051	0.051	0.0	126	0.00
25	Phenanthrene	1.099	1.105	-0.5	105	0.00
26	Anthracene	0.912	0.872	4.4	106	0.00
27 SURR	Fluoranthene-d10	1.195	1.182	1.1	106	0.00
28	Fluoranthene	1.179	1.195	-1.4	106	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	105	0.01
30	Pyrene	1.437	1.452	-1.0	106	0.00
31 S	Terphenyl-d14	0.982	1.030	-4.9	106	0.00
32	Benzo(a)anthracene	1.238	1.218	1.6	110	0.00
33	Chrysene	1.274	1.272	0.2	105	0.00
34	Bis(2-ethylhexyl)phthalate	0.647	0.534	17.5	102	0.01
35 I	Perylene-d12	1.000	1.000	0.0	107	0.00
36	Indeno(1,2,3-cd)pyrene	0.923	1.005	-8.9	134	0.00
37	Benzo(b)fluoranthene	1.502	1.564	-4.1	114	0.00
38	Benzo(k)fluoranthene	1.381	1.399	-1.3	111	0.00
39 C	Benzo(a)pyrene	1.165	1.196	-2.7	116	0.00
40	Dibenzo(a,h)anthracene	0.702	0.792	-12.8	141	0.00
41	Benzo(g,h,i)perylene	0.868	0.934	-7.6	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN111621