



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111621\
 Data File : BN017465.D
 Acq On : 16 Nov 2021 07:31
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 16 14:06:59 2021






AutoFind: Scans 2499, 2500, 2501; Background Corrected with Scan 2492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.1	217323	PASS
68	69	0.00	2	1.9	4151	PASS
69	198	0.00	100	39.9	221461	PASS
70	69	0.00	2	0.6	1287	PASS
127	198	10	80	47.6	264149	PASS
197	198	0.00	2	0.6	3065	PASS
198	198	100	100	100.0	555456	PASS
199	198	5	9	6.9	38397	PASS
275	198	10	60	24.5	135845	PASS
365	198	1	100	2.9	16284	PASS
441	198	0.01	100	9.2	51360	PASS
442	442	50	100	100.0	328789	PASS
443	442	15	24	17.7	58181	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111621\
 Data File : BN017465.D
 Acq On : 16 Nov 2021 07:31
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Nov 16 15:53:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 16 15:52:55 2021
 Response via : Initial Calibration



TIC: BN017465.D\data.ms

(70) Pentachlorophenol (C)

16.886min (0.000) 20465.43 ng




response 610148

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.05
264.00	61.60	62.13
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111621\
Data File : BN017465.D
Acq On : 16 Nov 2021 07:31
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Nov 16 15:53:11 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 16 15:52:55 2021
Response via : Initial Calibration



TIC: BN017465.D\data.ms

(77) Benzidine

19.480min (0.000) 0.00 ng

response 3383978

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.82
183.00	13.20	11.62
0.00	0.00	0.00

DDT Breakdown

Instrument :
BNA_N
ClientSampleId :
DFTPP

Date	Instrument Name	DFTPP Data File
11/16/2021	BNA_N	<u>BN017465.D</u>
Compound Name	Response	Retention Time
DDT	1885114	20.72
DDD	30803	20.28
DDE	1426	19.77
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
32229	1917343	1.68