

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112520\
 Data File : BN012742.D
 Acq On : 25 Nov 2020 21:31
 Operator : CG/JU
 Sample : PB133233BL
 Misc : MDL-WATER-BL-03
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK233

Quant Time: Nov 27 11:04:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN112520.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 25 18:19:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.42	152	106829	20.00	ng/ul	0.00
20) Naphthalene-d8	10.19	136	434613	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.08	164	278590	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.84	188	614412	20.00	ng/ul	0.00
79) Chrysene-d12	21.06	240	644371	20.00	ng/ul	0.00
88) Perylene-d12	23.17	264	712005	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	2.96	96	22589	7.61	ng/uL	0.00
4) Pyridine-d5	3.36	84	324852	40.52	ng/ul	0.00
7) Phenol-d5	6.61	99	388262	40.60	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.78	67	247374	41.47	ng/ul	0.00
11) 2-Chlorophenol-d4	6.96	132	314349	41.81	ng/ul	0.00
15) 4-Methylphenol-d8	8.13	113	307448	39.47	ng/ul	0.00
21) Nitrobenzene-d5	8.57	128	152984	43.99	ng/ul	0.00
24) 2-Nitrophenol-d4	9.29	143	177497	46.24	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.82	165	297361	40.78	ng/ul	0.00
31) 4-Chloroaniline-d4	10.33	131	415996	40.78	ng/ul	0.00
46) Dimethylphthalate-d6	13.51	166	1012421	44.69	ng/ul	0.00
49) Acenaphthylene-d8	13.77	160	1103145	41.74	ng/ul	0.00
54) 4-Nitrophenol-d4	14.30	143	155937	38.91	ng/ul	0.00
60) Fluorene-d10	15.09	176	839977	43.92	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.22	200	156578	37.27	ng/ul	0.00
73) Anthracene-d10	16.94	188	1343929	43.77	ng/ul	0.00
81) Pyrene-d10	19.25	212	1412548	43.97	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.03	264	1673537	43.51	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112520\
Data File : BN012742.D
Acq On : 25 Nov 2020 21:31
Operator : CG/JU
Sample : PB133233BL
Misc : MDL-WATER-BL-03
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK233

Quant Time: Nov 27 11:04:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN112520.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 25 18:19:04 2020
Response via : Initial Calibration

