

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN112822\  
 Data File : BN022910.D  
 Acq On : 28 Nov 2022 09:28  
 Operator : CG/JU  
 Sample : PB149071BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SBLK071

Quant Time: Nov 29 01:09:46 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN112622.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Nov 28 03:16:56 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 8.061  | 152  | 10335    | 0.400 | ng/ul | 0.00     |
| 4) Naphthalene-d8           | 10.881 | 136  | 30548    | 0.400 | ng/ul | -0.01    |
| 9) Acenaphthene-d10         | 14.706 | 164  | 17533    | 0.400 | ng/ul | 0.00     |
| 13) Phenanthrene-d10        | 17.450 | 188  | 37227    | 0.400 | ng/ul | 0.00     |
| 17) Chrysene-d12            | 21.634 | 240  | 23948    | 0.400 | ng/ul | 0.00     |
| 23) Perylene-d12            | 24.125 | 264  | 21185    | 0.400 | ng/ul | # 0.00   |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.407  | 96   | 76954    | 6.205 | ng/ul | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.465 | 152  | 15014    | 0.323 | ng/ul | -0.01    |
| 18) Fluoranthene-d10        | 19.472 | 212  | 30566    | 0.345 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN112822\  
Data File : BN022910.D  
Acq On : 28 Nov 2022 09:28  
Operator : CG/JU  
Sample : PB149071BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SBLK071

Quant Time: Nov 29 01:09:46 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN112622.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Nov 28 03:16:56 2022  
Response via : Initial Calibration

