

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112822\
 Data File : BN022922.D
 Acq On : 28 Nov 2022 16:49
 Operator : CG/JU
 Sample : N5591-18
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C9830

Quant Time: Nov 29 01:12:32 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 28 03:16:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.065	152	8374	0.400	ng/ul	0.00
4) Naphthalene-d8	10.887	136	25638	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.705	164	14926	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.449	188	32799	0.400	ng/ul	0.00
17) Chrysene-d12	21.636	240	21589	0.400	ng/ul	0.00
23) Perylene-d12	24.124	264	16332	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.408	96	51041	5.079	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.471	152	11773	0.301	ng/ul	0.00
18) Fluoranthene-d10	19.476	212	25210	0.316	ng/ul	0.00
Target Compounds						
					Qvalue	
15) Phenanthrene	17.492	178	3274	0.029	ng/ul#	92
19) Fluoranthene	19.504	202	4405	0.039	ng/ul#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112822\
Data File : BN022922.D
Acq On : 28 Nov 2022 16:49
Operator : CG/JU
Sample : N5591-18
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C9830

Quant Time: Nov 29 01:12:32 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 28 03:16:56 2022
Response via : Initial Calibration

