

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112922\
 Data File : BN022938.D
 Acq On : 29 Nov 2022 05:11
 Operator : CG/JU
 Sample : N5672-08
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D1-20221116

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 11/29/2022
 Supervised By :mohammad ahmed 12/01/2022

Quant Time: Nov 29 06:04:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 05:14:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.064	152	7166	0.400	ng	0.00
7) Naphthalene-d8	10.882	136	20177	0.400	ng	0.00
13) Acenaphthene-d10	14.698	164	10558	0.400	ng	0.00
19) Phenanthrene-d10	17.451	188	22683m	0.400	ng	0.00
29) Chrysene-d12	21.629	240	12861	0.400	ng	# 0.00
35) Perylene-d12	24.116	264	8288	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.601	112	2375	0.128	ng	0.00
5) Phenol-d6	7.211	99	1906	0.080	ng	0.00
8) Nitrobenzene-d5	9.227	82	4738	0.302	ng	0.00
11) 2-Methylnaphthalene-d10	12.469	152	12698	0.290	ng	0.00
14) 2,4,6-Tribromophenol	16.185	330	1329	0.261	ng	0.00
15) 2-Fluorobiphenyl	13.330	172	16637	0.344	ng	0.00
27) Fluoranthene-d10	19.473	212	21980m	0.357	ng	0.00
31) Terphenyl-d14	20.071	244	14085	0.466	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.441	88	45650	5.019	ng	97
34) Bis(2-ethylhexyl)phtha...	21.540	149	2945	0.141	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112922\
Data File : BN022938.D
Acq On : 29 Nov 2022 05:11
Operator : CG/JU
Sample : N5672-08
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D1-20221116

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 11/29/2022
Supervised By :mohammad ahmed 12/01/2022

Quant Time: Nov 29 06:04:20 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 05:14:21 2022
Response via : Initial Calibration

