

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112922\
 Data File : BN022945.D
 Acq On : 29 Nov 2022 09:29
 Operator : CG/JU
 Sample : PB149298BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB149298BL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 11/30/2022
 Supervised By : mohammad ahmed 12/01/2022

Quant Time: Nov 30 00:21:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 05:14:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.064	152	7149	0.400	ng	0.00
7) Naphthalene-d8	10.882	136	18944	0.400	ng	0.00
13) Acenaphthene-d10	14.698	164	9394	0.400	ng	0.00
19) Phenanthrene-d10	17.451	188	18715	0.400	ng	# 0.00
29) Chrysene-d12	21.634	240	9616	0.400	ng	0.00
35) Perylene-d12	24.114	264	6408	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.601	112	6821	0.367	ng	0.00
5) Phenol-d6	7.211	99	8720	0.369	ng	0.00
8) Nitrobenzene-d5	9.227	82	5074m	0.345	ng	0.00
11) 2-Methylnaphthalene-d10	12.465	152	13806	0.336	ng	0.00
14) 2,4,6-Tribromophenol	16.185	330	1256	0.277	ng	0.00
15) 2-Fluorobiphenyl	13.330	172	16793	0.391	ng	0.00
27) Fluoranthene-d10	19.473	212	15398	0.303	ng	0.00
31) Terphenyl-d14	20.067	244	10303	0.456	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112922\
Data File : BN022945.D
Acq On : 29 Nov 2022 09:29
Operator : CG/JU
Sample : PB149298BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB149298BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 11/30/2022
Supervised By :mohammad ahmed 12/01/2022

Quant Time: Nov 30 00:21:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 05:14:21 2022
Response via : Initial Calibration

