

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112922\
 Data File : BN022958.D
 Acq On : 29 Nov 2022 18:18
 Operator : CG/JU
 Sample : N5671-13
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP04-20221115

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 11/30/2022
 Supervised By :mohammad ahmed 12/01/2022

Quant Time: Nov 30 08:21:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 30 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.064	152	7058	0.400	ng	0.00
7) Naphthalene-d8	10.882	136	18880	0.400	ng	0.00
13) Acenaphthene-d10	14.698	164	8950	0.400	ng	0.00
19) Phenanthrene-d10	17.452	188	17459	0.400	ng	# 0.01
29) Chrysene-d12	21.625	240	9390	0.400	ng	# 0.00
35) Perylene-d12	24.118	264	6589	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.601	112	1596	0.087	ng	0.00
5) Phenol-d6	7.212	99	1324	0.057	ng	0.00
8) Nitrobenzene-d5	9.228	82	4299m	0.293	ng	0.00
11) 2-Methylnaphthalene-d10	12.465	152	7634	0.186	ng	0.00
14) 2,4,6-Tribromophenol	16.186	330	10	0.002	ng	0.00
15) 2-Fluorobiphenyl	13.330	172	12944	0.316	ng	0.00
27) Fluoranthene-d10	19.473	212	12215	0.258	ng	0.00
31) Terphenyl-d14	20.067	244	9058	0.411	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.442	88	46074	5.143	ng	98
34) Bis(2-ethylhexyl)phtha...	21.536	149	467	0.031	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112922\
Data File : BN022958.D
Acq On : 29 Nov 2022 18:18
Operator : CG/JU
Sample : N5671-13
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP04-20221115

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 11/30/2022
Supervised By :mohammad ahmed 12/01/2022

Quant Time: Nov 30 08:21:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 30 00:45:21 2022
Response via : Initial Calibration

