

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120320\
 Data File : BN012844.D
 Acq On : 03 Dec 2020 13:02
 Operator : CG/JU
 Sample : SST0.402
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST0.402

Quant Time: Dec 03 13:33:44 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN120320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 03 13:32:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.41	152	501	0.40	ng/ul	0.00
2) Naphthalene-d8	10.17	136	1850	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.07	164	1068	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.82	188	2337	0.40	ng/ul	0.00
16) Chrysene-d12	21.04	240	2011	0.40	ng/ul	0.00
20) Perylene-d12	23.15	264	2005	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.78	152	1096	0.41	ng/ul	0.00
14) Fluoranthene-d10	18.87	212	2526	0.41	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.22	128	2086	0.375	ng/ul	100
5) 2-Methylnaphthalene	11.85	142	1360	0.382	ng/ul	100
7) Acenaphthylene	13.78	152	2003	0.385	ng/ul	100
8) Acenaphthene	14.13	153	1641	0.374	ng/ul	100
9) Fluorene	15.13	166	1879	0.389	ng/ul	100
11) Pentachlorophenol	16.49	266	247	0.333	ng/ul	100
12) Phenanthrene	16.87	178	3091	0.380	ng/ul	100
13) Anthracene	16.96	178	2643	0.374	ng/ul	100
15) Fluoranthene	18.89	202	3635	0.387	ng/ul	100
17) Pyrene	19.26	202	3766	0.419	ng/ul	100
18) Benzo(a)anthracene	21.03	228	3296	0.457	ng/ul	100
19) Chrysene	21.08	228	3816	0.424	ng/ul	100
21) Benzo(b)fluoranthene	22.52	252	3676	0.399	ng/ul	100
22) Benzo(k)fluoranthene	22.57	252	3988	0.375	ng/ul	100
23) Benzo(a)pyrene	23.06	252	3294	0.379	ng/ul	100
24) Indeno(1,2,3-cd)pyrene	25.21	276	4453	0.392	ng/ul#	100
25) Dibenzo(a,h)anthracene	25.23	278	3548	0.389	ng/ul	100
26) Benzo(g,h,i)perylene	25.84	276	4066	0.398	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120320\
Data File : BN012844.D
Acq On : 03 Dec 2020 13:02
Operator : CG/JU
Sample : SST0.402
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleID :
SSTD0.402

Quant Time: Dec 03 13:33:44 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN120320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 03 13:32:50 2020
Response via : Initial Calibration

