

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120320\
 Data File : BN012847.D
 Acq On : 03 Dec 2020 14:50
 Operator : CG/JU
 Sample : SSTD3.205
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.205

Quant Time: Dec 03 15:25:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN120320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 03 13:32:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.41	152	490	0.40	ng/ul	0.00
2) Naphthalene-d8	10.16	136	1897	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.06	164	1129	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.82	188	2417	0.40	ng/ul	0.00
16) Chrysene-d12	21.04	240	2159	0.40	ng/ul	0.00
20) Perylene-d12	23.15	264	2043	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.48	266	2552	3.325	ng/ul	Ovalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120320\
Data File : BN012847.D
Acq On : 03 Dec 2020 14:50
Operator : CG/JU
Sample : SST3.205
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.205

Quant Time: Dec 03 15:25:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN120320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 03 13:32:50 2020
Response via : Initial Calibration

