


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120320\
 Data File : BN012841.D
 Acq On : 03 Dec 2020 11:14
 Operator : CG/JU
 Sample : DFTPP24
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP24

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN120320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Dec 04 10:02:32 2020



AutoFind: Scans 2515, 2516, 2517; Background Corrected with Scan 2508

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.9	79725	PASS
68	69	0.00	2	2.0	1479	PASS
69	198	0.00	100	47.4	74224	PASS
70	69	0.00	2	0.7	514	PASS
127	198	10	80	47.9	75098	PASS
197	198	0.00	2	0.6	970	PASS
198	198	100	100	100.0	156733	PASS
199	198	5	9	6.4	10010	PASS
275	198	10	60	29.1	45546	PASS
365	198	1	100	4.5	7036	PASS
441	198	0.01	100	13.3	20922	PASS
442	198	50	100	92.2	144525	PASS
443	442	15	24	19.4	28085	PASS

Scan 2508 (17.239 min): BN012841.D
DFTPP24

m/z	Abundance
40.10	399.0
44.00	1219.0
72.90	357.0
83.80	374.0
206.90	595.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

m/z	Abundance
37.10	1014.0
38.10	1206.0
39.10	7895.0
40.00	771.0
41.10	311.0
44.00	1635.0
45.10	311.0
49.00	1259.0
50.00	32408.0
51.00	108904.0
52.00	5563.0
55.00	766.0
56.00	3124.0
57.00	6685.0
61.00	1686.0
62.00	1604.0
63.00	4759.0
64.10	436.0
64.90	1925.0
68.10	2168.0
69.00	100752.0
70.00	826.0
73.10	1461.0
74.00	13223.0
75.00	19256.0
76.00	5582.0
77.10	110440.0
78.10	7948.0
79.00	9871.0
80.00	6501.0
81.00	8604.0
81.90	2410.0
83.00	2046.0
84.90	1452.0
86.00	2186.0
86.90	1279.0
87.70	307.0
90.90	1856.0
92.00	2463.0
93.00	17824.0
93.90	884.0
95.80	459.0
98.00	14110.0
99.00	10638.0
100.00	830.0
101.00	4327.0
103.00	1509.0
104.00	3538.0
105.00	2774.0
106.10	1264.0
107.00	32800.0
108.00	5164.0
108.90	804.0
110.00	53544.0
111.00	8267.0
111.90	1199.0
112.90	467.0
116.00	2151.0
117.00	43272.0
118.00	3504.0
118.90	363.0
120.00	473.0
122.00	2649.0
123.00	4685.0
123.90	1863.0
124.90	1471.0
126.20	608.0
127.00	98520.0
128.00	8296.0
129.00	44080.0
130.00	3678.0
131.00	794.0
131.90	790.0
132.80	331.0
134.00	1287.0
135.00	4297.0
135.90	1607.0
137.00	2364.0
138.10	502.0
139.00	404.0
140.00	605.0
141.00	6378.0
142.00	1699.0
143.00	1342.0
144.00	456.0
145.10	315.0
145.90	1250.0
147.00	3422.0
148.00	9975.0
149.00	2276.0
149.90	562.0
150.90	744.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

151.90	481.0
153.00	1873.0
153.90	1754.0
155.00	3584.0
156.00	4606.0
157.00	786.0
157.90	1444.0
158.90	1288.0
159.90	1517.0
160.90	2751.0
162.10	1058.0
163.00	313.0
163.90	353.0
165.00	2078.0
165.90	1835.0
167.00	14254.0
168.00	9027.0
169.00	1115.0
170.00	502.0
171.00	506.0
172.00	879.0
172.90	1558.0
173.90	2535.0
175.00	4024.0
175.90	726.0
177.00	2181.0
178.10	638.0
179.00	8809.0
180.00	6471.0
181.00	2623.0
182.00	503.0
184.00	930.0
185.10	3204.0
186.00	28768.0
187.00	9053.0
188.00	1055.0
188.90	2312.0
190.10	451.0
190.90	1431.0
192.00	3674.0
193.00	2900.0
194.00	529.0
195.00	424.0
196.00	5618.0
197.10	1425.0
197.90	189824.0
199.00	13061.0
200.00	1241.0
201.40	679.0
203.00	2017.0
204.10	7967.0
205.00	14232.0
206.00	52800.0
207.00	9173.0
208.00	2502.0
209.00	1330.0
210.20	998.0
211.00	2127.0
212.10	416.0
215.00	819.0
216.00	1106.0
217.00	15332.0
217.90	2106.0
220.10	330.0
221.00	6558.0
222.00	1637.0
223.00	3490.0
224.10	29432.0
225.00	7908.0
226.10	1071.0
227.00	15695.0
228.00	2415.0
229.00	3389.0
229.90	449.0
231.00	1268.0
234.00	1318.0
235.00	1385.0
235.90	738.0
237.00	1757.0
238.90	749.0
239.60	389.0
241.00	868.0
242.00	1317.0
243.00	1995.0
244.00	22384.0
245.00	2765.0
246.00	5842.0
247.00	1158.0
249.10	615.0
251.90	302.0
252.90	669.0
255.00	116672.0
256.00	16912.0
257.00	1660.0
258.00	7902.0
258.90	1323.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

265.00	2732.0
266.00	456.0
273.00	3201.0
274.00	10369.0
275.00	53624.0
276.10	7942.0
277.00	5133.0
278.10	1221.0
279.00	395.0
280.90	322.0
283.00	802.0
284.00	604.0
285.00	840.0
293.10	1227.0
294.10	377.0
295.10	697.0
296.00	17520.0
297.00	2609.0
301.10	319.0
301.90	302.0
303.00	1677.0
304.00	457.0
313.00	421.0
314.00	745.0
315.00	1992.0
316.00	809.0
320.90	436.0
321.90	325.0
323.10	3736.0
324.10	800.0
326.90	631.0
327.90	312.0
332.00	479.0
332.90	511.0
334.00	3691.0
335.10	1352.0
340.90	459.0
345.90	1365.0
352.00	980.0
352.80	638.0
354.10	1330.0
365.00	7506.0
365.90	968.0
372.00	1618.0
373.00	461.0
382.90	680.0
401.90	1050.0
403.00	1338.0
420.90	1045.0
422.10	1052.0
423.00	7019.0
423.90	2049.0
441.00	20288.0
442.00	116424.0
443.10	28464.0
443.90	3099.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

m/z	Abundance
37.00	605.0
38.10	1857.0
39.10	7334.0
40.00	735.0
41.10	441.0
44.00	1643.0
49.00	1232.0
50.00	26760.0
51.00	91472.0
52.00	4769.0
55.00	575.0
56.00	2919.0
57.00	4689.0
58.00	346.0
61.00	1230.0
62.00	1096.0
63.00	3059.0
64.10	772.0
65.00	1262.0
68.00	1625.0
69.00	84872.0
69.90	387.0
73.10	1005.0
74.00	12058.0
75.00	14091.0
76.10	4748.0
77.10	91984.0
78.00	5240.0
79.00	7964.0
80.00	5708.0
81.00	7769.0
82.00	1387.0
82.90	1480.0
83.90	413.0
85.10	1062.0
85.90	2366.0
87.00	1212.0
87.90	397.0
91.00	2048.0
92.00	1771.0
93.00	13396.0
94.00	680.0
96.10	402.0
98.00	12945.0
99.00	8408.0
100.10	686.0
101.00	3903.0
101.90	513.0
103.00	1258.0
104.00	2890.0
105.00	2361.0
106.10	770.0
107.00	31128.0
108.00	3988.0
108.90	1201.0
109.20	1160.0
110.00	48104.0
111.00	7525.0
112.00	993.0
115.90	2093.0
117.00	38376.0
118.00	3022.0
120.00	481.0
122.00	2544.0
123.00	3503.0
124.00	1891.0
124.90	1303.0
127.00	87912.0
128.00	7097.0
129.00	40520.0
130.00	3646.0
130.90	696.0
134.00	2208.0
135.10	2696.0
136.00	1560.0
137.00	2015.0
138.90	330.0
140.00	709.0
141.00	6151.0
142.00	2002.0
142.90	896.0
146.00	1238.0
147.00	3087.0
148.00	8903.0
149.00	2103.0
149.80	404.0
151.00	980.0
152.00	319.0
153.10	2321.0
154.00	1344.0
155.00	3130.0
156.00	4143.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

157.00	778.0
158.00	1435.0
159.00	1026.0
160.00	2011.0
161.00	2352.0
162.10	768.0
165.10	2073.0
166.00	1699.0
167.00	13381.0
168.00	7391.0
169.00	762.0
170.00	336.0
171.00	617.0
171.90	735.0
173.00	1289.0
174.10	2020.0
175.00	4899.0
176.00	1125.0
177.00	2617.0
177.90	678.0
179.00	7781.0
180.00	5735.0
181.00	2552.0
182.00	518.0
183.90	654.0
185.00	4609.0
186.00	29720.0
187.00	8529.0
188.00	1166.0
189.00	2717.0
191.00	1078.0
192.00	2766.0
193.00	3595.0
193.90	746.0
196.00	4810.0
197.20	1485.0
198.00	191232.0
199.00	11523.0
200.00	1260.0
201.00	442.0
201.60	584.0
203.00	1760.0
204.00	7315.0
205.00	13764.0
206.00	50304.0
207.00	7234.0
208.00	2642.0
209.00	793.0
210.20	775.0
210.90	2471.0
215.10	872.0
216.10	1154.0
217.00	16432.0
218.00	1693.0
219.00	418.0
221.00	6704.0
221.90	1743.0
222.90	2925.0
224.00	27344.0
225.10	8001.0
225.90	769.0
227.00	15117.0
227.90	2741.0
229.00	3125.0
230.10	357.0
230.90	931.0
231.90	327.0
234.00	812.0
234.90	1542.0
236.00	887.0
237.00	1329.0
238.90	741.0
240.00	344.0
241.00	881.0
241.90	1596.0
243.10	2266.0
244.10	23344.0
245.00	2643.0
246.00	5405.0
247.00	1409.0
249.00	827.0
251.00	336.0
252.90	488.0
255.00	122480.0
256.00	17328.0
257.10	1914.0
258.00	8786.0
258.90	1373.0
261.00	340.0
265.00	3502.0
266.00	594.0
272.10	378.0
273.00	3726.0
274.00	11019.0
275.00	54224.0
276.00	6335.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

277.00	5984.0
278.00	1101.0
281.00	510.0
283.00	1222.0
284.00	581.0
284.90	1205.0
292.00	371.0
292.90	1097.0
296.00	19640.0
297.00	2833.0
302.10	367.0
303.00	2795.0
304.00	847.0
307.90	391.0
313.90	552.0
314.30	878.0
315.00	2049.0
316.10	670.0
320.90	621.0
323.00	5416.0
324.00	814.0
327.10	1020.0
328.10	863.0
331.80	546.0
333.00	479.0
334.00	3389.0
335.00	649.0
340.90	707.0
345.90	1718.0
352.00	1775.0
353.00	1011.0
354.00	1954.0
354.90	408.0
364.90	8100.0
365.90	1428.0
372.10	3577.0
373.00	1064.0
383.00	709.0
390.20	539.0
401.90	1129.0
403.00	1515.0
403.80	535.0
420.90	1298.0
421.90	1328.0
423.00	9627.0
424.00	3275.0
441.00	25592.0
442.00	198400.0
443.00	32960.0
444.00	2949.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

m/z	Abundance
38.00	1044.0
39.10	2430.0
39.90	739.0
44.00	1242.0
49.10	487.0
50.10	12279.0
51.10	38800.0
52.00	2014.0
54.90	345.0
56.00	1419.0
57.10	2632.0
60.90	576.0
62.00	708.0
63.00	1377.0
64.90	573.0
68.00	645.0
69.00	37048.0
69.90	331.0
73.10	965.0
74.00	5066.0
75.00	6652.0
76.00	2606.0
77.10	41864.0
78.00	3105.0
79.00	3137.0
80.00	2768.0
81.00	3259.0
81.90	832.0
83.00	713.0
85.10	641.0
86.00	1507.0
87.00	455.0
91.00	609.0
92.10	968.0
93.00	6356.0
94.10	483.0
98.00	5986.0
99.00	3724.0
100.90	1658.0
103.00	744.0
104.00	1441.0
105.00	1390.0
107.00	13933.0
108.00	1343.0
109.00	533.0
110.00	20112.0
110.90	3458.0
112.00	594.0
116.10	919.0
117.00	18424.0
117.90	1603.0
122.10	793.0
123.00	1687.0
123.90	632.0
125.10	544.0
127.00	38864.0
128.00	3013.0
129.00	18136.0
130.00	1384.0
134.00	533.0
135.10	1385.0
136.00	719.0
137.10	761.0
139.90	343.0
141.00	2427.0
141.90	968.0
142.70	436.0
146.10	631.0
147.00	1502.0
148.00	3998.0
149.00	614.0
151.00	480.0
152.00	343.0
153.00	848.0
153.80	505.0
155.00	1627.0
156.00	2197.0
158.00	645.0
158.90	683.0
159.90	1115.0
161.00	1322.0
161.80	778.0
165.00	1201.0
166.00	865.0
167.10	6237.0
168.00	3937.0
169.10	466.0
171.00	318.0
171.90	778.0
172.90	539.0
174.00	607.0
175.00	1960.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

176.10	655.0
177.00	1465.0
179.00	3902.0
180.00	2747.0
181.00	1294.0
183.90	791.0
185.00	2411.0
186.00	14358.0
187.00	3637.0
189.00	884.0
190.00	492.0
190.90	457.0
191.90	1572.0
192.90	1325.0
194.10	522.0
196.00	2650.0
198.00	89144.0
199.00	5448.0
200.00	580.0
201.30	323.0
201.60	319.0
202.90	1123.0
204.00	4425.0
205.10	6342.0
206.00	25880.0
207.10	4380.0
208.00	1423.0
209.90	854.0
211.00	1290.0
216.10	625.0
217.00	9174.0
218.00	1065.0
221.10	3569.0
221.90	1390.0
223.10	1714.0
224.10	14502.0
225.10	3463.0
226.10	322.0
227.00	8106.0
227.90	1104.0
228.90	1057.0
231.10	394.0
234.10	589.0
234.80	425.0
235.10	356.0
236.00	418.0
237.10	381.0
238.90	325.0
240.00	323.0
240.70	322.0
241.20	361.0
242.00	1117.0
242.90	691.0
244.00	12479.0
245.00	1968.0
246.00	2422.0
247.00	394.0
252.90	454.0
255.00	68064.0
256.00	9322.0
256.90	676.0
258.00	4875.0
258.90	565.0
265.00	2065.0
271.80	422.0
272.00	422.0
273.10	1364.0
274.10	5345.0
275.00	28792.0
276.00	3766.0
277.00	3113.0
277.90	785.0
281.00	644.0
283.00	399.0
284.90	407.0
292.90	718.0
296.00	11978.0
297.00	1170.0
303.00	1055.0
313.90	664.0
315.00	726.0
315.80	527.0
323.00	2842.0
323.90	677.0
326.90	531.0
327.90	383.0
333.00	680.0
334.00	1909.0
335.00	674.0
340.90	590.0
345.90	799.0
351.90	989.0
352.90	767.0
354.00	944.0
355.00	442.0
365.00	5503.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

365.90	684.0
371.00	303.0
372.00	1840.0
373.10	779.0
383.00	429.0
389.90	318.0
402.00	827.0
403.00	1120.0
403.80	422.0
420.90	900.0
422.00	1012.0
423.00	5686.0
424.00	1787.0
441.00	16888.0
442.00	118752.0
443.00	22832.0
444.00	2033.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24