

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120522\
 Data File : BN023065.D
 Acq On : 05 Dec 2022 19:42
 Operator : CG/JU
 Sample : PB149410BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB149410BL

Quant Time: Dec 06 04:28:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 06 04:02:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.049	152	4483	0.400	ng	0.00
7) Naphthalene-d8	10.872	136	12318	0.400	ng	# 0.01
13) Acenaphthene-d10	14.687	164	6262	0.400	ng	0.00
19) Phenanthrene-d10	17.439	188	15421	0.400	ng	# 0.01
29) Chrysene-d12	21.620	240	11860	0.400	ng	0.00
35) Perylene-d12	24.098	264	9975	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.586	112	4176	0.417	ng	0.00
5) Phenol-d6	7.197	99	5390	0.405	ng	0.00
8) Nitrobenzene-d5	9.206	82	3205	0.345	ng	0.00
11) 2-Methylnaphthalene-d10	12.450	152	7396	0.347	ng	0.00
14) 2,4,6-Tribromophenol	16.173	330	963	0.371	ng	0.00
15) 2-Fluorobiphenyl	13.319	172	10933	0.390	ng	0.00
27) Fluoranthene-d10	19.464	212	14578	0.331	ng	0.00
31) Terphenyl-d14	20.053	244	9335	0.385	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120522\
Data File : BN023065.D
Acq On : 05 Dec 2022 19:42
Operator : CG/JU
Sample : PB149410BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB149410BL

Quant Time: Dec 06 04:28:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 06 04:02:46 2022
Response via : Initial Calibration

