



Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN120522\  
Data File : BN023055.D  
Acq On : 05 Dec 2022 12:06  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN120522.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Tue Dec 06 04:02:46 2022






AutoFind: Scans 2513, 2514, 2515; Background Corrected with Scan 2506

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 42.3         | 109034     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.9          | 2179       | PASS                |
| 69             | 198             | 0.00            | 100             | 45.2         | 116645     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 603        | PASS                |
| 127            | 198             | 10              | 80              | 52.2         | 134669     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.4          | 933        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 257792     | PASS                |
| 199            | 198             | 5               | 9               | 6.7          | 17318      | PASS                |
| 275            | 198             | 10              | 60              | 28.5         | 73501      | PASS                |
| 365            | 198             | 1               | 100             | 4.4          | 11442      | PASS                |
| 441            | 198             | 0.01            | 100             | 16.7         | 43139      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 302848     | PASS                |
| 443            | 442             | 15              | 24              | 19.4         | 58667      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN120522\  
 Data File : BN023055.D  
 Acq On : 05 Dec 2022 12:06  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Quant Time: Dec 06 06:15:31 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Dec 06 06:15:26 2022  
 Response via : Initial Calibration



TIC: BN023055.D\data.ms

**(70) Pentachlorophenol (C)**

17.080min ( 0.000) 14226.77 ng




response 212578

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 67.18  |
| 264.00 | 61.60  | 63.32  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN120522\  
Data File : BN023055.D  
Acq On : 05 Dec 2022 12:06  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Dec 06 06:15:31 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Dec 06 06:15:26 2022  
Response via : Initial Calibration



TIC: BN023055.D\data.ms

(77) Benzidine

19.680min ( 0.000) 0.00 ng

response 1459800

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 13.22  |
| 183.00 | 13.20  | 11.21  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 12/5/2022     | BNA_N            | <u>BN023055.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 642197           | 20.91              |
| DDD           | 6318             | 20.521             |
| DDE           | 65               | 19.963             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 6383          | 648580           | 0.98               |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP