

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120822\
 Data File : BN023098.D
 Acq On : 08 Dec 2022 17:03
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Dec 09 07:28:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 09 07:25:53 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.020	152	8090	0.400	ng	0.00
7) Naphthalene-d8	10.840	136	23133	0.400	ng	0.00
13) Acenaphthene-d10	14.666	164	13470	0.400	ng	0.00
19) Phenanthrene-d10	17.414	188	28191	0.400	ng	0.00
29) Chrysene-d12	21.589	240	26247	0.400	ng	# 0.00
35) Perylene-d12	24.056	264	17877	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.558	112	49643	2.653	ng	0.00
5) Phenol-d6	7.175	99	65990	2.805	ng	0.00
8) Nitrobenzene-d5	9.185	82	52766	3.040	ng	0.00
11) 2-Methylnaphthalene-d10	12.427	152	138357	3.168	ng	0.00
14) 2,4,6-Tribromophenol	16.148	330	18487	3.249	ng	0.00
15) 2-Fluorobiphenyl	13.298	172	173988	2.910	ng	0.00
27) Fluoranthene-d10	19.439	212	234419	3.037	ng	0.00
31) Terphenyl-d14	20.031	244	139156	2.793	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.398	88	25353	2.527	ng	98
3) n-Nitrosodimethylamine	3.723	42	28858	2.941	ng	# 96
6) bis(2-Chloroethyl)ether	7.435	93	69322	2.669	ng	98
9) Naphthalene	10.893	128	195994	2.856	ng	99
10) Hexachlorobutadiene	11.181	225	36427	2.827	ng	# 100
12) 2-Methylnaphthalene	12.117	142	40229	3.831	ng	# 91
16) Acenaphthylene	14.388	152	201290	3.180	ng	100
17) Acenaphthene	14.730	154	136104	2.951	ng	97
18) Fluorene	15.714	166	154549	2.988	ng	99
20) 4,6-Dinitro-2-methylph...	15.776	198	16007	4.197	ng	# 81
21) 4-Bromophenyl-phenylether	16.595	248	52085	3.067	ng	98
22) Hexachlorobenzene	16.719	284	66183	3.001	ng	100
23) Atrazine	16.868	200	40499	3.278	ng	98
24) Pentachlorophenol	17.054	266	26196	4.044	ng	99
25) Phenanthrene	17.452	178	287541	3.026	ng	100
26) Anthracene	17.538	178	247626	3.234	ng	100
28) Fluoranthene	19.469	202	324068	3.126	ng	100
30) Pyrene	19.831	202	320665	2.930	ng	100
32) Benzo(a)anthracene	21.580	228	298657	3.082	ng	99
33) Chrysene	21.634	228	307277	2.838	ng	99
34) Bis(2-ethylhexyl)phtha...	21.500	149	124141	2.965	ng	99
36) Indeno(1,2,3-cd)pyrene	26.626	276	293290	3.087	ng	100
37) Benzo(b)fluoranthene	23.302	252	268022	3.280	ng	# 94
38) Benzo(k)fluoranthene	23.351	252	279632	3.339	ng	# 94
39) Benzo(a)pyrene	23.945	252	206454	3.118	ng	# 90
40) Dibenzo(a,h)anthracene	26.649	278	236723	3.149	ng	97
41) Benzo(g,h,i)perylene	27.421	276	245390	3.189	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120822\
 Data File : BN023098.D
 Acq On : 08 Dec 2022 17:03
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Dec 09 07:28:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 09 07:25:53 2022
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 8.020 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN023098.D
 Acq: 08 Dec 2022 17:03




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:152 Resp: 8090
 Ion Ratio Lower Upper
 152 100
 150 158.7 125.6 188.4
 115 62.1 49.0 73.4



#2
 1,4-Dioxane
 Concen: 2.527 ng
 RT: 3.398 min Scan# 44
 Delta R.T. 0.000 min
 Lab File: BN023098.D
 Acq: 08 Dec 2022 17:03

Tgt Ion: 88 Resp: 25353
 Ion Ratio Lower Upper
 88 100
 43 28.7 23.3 34.9
 58 70.4 58.0 87.0









#3
n-Nitrosodimethylamine
Concen: 2.941 ng
RT: 3.723 min Scan# 89
Delta R.T. -0.014 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 28858
Ion Ratio Lower Upper
42 100
74 123.3 95.8 143.6
44 7.1 8.4 12.6#



#4
2-Fluorophenol
Concen: 2.653 ng
RT: 5.558 min Scan# 343
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion: 112 Resp: 49643
Ion Ratio Lower Upper
112 100
64 55.8 44.4 66.6
63 29.8 23.7 35.5









#5
Phenol-d6
Concen: 2.805 ng
RT: 7.175 min Scan# 567
Delta R.T. 0.007 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 65990
Ion Ratio Lower Upper
99 100
42 20.8 16.3 24.5
71 32.7 26.5 39.7



#6
bis(2-Chloroethyl)ether
Concen: 2.669 ng
RT: 7.435 min Scan# 603
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion: 93 Resp: 69322
Ion Ratio Lower Upper
93 100
63 74.9 58.1 87.1
95 32.3 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.840 min Scan# 961
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	23133
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.0	13.4
54	7.9	6.5	9.7
68	6.5	5.4	8.2



#8
Nitrobenzene-d5
Concen: 3.040 ng
RT: 9.185 min Scan# 806
Delta R.T. -0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:	82	Resp:	52766
Ion Ratio	Lower	Upper	
82	100		
128	39.1	31.4	47.2
54	48.7	41.0	61.4









#9
Naphthalene
Concen: 2.856 ng
RT: 10.893 min Scan# 966
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 195994
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 12.7 10.5 15.7



#10
Hexachlorobutadiene
Concen: 2.827 ng
RT: 11.181 min Scan# 993
Delta R.T. -0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:225 Resp: 36427
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.1 76.7









#11
2-Methylnaphthalene-d10
Concen: 3.168 ng
RT: 12.427 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 138357
Ion Ratio Lower Upper
152 100
151 19.7 15.1 22.7



#12
2-Methylnaphthalene
Concen: 3.831 ng
RT: 12.117 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:142 Resp: 40229
Ion Ratio Lower Upper
142 100
141 10.8 10.9 16.3#
115 2.2 5.7 8.5#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.666 min Scan# 1489
Delta R.T. -0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 13470
Ion Ratio Lower Upper
164 100
162 103.7 83.4 125.0
160 52.0 41.8 62.8



#14
2,4,6-Tribromophenol
Concen: 3.249 ng
RT: 16.148 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:330 Resp: 18487
Ion Ratio Lower Upper
330 100
332 96.4 77.3 115.9
141 41.8 33.5 50.3









#15
2-Fluorobiphenyl
Concen: 2.910 ng
RT: 13.298 min Scan# 1361
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 173988
Ion Ratio Lower Upper
172 100
171 33.5 27.4 41.0
170 22.5 17.9 26.9



#16
Acenaphthylene
Concen: 3.180 ng
RT: 14.388 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:152 Resp: 201290
Ion Ratio Lower Upper
152 100
151 19.4 15.4 23.2
153 13.0 10.3 15.5









#17
Acenaphthene
Concen: 2.951 ng
RT: 14.730 min Scan# 1495
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 136104
Ion Ratio Lower Upper
154 100
153 108.9 88.6 132.8
152 56.7 48.1 72.1



#18
Fluorene
Concen: 2.988 ng
RT: 15.714 min Scan# 1587
Delta R.T. 0.011 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:166 Resp: 154549
Ion Ratio Lower Upper
166 100
165 98.3 79.8 119.6
167 12.9 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.414 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 28191
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.7 6.1 9.1



#20
4,6-Dinitro-2-methylphenol
Concen: 4.197 ng
RT: 15.776 min Scan# 1592
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:198 Resp: 16007
Ion Ratio Lower Upper
198 100
51 49.8 57.0 85.4#
105 51.4 47.2 70.8









#21
4-Bromophenyl-phenylether
Concen: 3.067 ng
RT: 16.595 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 52085
Ion Ratio Lower Upper
248 100
250 92.1 74.3 111.5
141 83.7 65.0 97.6



#22
Hexachlorobenzene
Concen: 3.001 ng
RT: 16.719 min Scan# 1668
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:284 Resp: 66183
Ion Ratio Lower Upper
284 100
142 39.1 31.0 46.4
249 30.7 24.4 36.6









#23
Atrazine
Concen: 3.278 ng
RT: 16.868 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 40499
Ion Ratio Lower Upper
200 100
173 21.6 18.2 27.4
215 46.4 38.0 57.0



#24
Pentachlorophenol
Concen: 4.044 ng
RT: 17.054 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:266 Resp: 26196
Ion Ratio Lower Upper
266 100
264 62.6 50.1 75.1
268 63.9 49.7 74.5





#25

Phenanthrene

Concen: 3.026 ng

RT: 17.452 min Scan# 1727

Delta R.T. 0.000 min

Lab File: BN023098.D

Acq: 08 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:178 Resp: 287541

Ion Ratio Lower Upper

178 100

176 19.3 15.4 23.2

179 15.2 12.2 18.2



#26

Anthracene

Concen: 3.234 ng

RT: 17.538 min Scan# 1734

Delta R.T. 0.000 min

Lab File: BN023098.D

Acq: 08 Dec 2022 17:03




Tgt Ion:178 Resp: 247626


Ion Ratio Lower Upper

178 100

176 18.7 15.1 22.7

179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.037 ng
RT: 19.439 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 234419
Ion Ratio Lower Upper
212 100
106 16.2 13.0 19.4
104 9.2 7.5 11.3



#28
Fluoranthene
Concen: 3.126 ng
RT: 19.469 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:202 Resp: 324068
Ion Ratio Lower Upper
202 100
101 12.3 9.7 14.5
203 17.4 13.8 20.6









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.589 min Scan# 2423
Delta R.T. -0.009 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 26247
Ion Ratio Lower Upper
240 100
120 18.7 10.1 15.1#
236 29.3 22.2 33.4



#30
Pyrene
Concen: 2.930 ng
RT: 19.831 min Scan# 2204
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:202 Resp: 320665
Ion Ratio Lower Upper
202 100
200 21.1 16.9 25.3
203 17.8 14.2 21.4









#31
 Terphenyl-d14
 Concen: 2.793 ng
 RT: 20.031 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN023098.D
 Acq: 08 Dec 2022 17:03

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:244 Resp: 139156




Ion	Ratio	Lower	Upper
244	100		
212	8.4	7.6	11.4
122	12.8	12.6	18.8




#32
 Benzo(a)anthracene
 Concen: 3.082 ng
 RT: 21.580 min Scan# 2421
 Delta R.T. 0.000 min
 Lab File: BN023098.D
 Acq: 08 Dec 2022 17:03

Tgt Ion:228 Resp: 298657

Ion	Ratio	Lower	Upper
228	100		
226	26.8	22.0	33.0
229	19.8	15.8	23.8









#33
Chrysene
Concen: 2.838 ng
RT: 21.634 min Scan# 2427
Delta R.T. -0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 307277
Ion Ratio Lower Upper
228 100
226 29.7 24.4 36.6
229 19.7 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 2.965 ng
RT: 21.500 min Scan# 2412
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:149 Resp: 124141
Ion Ratio Lower Upper
149 100
167 25.7 20.2 30.2
279 2.5 2.3 3.5





#35

Perylene-d12

Concen: 0.400 ng

RT: 24.056 min Scan# 3079

Delta R.T. -0.003 min

Lab File: BN023098.D

Acq: 08 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 17877

Ion Ratio Lower Upper

264 100

260 26.7 21.7 32.5

265 50.1 43.2 64.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.087 ng

RT: 26.626 min Scan# 3957

Delta R.T. -0.006 min

Lab File: BN023098.D

Acq: 08 Dec 2022 17:03




Tgt Ion:276 Resp: 293290


Ion Ratio Lower Upper

276 100

138 30.9 25.0 37.6

277 24.7 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 3.280 ng
RT: 23.302 min Scan# 2820
Delta R.T. -0.003 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 268022
Ion Ratio Lower Upper
252 100
253 21.6 19.0 28.4
125 12.1 12.8 19.2#



#38
Benzo(k)fluoranthene
Concen: 3.339 ng
RT: 23.351 min Scan# 2837
Delta R.T. 0.000 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03


Tgt Ion:252 Resp: 279632
Ion Ratio Lower Upper
252 100
253 21.6 19.1 28.7
125 11.9 12.5 18.7#








#39
Benzo(a)pyrene
Concen: 3.118 ng
RT: 23.945 min Scan# 3040
Delta R.T. -0.003 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 206454
Ion Ratio Lower Upper
252 100
253 22.0 20.6 30.8
125 13.5 15.8 23.8#



#40
Dibenzo(a,h)anthracene
Concen: 3.149 ng
RT: 26.649 min Scan# 3965
Delta R.T. -0.003 min
Lab File: BN023098.D
Acq: 08 Dec 2022 17:03

Tgt Ion:278 Resp: 236723
Ion Ratio Lower Upper
278 100
139 20.9 17.5 26.3
279 23.2 20.5 30.7





#41
 Benzo(g,h,i)perylene
 Concen: 3.189 ng
 RT: 27.421 min Scan# 41
 Delta R.T. -0.003 min
 Lab File: BN023098.D
 Acq: 08 Dec 2022 17:03

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:276 Resp: 245390
 Ion Ratio Lower Upper
 276 100
 277 23.7 19.9 29.9
 138 26.5 22.2 33.2

