

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120921\
 Data File : BN017852.D
 Acq On : 10 Dec 2021 02:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 10 06:50:49 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 08 01:18:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.147	0.133	9.5	85	-0.04
3	n-Nitrosodimethylamine	0.413	0.375	9.2	84	0.00
4 S	2-Fluorophenol	0.966	0.821	15.0	79	0.00
5 S	Phenol-d6	0.914	0.787	13.9	85	0.00
6	bis(2-Chloroethyl)ether	1.035	1.028	0.7	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.266	0.230	13.5	95	0.00
9	Naphthalene	1.047	1.014	3.2	100	0.00
10	Hexachlorobutadiene	0.304	0.296	2.6	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.798	0.728	8.8	89	0.00
12	2-Methylnaphthalene	0.755	0.696	7.8	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	-0.01
14 S	2,4,6-Tribromophenol	0.209	0.092	56.0#	43#	0.00
15 S	2-Fluorobiphenyl	1.465	1.451	1.0	85	0.00
16	Acenaphthylene	1.571	1.443	8.1	83	0.00
17	Acenaphthene	1.047	1.073	-2.5	88	0.00
18	Fluorene	1.337	1.292	3.4	83	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
20	4,6-Dinitro-2-methylphenol	0.007	0.001	85.7#	26#	0.03
21	4-Bromophenyl-phenylether	0.242	0.226	6.6	80	0.00
22	Hexachlorobenzene	0.288	0.319	-10.8	91	0.00
23	Atrazine	0.175	0.136	22.3	72	0.01
24	Pentachlorophenol	0.072	0.011	84.7#	19#	0.01
25	Phenanthrene	1.051	1.063	-1.1	83	0.00
26	Anthracene	0.924	0.851	7.9	79	0.00
27 SURR	Fluoranthene-d10	1.337	1.345	-0.6	83	0.00
28	Fluoranthene	1.279	1.363	-6.6	85	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
30	Pyrene	1.268	1.598	-26.0#	85	0.00
31 S	Terphenyl-d14	0.859	0.995	-15.8	79	0.00
32	Benzo(a)anthracene	1.231	1.098	10.8	64	0.00
33	Chrysene	1.286	1.292	-0.5	66	0.00
34	Bis(2-ethylhexyl)phthalate	0.558	0.404	27.6#	57	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	56	0.00
36	Indeno(1,2,3-cd)pyrene	1.109	0.675	39.1#	36#	0.00
37	Benzo(b)fluoranthene	1.360	1.292	5.0	54	0.00
38	Benzo(k)fluoranthene	1.283	1.662	-29.5#	73	0.00
39 C	Benzo(a)pyrene	1.222	1.150	5.9	53	0.00
40	Dibenzo(a,h)anthracene	0.859	0.523	39.1#	36#	0.02
41	Benzo(g,h,i)perylene	0.923	0.646	30.0#	40#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4