


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120921\
 Data File : BN017834.D
 Acq On : 09 Dec 2021 13:47
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 08 01:18:09 2021






AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2469

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.7	151027	PASS
68	69	0.00	2	1.9	2911	PASS
69	198	0.00	100	33.8	156104	PASS
70	69	0.00	2	0.6	859	PASS
127	198	10	80	38.3	176816	PASS
197	198	0.00	2	0.3	1240	PASS
198	198	100	100	100.0	461440	PASS
199	198	5	9	6.8	31568	PASS
275	198	10	60	27.8	128392	PASS
365	198	1	100	3.5	16290	PASS
441	198	0.01	100	12.1	55984	PASS
442	442	50	100	100.0	356907	PASS
443	442	15	24	18.4	65709	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120921\
Data File : BN017834.D
Acq On : 09 Dec 2021 13:47
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 10 08:38:50 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 09 12:31:20 2021
Response via : Initial Calibration



TIC: BN017834.D\data.ms

(70) Pentachlorophenol (C)

16.751min (-0.121) 9192.41 ng




response 249387

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.16
264.00	61.60	63.30
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120921\
Data File : BN017834.D
Acq On : 09 Dec 2021 13:47
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 10 08:38:50 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 09 12:31:20 2021
Response via : Initial Calibration



TIC: BN017834.D\data.ms

(77) Benzidine

19.351min (-0.115) 0.00 ng

response 2663192

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.46
183.00	13.20	11.18
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/9/2021	BNA_N	<u>BN017834.D</u>
Compound Name	Response	Retention Time
DDT	1412427	20.592
DDD	89206	20.151
DDE	4218	19.639
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
93424	1505851	6.20

Instrument :
BNA_N
ClientSampleId :
DFTPP