



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120921\
 Data File : BN017851.D
 Acq On : 10 Dec 2021 01:42
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 08 01:18:09 2021






AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.6	233592	PASS
68	69	0.00	2	1.9	4449	PASS
69	198	0.00	100	36.3	238165	PASS
70	69	0.00	2	0.6	1445	PASS
127	198	10	80	38.5	252288	PASS
197	198	0.00	2	0.5	3509	PASS
198	198	100	100	100.0	655701	PASS
199	198	5	9	6.6	43352	PASS
275	198	10	60	27.7	181845	PASS
365	198	1	100	3.5	23091	PASS
441	198	0.01	100	12.0	78704	PASS
442	442	50	100	100.0	471445	PASS
443	442	15	24	20.5	96576	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120921\
 Data File : BN017851.D
 Acq On : 10 Dec 2021 01:42
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Dec 10 08:40:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 09 12:31:20 2021
 Response via : Initial Calibration



TIC: BN017851.D\data.ms

(70) Pentachlorophenol (C)

16.745min (-0.127) 13697.44 ng




response 467599

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.85
264.00	61.60	58.71
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120921\
Data File : BN017851.D
Acq On : 10 Dec 2021 01:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 10 08:40:43 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 09 12:31:20 2021
Response via : Initial Calibration



TIC: BN017851.D\data.ms

(77) Benzidine

19.345min (-0.121) 0.00 ng

response 2802502

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.31
183.00	13.20	12.02
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/9/2021	BNA_N	<u>BN017851.D</u>
Compound Name	Response	Retention Time
DDT	1850297	20.592
DDD	91100	20.145
DDE	5326	19.639
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
96426	1946723	4.95

Instrument :
BNA_N
ClientSampleId :
DFTPP