

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120922\
 Data File : BN023104.D
 Acq On : 09 Dec 2022 13:09
 Operator : CG/JU
 Sample : PB149020BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB149020BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 12/10/2022
 Supervised By :Jagrut Upadhyay 12/13/2022

Quant Time: Dec 10 03:15:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 09 07:44:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.021	152	9681	0.400	ng	0.00
7) Naphthalene-d8	10.840	136	28505	0.400	ng	0.00
13) Acenaphthene-d10	14.666	164	14796	0.400	ng	0.00
19) Phenanthrene-d10	17.402	188	33614	0.400	ng	#-0.01
29) Chrysene-d12	21.589	240	26785	0.400	ng	# 0.00
35) Perylene-d12	24.053	264	24932	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.565	112	8439	0.468	ng	0.00
5) Phenol-d6	7.176	99	10764	0.470	ng	0.00
8) Nitrobenzene-d5	9.185	82	6630m	0.353	ng	0.00
11) 2-Methylnaphthalene-d10	12.424	152	13256	0.274	ng	0.00
14) 2,4,6-Tribromophenol	16.148	330	1093	0.204	ng	0.00
15) 2-Fluorobiphenyl	13.287	172	22244	0.376	ng	-0.01
27) Fluoranthene-d10	19.435	212	27840	0.354	ng	0.00
31) Terphenyl-d14	20.031	244	18927	0.435	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120922\
Data File : BN023104.D
Acq On : 09 Dec 2022 13:09
Operator : CG/JU
Sample : PB149020BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB149020BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 12/10/2022
Supervised By :Jagrut Upadhyay 12/13/2022

Quant Time: Dec 10 03:15:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 09 07:44:40 2022
Response via : Initial Calibration

