

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121621\
 Data File : BN018059.D
 Acq On : 16 Dec 2021 13:12
 Operator : CG/JU
 Sample : M5037-05
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C00W1

Quant Time: Dec 17 02:12:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN121621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 15 17:47:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.662	152	4567	0.400	ng/ul	0.00
4) Naphthalene-d8	10.447	136	14962	0.400	ng/ul	#-0.02
9) Acenaphthene-d10	14.308	164	9713	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.053	188	22073	0.400	ng/ul	-0.01
17) Chrysene-d12	21.263	240	18522	0.400	ng/ul	# 0.00
23) Perylene-d12	23.528	264	22091	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.139	96	24248	5.114	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.059	152	6518	0.374	ng/ul	0.00
18) Fluoranthene-d10	19.091	212	24224	0.459	ng/ul	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121621\
Data File : BN018059.D
Acq On : 16 Dec 2021 13:12
Operator : CG/JU
Sample : M5037-05
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C00W1

Quant Time: Dec 17 02:12:16 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN121621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 15 17:47:27 2021
Response via : Initial Calibration

