

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121624\
 Data File : BN035622.D
 Acq On : 16 Dec 2024 18:13
 Operator : RC/JU
 Sample : P5232-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E28R5

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/18/2024
 Supervised By :mohammad ahmed 12/21/2024

Quant Time: Dec 16 21:56:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN121624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 16 16:26:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.284	152	2389	0.400	ng/ul	0.00
4) Naphthalene-d8	10.031	136	7279	0.400	ng/ul	0.00
9) Acenaphthene-d10	13.944	164	5148	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.713	188	11153	0.400	ng/ul	0.00
17) Chrysene-d12	20.958	240	9210	0.400	ng/ul	# 0.00
23) Perylene-d12	23.045	264	10406	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.955	96	4205	1.985	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.631	152	993	0.096	ng/ul	0.00
18) Fluoranthene-d10	18.767	212	2264	0.084	ng/ul	0.00
Target Compounds						
						Qvalue
5) Naphthalene	10.080	128	2382m	0.133	ng/ul	
7) 2-Methylnaphthalene	11.708	142	1448	0.120	ng/ul	99
8) 1-Methylnaphthalene	11.934	142	580	0.047	ng/ul	100
10) Acenaphthylene	13.657	152	1180	0.059	ng/ul#	80
11) Acenaphthene	14.009	153	324	0.021	ng/ul#	92
14) Pentachlorophenol	16.379	266	59	0.021	ng/ul	86
15) Phenanthrene	16.755	178	8791	0.315	ng/ul	99
16) Anthracene	16.848	178	1982	0.076	ng/ul#	86
19) Fluoranthene	18.795	202	21198	0.644	ng/ul	100
20) Pyrene	19.163	202	21719	0.599	ng/ul	99
21) Benzo(a)anthracene	20.944	228	14639	0.479	ng/ul	95
22) Chrysene	20.993	228	19334m	0.610	ng/ul	
24) Benzo(b)fluoranthene	22.437	252	31771m	1.036	ng/ul	
25) Benzo(k)fluoranthene	22.472	252	10862m	0.320	ng/ul	
26) Benzo(a)pyrene	22.957	252	23287m	0.784	ng/ul	
27) Indeno(1,2,3-cd)pyrene	25.074	276	18076m	0.500	ng/ul	
28) Dibenzo(a,h)anthracene	25.084	278	4486m	0.162	ng/ul	
29) Benzo(g,h,i)perylene	25.688	276	22659	0.781	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121624\
Data File : BN035622.D
Acq On : 16 Dec 2024 18:13
Operator : RC/JU
Sample : P5232-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E28R5

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/18/2024
Supervised By :mohammad ahmed 12/21/2024

Quant Time: Dec 16 21:56:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN121624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 16 16:26:08 2024
Response via : Initial Calibration

