

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041024\
 Data File : BN030830.D
 Acq On : 11 Apr 2024 00:26
 Operator : MA/JU
 Sample : P2014-01
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E0MC7

Quant Time: Apr 11 00:56:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 09 12:10:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.681	152	170718	20.000	ng/u1	0.00
20) Naphthalene-d8	10.457	136	801922	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	559774	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.069	188	1243929	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	1219205	20.000	ng/u1	0.00
88) Perylene-d12	24.233	264	1352461	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.187	96	15035	4.129	ng/uL	0.00
4) Pyridine-d5	3.593	84	110375	10.110	ng/u1	0.00
7) Phenol-d5	6.846	99	105576	7.645	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.016	67	292910	34.339	ng/u1	0.00
11) 2-Chlorophenol-d4	7.210	132	299403	28.578	ng/u1	0.00
15) 4-Methylphenol-d8	8.381	113	211081	18.237	ng/u1	0.00
21) Nitrobenzene-d5	8.834	128	209096	37.588	ng/u1	0.00
24) 2-Nitrophenol-d4	9.551	143	204457	40.176	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.081	165	376498	33.794	ng/u1	0.00
31) 4-Chloroaniline-d4	10.598	131	542993	29.605	ng/u1	0.00
46) Dimethylphthalate-d6	13.734	166	1499173	40.113	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	1650221	38.403	ng/u1	0.00
54) 4-Nitrophenol-d4	14.534	143	40269	5.633	ng/u1	0.00
60) Fluorene-d10	15.316	176	1324623	40.502	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.433	200	185016	33.697	ng/u1	0.00
73) Anthracene-d10	17.169	188	2270807	42.369	ng/u1	0.00
81) Pyrene-d10	19.469	212	2736105	41.741	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.039	264	2719303	43.636	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041024\
Data File : BN030830.D
Acq On : 11 Apr 2024 00:26
Operator : MA/JU
Sample : P2014-01
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E0MC7

Quant Time: Apr 11 00:56:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 09 12:10:11 2024
Response via : Initial Calibration

