

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041024\
 Data File : BN030831.D
 Acq On : 11 Apr 2024 01:06
 Operator : MA/JU
 Sample : P2014-02
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E0MC9

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/12/2024
 Supervised By :mohammad ahmed 04/12/2024

Quant Time: Apr 11 01:39:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 09 12:10:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	167174	20.000	ng/u1	0.00
20) Naphthalene-d8	10.457	136	789786	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	565652	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.068	188	1213478	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	1249803	20.000	ng/u1	0.00
88) Perylene-d12	24.239	264	1306839	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.187	96	15214	4.267	ng/uL	0.00
4) Pyridine-d5	3.604	84	42298m	3.957	ng/u1	0.01
7) Phenol-d5	6.851	99	89811	6.641	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.016	67	265423	31.777	ng/u1	0.00
11) 2-Chlorophenol-d4	7.210	132	271367	26.452	ng/u1	0.00
15) 4-Methylphenol-d8	8.381	113	178772	15.773	ng/u1	0.00
21) Nitrobenzene-d5	8.834	128	186877	34.110	ng/u1	0.00
24) 2-Nitrophenol-d4	9.551	143	182985	36.510	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.081	165	327511	29.849	ng/u1	0.00
31) 4-Chloroaniline-d4	10.604	131	342511	18.961	ng/u1	0.00
46) Dimethylphthalate-d6	13.733	166	1278550	33.854	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	1474375	33.954	ng/u1	0.00
54) 4-Nitrophenol-d4	14.533	143	31787m	4.400	ng/u1	0.00
60) Fluorene-d10	15.316	176	1179585	35.693	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.433	200	160148	29.899	ng/u1	0.00
73) Anthracene-d10	17.168	188	2085511	39.888	ng/u1	0.00
81) Pyrene-d10	19.468	212	2544676	37.870	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.033	264	2414916	40.104	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041024\
Data File : BN030831.D
Acq On : 11 Apr 2024 01:06
Operator : MA/JU
Sample : P2014-02
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E0MC9

Quant Time: Apr 11 01:39:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 09 12:10:11 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/12/2024
Supervised By :mohammad ahmed 04/12/2024

