

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041024\
 Data File : BN030841.D
 Acq On : 11 Apr 2024 08:46
 Operator : MA/JU
 Sample : PB160146BL
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK146

Quant Time: Apr 11 10:40:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 09 12:10:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	155917	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.451	136	727587	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.316	164	511038	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.069	188	1194631	20.000	ng/u1	0.00
79) Chrysene-d12	21.310	240	1302422	20.000	ng/u1	0.00
88) Perylene-d12	24.239	264	1451154	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	21208	6.377	ng/uL	0.00
4) Pyridine-d5	3.587	84	326388	32.734	ng/u1	0.00
7) Phenol-d5	6.840	99	400460	31.750	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.010	67	245934	31.569	ng/u1	0.00
11) 2-Chlorophenol-d4	7.205	132	325623	34.032	ng/u1	0.00
15) 4-Methylphenol-d8	8.375	113	339954	32.159	ng/u1	-0.01
21) Nitrobenzene-d5	8.828	128	168322	33.349	ng/u1	0.00
24) 2-Nitrophenol-d4	9.546	143	171851	37.219	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.075	165	336483	33.288	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.599	131	531040	31.911	ng/u1	0.00
46) Dimethylphthalate-d6	13.734	166	1178710	34.546	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	1321787	33.693	ng/u1	0.00
54) 4-Nitrophenol-d4	14.528	143	196831	30.159	ng/u1	0.00
60) Fluorene-d10	15.316	176	1032887	34.594	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.434	200	132200	25.071	ng/u1	0.00
73) Anthracene-d10	17.169	188	1766333	34.316	ng/u1	0.00
81) Pyrene-d10	19.469	212	2227946	31.817	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.045	264	2361554	35.318	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041024\
Data File : BN030841.D
Acq On : 11 Apr 2024 08:46
Operator : MA/JU
Sample : PB160146BL
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK146

Quant Time: Apr 11 10:40:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 09 12:10:11 2024
Response via : Initial Calibration

