


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041024\
Data File : BN030809.D
Acq On : 10 Apr 2024 09:10
Operator : MA/JU
Sample : DFTPP316
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP316

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Apr 09 12:10:11 2024



AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.1	148941	PASS
68	69	0.00	2	1.7	2696	PASS
69	198	0.00	100	35.5	159696	PASS
70	69	0.00	2	0.5	824	PASS
127	198	10	80	47.0	211221	PASS
197	198	0.00	2	0.4	1728	PASS
198	198	100	100	100.0	449429	PASS
199	198	5	9	6.5	29317	PASS
275	198	10	60	25.7	115480	PASS
365	198	1	100	3.7	16765	PASS
441	198	0.01	100	11.7	52483	PASS
442	198	50	100	70.7	317952	PASS
443	442	15	24	19.6	62195	PASS

m/z	Abundance
40.00	1541.0
41.00	207.0
43.00	328.0
44.00	4752.0
44.90	283.0
48.90	376.0
50.90	182.0
55.00	229.0
73.00	537.0
75.00	191.0
77.00	391.0
78.00	162.0
83.90	326.0
85.80	276.0
96.00	404.0
118.90	158.0
133.00	551.0
146.90	307.0
176.80	174.0
190.80	538.0
192.90	426.0
207.00	4549.0
207.90	1031.0
208.90	450.0
281.00	606.0
282.10	202.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

m/z	Abundance
35.90	181.0
37.00	1117.0
38.10	2347.0
39.00	14212.0
40.00	1516.0
40.90	404.0
43.10	188.0
44.00	3794.0
45.00	501.0
49.10	1686.0
50.00	51856.0
51.00	190592.0
52.00	9581.0
53.00	419.0
55.00	1214.0
56.00	5560.0
56.90	12225.0
57.90	611.0
58.90	257.0
59.90	338.0
61.00	2432.0
62.00	2717.0
63.00	8053.0
64.00	1087.0
65.00	3760.0
66.00	251.0
66.90	367.0
68.00	3381.0
68.90	204416.0
70.00	1108.0
73.00	2071.0
74.00	22696.0
75.00	31752.0
76.00	11316.0
77.00	238912.0
78.00	15787.0
79.00	17232.0
79.90	11867.0
81.00	16704.0
82.00	3797.0
83.00	3707.0
83.90	818.0
84.90	2808.0
86.00	5080.0
87.00	2436.0
88.00	1130.0
88.90	493.0
91.00	3530.0
91.90	4021.0
93.00	29056.0
93.90	2136.0
94.90	440.0
95.90	1339.0
96.90	589.0
98.00	22472.0
99.00	19264.0
100.00	1411.0
101.00	9051.0
101.90	607.0
103.00	3691.0
104.00	6395.0
105.00	5950.0
106.00	2032.0
107.00	75056.0
108.00	10904.0
109.00	2345.0
110.00	129832.0
111.00	21088.0
112.00	2555.0
113.00	1083.0
114.50	265.0
114.90	302.0
116.00	4207.0
117.00	68768.0
118.00	4753.0
118.90	695.0
120.00	943.0
120.90	446.0
121.90	4589.0
122.90	7580.0
123.90	3644.0
125.00	2962.0
127.00	257600.0
128.00	20640.0
129.00	96616.0
129.90	8779.0
130.90	1545.0
132.00	839.0
132.90	618.0
134.00	2766.0
135.00	7880.0
136.00	2939.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

137.00	5491.0
137.90	893.0
138.90	459.0
140.00	1520.0
141.00	12677.0
141.90	3891.0
143.00	3012.0
144.00	850.0
144.90	727.0
145.90	2649.0
147.00	6971.0
148.00	16011.0
149.00	3861.0
150.00	1010.0
151.00	1725.0
151.90	799.0
153.00	4339.0
154.00	2986.0
155.00	6948.0
156.00	9906.0
157.00	1994.0
158.00	2383.0
159.00	1670.0
160.00	4288.0
161.00	5834.0
162.00	1841.0
162.80	319.0
163.90	779.0
165.00	4813.0
166.00	3695.0
167.00	26816.0
168.00	18912.0
169.00	2743.0
169.90	818.0
171.00	1163.0
171.90	2227.0
173.00	2966.0
174.00	6012.0
175.00	10622.0
176.00	2899.0
177.00	5472.0
178.00	1404.0
179.00	18728.0
180.00	14645.0
181.00	5821.0
181.90	932.0
182.90	517.0
184.00	1582.0
185.00	9835.0
186.00	72960.0
187.00	19992.0
188.00	2598.0
189.00	4717.0
189.90	869.0
191.00	2538.0
192.00	6101.0
193.00	6677.0
194.00	1495.0
195.10	831.0
196.00	13987.0
198.00	511552.0
199.00	32824.0
200.00	2272.0
201.40	2117.0
202.90	3861.0
204.00	18496.0
205.00	32472.0
206.00	130216.0
207.00	21560.0
208.00	5267.0
209.00	2020.0
210.10	2076.0
211.00	5564.0
213.00	433.0
215.00	1374.0
216.00	2668.0
216.90	33120.0
218.00	4485.0
218.90	374.0
221.00	18752.0
221.80	4779.0
223.00	7657.0
224.00	71528.0
225.00	19232.0
226.00	2408.0
227.00	31520.0
228.00	4180.0
229.00	6355.0
229.90	796.0
231.00	2848.0
231.80	404.0
232.90	615.0
233.90	1946.0
235.00	2472.0
235.90	1446.0
237.00	3555.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

238.00	511.0
239.00	1194.0
240.00	1032.0
241.00	1963.0
242.00	3537.0
243.00	4225.0
244.00	53976.0
245.00	7476.0
246.00	10833.0
246.90	2509.0
247.90	584.0
248.90	2317.0
249.90	430.0
251.00	507.0
252.00	386.0
253.00	1306.0
255.00	289280.0
256.00	41008.0
257.00	3597.0
257.90	17384.0
258.90	2800.0
259.90	454.0
260.90	449.0
262.80	178.0
264.00	457.0
264.90	6498.0
265.90	1066.0
266.80	169.0
268.00	170.0
270.00	251.0
271.00	803.0
272.00	764.0
273.00	8299.0
274.00	21088.0
275.00	120856.0
276.00	15435.0
277.00	11237.0
277.90	2353.0
279.00	382.0
281.00	824.0
282.10	501.0
283.10	1158.0
284.00	693.0
285.00	1901.0
286.00	343.0
288.00	153.0
289.00	466.0
289.90	257.0
291.00	314.0
291.90	666.0
293.00	2602.0
293.90	696.0
295.00	845.0
296.00	38904.0
297.00	4956.0
297.80	340.0
300.90	499.0
302.00	683.0
303.00	4146.0
303.90	999.0
304.90	200.0
307.90	672.0
308.90	286.0
309.90	489.0
311.80	180.0
312.90	237.0
314.00	1883.0
315.00	4239.0
316.00	2570.0
317.00	557.0
320.00	201.0
321.00	974.0
322.10	534.0
323.00	10194.0
324.00	1749.0
325.00	220.0
325.90	248.0
326.90	2289.0
328.00	1118.0
328.90	204.0
332.00	814.0
333.00	1027.0
334.00	7595.0
335.00	1791.0
335.90	194.0
338.90	175.0
340.00	171.0
341.00	1330.0
342.00	398.0
342.90	154.0
345.90	2295.0
346.90	519.0
350.90	153.0
352.00	3106.0
353.00	2077.0
353.90	3102.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

355.00	673.0
358.80	205.0
364.90	15732.0
365.90	2334.0
366.90	252.0
369.90	332.0
370.90	943.0
372.00	5759.0
372.90	1622.0
373.90	244.0
382.90	1764.0
384.00	534.0
390.00	646.0
390.90	573.0
392.00	373.0
400.90	407.0
402.00	2565.0
403.00	3302.0
403.90	1265.0
404.90	283.0
421.00	3059.0
421.90	2879.0
423.00	17112.0
423.90	5408.0
425.10	628.0
439.00	167.0
441.00	47656.0
442.00	289536.0
443.00	57360.0
444.00	5126.0
445.00	355.0

Instrument :

BNA_N

ClientSampleId :

DFTPP316

m/z	Abundance
35.00	206.0
37.00	910.0
38.00	2324.0
39.00	13082.0
40.00	1664.0
41.00	552.0
42.90	277.0
44.00	3851.0
45.00	486.0
47.80	156.0
49.00	1753.0
50.00	47168.0
51.00	172416.0
52.00	8802.0
52.90	321.0
54.90	994.0
55.90	4793.0
57.00	10823.0
58.00	521.0
58.90	203.0
59.90	192.0
61.00	2189.0
62.00	2549.0
63.00	6784.0
64.00	958.0
65.00	3720.0
65.90	249.0
67.10	305.0
68.00	2991.0
69.00	182016.0
70.00	837.0
72.90	1688.0
74.00	19344.0
75.00	30568.0
76.00	10202.0
77.00	217984.0
78.00	14201.0
79.00	15036.0
80.00	11292.0
81.00	14616.0
82.00	3875.0
83.00	3350.0
84.00	551.0
85.00	2656.0
86.00	4622.0
86.90	2259.0
87.90	990.0
88.90	389.0
91.00	3876.0
92.00	3987.0
93.00	27408.0
94.00	1861.0
95.00	391.0
96.00	1826.0
96.90	473.0
98.00	22992.0
99.00	16704.0
99.90	1697.0
101.00	8788.0
102.00	550.0
103.00	2637.0
104.00	6080.0
104.90	5486.0
106.00	1707.0
107.00	68760.0
108.00	10985.0
109.00	2216.0
110.00	124336.0
111.00	18760.0
112.00	2358.0
112.90	779.0
115.00	367.0
116.00	4391.0
116.90	65392.0
117.90	4983.0
119.00	731.0
120.00	1065.0
120.80	409.0
121.90	4842.0
123.00	7885.0
124.00	3488.0
125.00	3207.0
127.00	243840.0
128.00	19184.0
129.00	95296.0
130.00	7983.0
131.00	1442.0
132.00	920.0
133.00	847.0
133.90	2546.0
134.90	8376.0
136.00	3023.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

136.90	4187.0
138.00	1020.0
138.90	603.0
139.90	1318.0
140.90	12964.0
142.00	4271.0
143.00	2653.0
144.00	871.0
145.00	594.0
146.00	2477.0
146.90	6566.0
148.00	16368.0
149.00	3598.0
150.00	881.0
151.00	1490.0
151.90	1160.0
152.90	3976.0
154.00	3166.0
155.00	7439.0
156.00	11117.0
156.90	1945.0
157.90	2375.0
158.90	1981.0
160.00	3703.0
161.00	6169.0
161.90	1776.0
163.00	636.0
164.00	704.0
165.00	5161.0
166.00	4163.0
167.00	29488.0
168.00	17600.0
169.00	2489.0
170.00	915.0
170.90	1271.0
172.00	2520.0
173.00	3086.0
173.90	5841.0
175.00	10175.0
176.00	2881.0
177.00	5694.0
178.10	1708.0
178.90	17880.0
180.00	14224.0
181.00	6373.0
182.00	839.0
182.90	461.0
184.00	1858.0
185.00	9379.0
186.00	74808.0
187.00	21920.0
188.00	2446.0
188.90	4404.0
189.90	954.0
191.00	2610.0
192.00	6680.0
193.00	7846.0
193.90	1493.0
195.00	758.0
196.00	15254.0
197.00	3330.0
197.90	539520.0
199.00	35072.0
200.00	2918.0
201.50	2176.0
203.00	3504.0
204.00	19928.0
205.00	32696.0
206.00	142528.0
207.00	21648.0
208.00	5416.0
208.90	1766.0
210.00	2364.0
211.00	5224.0
212.90	493.0
215.00	1855.0
216.00	2895.0
216.90	37952.0
217.90	5277.0
218.90	574.0
221.00	20704.0
221.80	5304.0
223.00	8003.0
224.00	78168.0
225.00	20656.0
226.00	2455.0
227.00	33992.0
228.00	4696.0
229.00	6427.0
230.00	826.0
231.10	2826.0
231.90	521.0
233.00	536.0
234.00	2034.0
235.00	2435.0
235.90	1497.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

237.00	2944.0
238.00	497.0
239.00	1433.0
240.00	1038.0
241.00	1752.0
242.00	4523.0
243.00	4471.0
244.00	59712.0
245.00	8050.0
245.90	12137.0
247.00	2346.0
248.10	519.0
249.00	2217.0
249.90	300.0
250.90	570.0
251.90	550.0
253.00	1352.0
255.00	321344.0
256.00	46144.0
257.00	3120.0
258.00	18704.0
259.00	2841.0
259.90	661.0
261.00	563.0
261.90	162.0
262.90	169.0
264.00	471.0
265.00	7672.0
265.90	1181.0
267.00	215.0
267.90	151.0
269.90	420.0
271.00	883.0
271.90	904.0
273.00	9475.0
274.00	24864.0
275.00	136384.0
276.00	19136.0
277.00	13068.0
278.00	2668.0
279.10	472.0
281.00	736.0
282.00	513.0
283.00	1549.0
283.90	1048.0
285.00	2403.0
285.90	559.0
287.90	184.0
289.00	575.0
289.80	517.0
291.00	437.0
292.00	637.0
292.90	2638.0
294.00	937.0
295.10	1142.0
295.90	39360.0
296.90	5462.0
298.00	304.0
301.00	564.0
302.00	740.0
303.00	4618.0
304.00	1480.0
305.00	307.0
307.90	500.0
308.90	386.0
309.80	529.0
311.80	203.0
313.00	436.0
314.00	2118.0
314.90	4664.0
316.00	2795.0
316.90	552.0
319.80	185.0
321.00	1210.0
322.00	792.0
323.00	12725.0
324.00	2426.0
325.00	342.0
325.90	264.0
327.00	2728.0
327.90	1302.0
328.90	180.0
332.00	1088.0
332.90	1303.0
334.00	8413.0
334.90	2031.0
335.90	220.0
339.00	225.0
340.00	275.0
340.90	1259.0
341.90	437.0
345.90	3407.0
347.00	637.0
351.00	317.0
351.90	4299.0
353.00	2936.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

354.00	4513.0
354.90	1090.0
358.90	287.0
363.90	244.0
365.00	20760.0
366.00	2723.0
366.90	166.0
369.80	310.0
370.90	1085.0
372.00	7267.0
373.00	1784.0
373.90	258.0
376.90	159.0
383.00	1992.0
384.00	715.0
390.00	1066.0
391.00	808.0
392.00	388.0
396.90	171.0
401.00	585.0
402.00	2673.0
403.00	4406.0
404.00	1681.0
404.90	399.0
415.00	239.0
421.00	3668.0
422.00	3834.0
422.90	26088.0
423.90	6151.0
424.90	537.0
439.00	175.0
441.00	62200.0
442.00	402112.0
443.00	78728.0
444.00	7259.0
444.90	313.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

m/z	Abundance
35.90	156.0
37.00	400.0
38.00	1080.0
39.00	5964.0
39.90	1531.0
41.00	342.0
43.00	164.0
44.00	4165.0
44.90	317.0
49.00	1145.0
50.00	22360.0
51.00	84360.0
52.00	4776.0
53.00	156.0
54.90	562.0
56.00	2314.0
57.00	5951.0
57.90	288.0
61.00	1089.0
62.00	1275.0
63.00	3535.0
64.00	487.0
64.90	1463.0
67.00	290.0
67.90	1716.0
69.00	92656.0
70.00	528.0
73.00	1069.0
74.00	9764.0
74.90	14711.0
76.00	5127.0
77.00	108832.0
78.00	7755.0
79.00	7135.0
80.00	5111.0
81.00	8081.0
82.00	1894.0
83.00	1851.0
83.90	482.0
85.00	1539.0
86.00	2310.0
87.00	1355.0
87.90	542.0
88.90	227.0
90.90	1923.0
91.90	1975.0
93.00	13400.0
93.90	851.0
94.90	203.0
96.00	1044.0
97.10	283.0
98.00	10271.0
99.00	9233.0
100.00	626.0
100.90	4457.0
102.00	317.0
102.90	1318.0
104.00	3304.0
105.00	2886.0
106.00	803.0
107.00	35520.0
108.00	5511.0
109.00	1113.0
110.00	66880.0
111.00	9879.0
111.90	1082.0
113.00	542.0
114.80	169.0
115.90	2189.0
117.00	34544.0
118.00	2531.0
118.90	402.0
119.90	531.0
122.00	2638.0
122.90	4025.0
123.90	1525.0
125.00	1411.0
127.00	132224.0
128.00	9208.0
129.00	50920.0
129.90	4269.0
130.90	994.0
132.00	528.0
132.90	513.0
134.00	1453.0
135.00	4066.0
135.90	1492.0
137.00	2662.0
137.90	389.0
138.70	268.0
140.10	544.0
141.00	6777.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

142.00	2061.0
143.00	1868.0
144.00	355.0
145.00	441.0
146.00	1380.0
147.00	3667.0
147.90	9549.0
149.00	2063.0
149.90	640.0
151.10	860.0
151.70	559.0
152.90	2309.0
153.90	2069.0
155.00	4287.0
156.00	5707.0
157.00	999.0
157.90	1254.0
158.80	982.0
160.00	2344.0
161.00	3674.0
161.90	950.0
162.90	397.0
163.90	405.0
164.90	2853.0
166.00	1803.0
167.00	14769.0
167.90	10501.0
169.00	1487.0
169.80	349.0
170.90	745.0
172.00	1480.0
173.00	1754.0
173.90	2986.0
175.00	5366.0
176.00	1445.0
177.00	3015.0
178.00	1025.0
178.90	10329.0
180.00	7433.0
181.00	3608.0
182.00	524.0
182.50	283.0
184.00	1003.0
184.90	5535.0
186.00	41776.0
187.00	11363.0
188.00	1275.0
188.90	2733.0
189.90	505.0
191.00	1957.0
192.00	3730.0
193.00	4461.0
193.90	838.0
196.00	8329.0
197.00	1854.0
197.90	297216.0
198.90	20056.0
200.00	1486.0
201.40	1417.0
202.90	2123.0
204.00	10399.0
205.00	17872.0
206.00	82056.0
207.00	15500.0
208.00	3659.0
209.00	1537.0
210.10	1604.0
211.00	3117.0
212.80	221.0
214.80	763.0
216.00	1723.0
216.90	20552.0
217.90	2975.0
218.80	196.0
220.10	473.0
221.00	12697.0
221.70	3345.0
223.00	4869.0
224.00	45888.0
225.00	11487.0
225.90	1066.0
227.00	18744.0
227.90	2373.0
229.00	3849.0
230.00	538.0
231.00	1728.0
231.90	220.0
232.90	296.0
233.90	1480.0
235.00	1626.0
235.90	987.0
237.00	1742.0
237.90	255.0
238.90	1040.0
240.00	639.0
241.00	1002.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

242.00	2467.0
243.00	2094.0
244.00	38944.0
245.00	4732.0
246.00	7560.0
247.00	1237.0
247.90	268.0
248.90	1196.0
249.90	194.0
250.90	330.0
251.90	414.0
252.90	949.0
255.00	192064.0
256.00	27408.0
256.90	2040.0
258.00	12128.0
259.00	1621.0
259.90	292.0
260.90	245.0
263.10	155.0
263.90	429.0
264.90	5003.0
266.00	737.0
267.00	215.0
268.00	155.0
270.00	268.0
270.90	551.0
271.90	612.0
273.00	5965.0
274.00	16282.0
275.00	89200.0
276.00	11446.0
277.00	7913.0
278.00	1400.0
279.00	346.0
280.90	565.0
282.00	409.0
283.00	664.0
284.00	643.0
285.10	1245.0
286.10	273.0
288.80	366.0
289.80	353.0
290.90	222.0
291.90	451.0
292.90	1722.0
294.00	456.0
296.00	26544.0
296.90	3438.0
297.90	322.0
300.90	286.0
302.00	409.0
303.00	2980.0
303.90	746.0
305.00	206.0
307.90	319.0
308.90	252.0
309.90	406.0
312.90	390.0
314.00	1472.0
314.90	2887.0
316.00	1696.0
317.10	237.0
321.00	855.0
322.00	493.0
323.00	8827.0
324.00	1535.0
325.00	179.0
326.10	184.0
326.90	1903.0
327.90	970.0
329.00	244.0
331.90	606.0
332.90	724.0
334.00	5421.0
334.90	1277.0
335.90	201.0
338.90	156.0
340.00	169.0
341.00	955.0
341.90	262.0
345.90	1937.0
346.90	324.0
351.10	191.0
352.00	2577.0
352.90	2045.0
354.00	3116.0
355.00	673.0
358.90	199.0
364.90	13802.0
365.90	2057.0
369.90	305.0
370.90	739.0
372.00	5095.0
372.90	1131.0
377.10	171.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316

382.90	1641.0
383.90	472.0
389.90	637.0
391.00	446.0
392.00	414.0
400.90	379.0
402.00	2338.0
403.00	3047.0
403.90	998.0
404.90	171.0
415.00	154.0
421.00	2605.0
422.00	2837.0
423.00	14762.0
424.00	4103.0
424.90	422.0
441.00	47592.0
442.00	262208.0
443.00	50496.0
443.90	5513.0
444.70	262.0

Instrument :
BNA_N
ClientSampleId :
DFTPP316