

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041321\
 Data File : BN014267.D
 Acq On : 13 Apr 2021 10:11
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4003

Quant Time: Apr 14 01:26:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN041221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 12 17:52:31 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	5790	0.40	ng/ul	0.00
4) Naphthalene-d8	10.48	136	19173	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.34	164	9293	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.08	188	18700	0.40	ng/ul	0.00
17) Chrysene-d12	21.28	240	15978	0.40	ng/ul	0.00
23) Perylene-d12	23.52	264	13795	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.23	96	2569	0.41	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.08	152	11083	0.37	ng/ul	0.00
18) Fluoranthene-d10	19.12	212	19379	0.39	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.27	88	2564	0.399	ng/ul#	88
5) Naphthalene	10.53	128	22014	0.388	ng/ul	100
7) 2-Methylnaphthalene	12.15	142	14213	0.393	ng/ul	99
8) 1-Methylnaphthalene	12.37	142	13718	0.394	ng/ul	100
10) Acenaphthylene	14.06	152	15819	0.364	ng/ul	99
11) Acenaphthene	14.40	153	13663	0.403	ng/ul	99
12) Fluorene	15.39	166	14730	0.389	ng/ul	99
14) Pentachlorophenol	16.73	266	1759	0.291	ng/ul	100
15) Phenanthrene	17.13	178	24009	0.398	ng/ul	99
16) Anthracene	17.22	178	19461	0.368	ng/ul	99
19) Fluoranthene	19.15	202	25238	0.394	ng/ul	99
20) Pyrene	19.51	202	25532	0.391	ng/ul	98
21) Benzo(a)anthracene	21.27	228	21630	0.366	ng/ul	100
22) Chrysene	21.32	228	24543	0.400	ng/ul	99
24) Benzo(b)fluoranthene	22.85	252	24379	0.392	ng/ul	99
25) Benzo(k)fluoranthene	22.89	252	25454	0.403	ng/ul	100
26) Benzo(a)pyrene	23.42	252	19963	0.388	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	25.76	276	29760	0.419	ng/ul	98
28) Dibenzo(a,h)anthracene	25.78	278	23542	0.415	ng/ul	98
29) Benzo(g,h,i)perylene	26.45	276	24341	0.407	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041321\
Data File : BN014267.D
Acq On : 13 Apr 2021 10:11
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4003

Quant Time: Apr 14 01:26:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN041221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 12 17:52:31 2021
Response via : Initial Calibration

