


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041321\
 Data File : BN014266.D
 Acq On : 13 Apr 2021 09:30
 Operator : CG/JU
 Sample : DFTPP31
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP031

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN041221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 14 11:00:32 2021



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.0	63117	PASS
68	69	0.00	2	1.9	1392	PASS
69	198	0.00	100	42.4	74480	PASS
70	69	0.00	2	0.7	506	PASS
127	198	10	80	51.7	90760	PASS
197	198	0.00	2	0.5	887	PASS
198	198	100	100	100.0	175530	PASS
199	198	5	9	7.2	12570	PASS
275	198	10	60	26.4	46418	PASS
365	198	1	100	3.4	5992	PASS
441	198	0.01	100	13.2	23140	PASS
442	198	50	100	81.4	142864	PASS
443	442	15	24	18.8	26864	PASS

Scan 2450 (17.498 min): BN014266.D
DFTPP31

m/z	Abundance
39.90	680.0
43.90	1251.0
48.90	321.0
207.00	974.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

m/z	Abundance
37.00	308.0
38.00	869.0
39.00	4550.0
39.90	961.0
40.90	329.0
44.00	1183.0
48.90	720.0
50.00	14469.0
51.00	55576.0
52.00	2882.0
55.00	424.0
56.00	1872.0
57.00	3864.0
60.90	820.0
61.90	939.0
63.00	2413.0
63.90	378.0
65.00	1023.0
68.00	1195.0
68.90	64872.0
69.90	404.0
73.00	1025.0
74.00	7107.0
75.00	11066.0
76.00	3935.0
77.00	77520.0
78.00	6178.0
78.90	4758.0
80.00	4137.0
81.00	5338.0
81.90	1492.0
83.00	1443.0
83.90	511.0
84.90	754.0
85.90	1870.0
87.00	793.0
87.90	349.0
91.00	1222.0
91.90	1543.0
92.90	9307.0
94.00	845.0
95.90	497.0
98.00	7162.0
98.90	7180.0
99.90	660.0
100.90	2989.0
103.00	891.0
103.90	1836.0
104.90	1639.0
105.90	552.0
107.00	22584.0
108.00	3112.0
109.00	749.0
109.90	37992.0
111.00	6076.0
112.00	772.0
115.90	1188.0
117.00	20760.0
118.00	1597.0
121.90	1227.0
122.90	2229.0
123.90	1051.0
124.90	1050.0
127.00	74552.0
128.00	5603.0
129.00	27864.0
129.90	2584.0
131.00	462.0
133.90	777.0
135.00	2345.0
136.00	873.0
136.90	1889.0
137.80	324.0
140.90	3723.0
141.90	1242.0
142.90	857.0
143.90	347.0
145.90	612.0
147.00	1954.0
147.90	4838.0
148.90	1311.0
149.90	386.0
151.00	530.0
153.00	1014.0
153.90	888.0
155.00	2029.0
156.00	3249.0
156.90	483.0
157.90	592.0
158.90	547.0
160.00	1128.0
161.00	1797.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

162.00	635.0
164.90	1427.0
165.90	939.0
167.00	8252.0
167.90	8668.0
169.00	919.0
169.90	305.0
170.80	309.0
171.90	679.0
172.90	875.0
174.00	1288.0
175.00	2791.0
176.00	681.0
176.90	1318.0
177.90	416.0
178.90	5341.0
180.00	4503.0
181.00	1657.0
183.90	427.0
185.00	2425.0
186.00	19016.0
187.00	5649.0
188.00	690.0
188.90	1298.0
191.00	696.0
192.00	1866.0
193.00	2132.0
193.90	477.0
196.00	3689.0
196.90	746.0
197.90	131584.0
198.90	9650.0
199.90	727.0
201.40	590.0
202.90	1117.0
204.00	5107.0
205.00	8558.0
206.00	34440.0
207.00	5996.0
208.00	1342.0
208.90	713.0
210.10	554.0
211.00	1491.0
214.90	385.0
216.00	885.0
216.90	9519.0
217.90	1418.0
221.00	5266.0
221.90	1213.0
223.00	2285.0
224.00	18136.0
225.00	5258.0
225.90	727.0
226.90	8599.0
228.00	1216.0
229.00	1829.0
229.80	307.0
231.00	731.0
233.90	451.0
235.00	648.0
235.90	380.0
236.90	1182.0
238.90	359.0
240.10	366.0
240.90	472.0
241.90	1044.0
243.00	1089.0
244.00	15259.0
245.00	2061.0
245.90	2531.0
246.90	643.0
248.90	594.0
252.90	331.0
253.90	405.0
255.00	74328.0
256.00	11558.0
257.00	997.0
258.00	4006.0
258.90	761.0
264.90	1758.0
273.00	2232.0
274.00	5465.0
275.00	29952.0
276.00	4182.0
276.90	2645.0
277.90	598.0
282.90	305.0
285.00	512.0
292.90	609.0
296.00	8186.0
297.00	1391.0
303.00	1039.0
304.00	305.0
314.00	542.0
315.00	869.0
315.90	636.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

322.90	2498.0
323.90	511.0
327.00	485.0
327.90	394.0
334.00	1867.0
335.00	452.0
341.00	351.0
345.90	585.0
352.00	641.0
353.00	471.0
354.00	787.0
364.90	3058.0
365.90	433.0
371.90	1239.0
372.90	327.0
401.90	441.0
402.90	646.0
403.90	383.0
421.00	566.0
422.00	732.0
423.00	4279.0
424.00	2143.0
425.00	306.0
441.10	11596.0
442.00	71920.0
443.00	14608.0
444.00	1492.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

m/z	Abundance
37.00	417.0
37.90	1087.0
39.00	6145.0
40.00	861.0
41.00	317.0
43.90	1170.0
48.90	868.0
50.00	20064.0
51.00	76640.0
52.00	4399.0
54.90	583.0
55.90	2317.0
57.00	5309.0
57.90	330.0
60.90	915.0
62.00	1253.0
63.00	3429.0
64.00	524.0
65.00	1775.0
68.00	1789.0
68.90	90544.0
70.00	556.0
73.00	918.0
74.00	9893.0
75.00	15474.0
76.00	5602.0
77.00	108168.0
78.00	7918.0
79.00	6879.0
79.90	5873.0
81.00	7443.0
82.00	1941.0
83.00	1858.0
83.80	513.0
85.00	1220.0
86.00	2318.0
87.00	940.0
87.90	445.0
91.00	1701.0
91.90	2098.0
93.00	12884.0
93.90	1001.0
95.00	358.0
95.90	713.0
97.90	9568.0
99.00	9301.0
99.90	711.0
101.00	4398.0
102.90	1256.0
103.90	2551.0
104.90	2828.0
105.90	774.0
107.00	33496.0
108.00	4875.0
109.00	817.0
110.00	56544.0
110.90	9244.0
111.90	1034.0
113.00	470.0
115.90	1543.0
117.00	31192.0
117.90	2534.0
119.00	318.0
120.00	459.0
122.00	2046.0
122.90	3121.0
123.90	1706.0
124.90	1385.0
126.00	409.0
127.00	109112.0
128.00	8346.0
129.00	41664.0
130.00	3519.0
131.00	734.0
132.00	415.0
134.00	1337.0
135.00	3461.0
136.00	1589.0
137.00	2253.0
137.90	543.0
139.90	534.0
140.90	5992.0
142.00	1808.0
142.90	1187.0
144.00	353.0
146.00	1064.0
147.00	2825.0
147.90	7847.0
149.00	1746.0
149.90	478.0
151.00	728.0
151.80	460.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

153.00	1791.0
154.00	1540.0
155.00	3382.0
156.00	4883.0
157.00	882.0
157.90	1054.0
158.90	876.0
159.90	1823.0
161.00	2612.0
162.00	773.0
164.90	2255.0
165.90	1412.0
167.00	12811.0
168.00	10220.0
168.90	1467.0
170.00	375.0
171.00	494.0
172.00	1049.0
172.90	1462.0
174.00	2730.0
175.00	4700.0
176.00	1000.0
176.90	2014.0
177.90	757.0
178.90	8709.0
180.00	6652.0
181.00	2791.0
181.90	398.0
183.90	605.0
185.00	4199.0
186.00	31736.0
187.00	8703.0
188.00	1101.0
189.00	2119.0
189.90	379.0
191.00	1153.0
191.90	2902.0
193.00	3214.0
194.00	868.0
194.90	355.0
195.90	5669.0
196.90	1032.0
197.90	218816.0
199.00	15579.0
199.90	1155.0
201.30	870.0
202.90	1694.0
204.00	8699.0
205.00	13867.0
206.00	60368.0
207.00	8132.0
208.00	2219.0
209.00	830.0
210.00	935.0
210.90	2195.0
211.90	424.0
214.90	598.0
216.00	1248.0
217.00	15250.0
217.90	2333.0
221.00	8854.0
221.90	2067.0
223.00	3488.0
224.00	35312.0
225.00	8840.0
226.00	1159.0
227.00	13620.0
228.00	2219.0
229.00	2819.0
229.90	401.0
231.00	1125.0
234.00	985.0
235.00	1001.0
236.00	762.0
237.00	1814.0
238.90	640.0
240.00	475.0
240.90	615.0
241.90	1573.0
243.00	1905.0
244.00	26424.0
245.00	3743.0
246.00	5287.0
246.90	1076.0
249.00	982.0
253.00	521.0
254.00	771.0
255.00	127208.0
256.00	18168.0
256.90	1416.0
257.90	7567.0
259.00	1267.0
265.00	3112.0
265.90	405.0
272.10	338.0
273.00	4225.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

274.00	10021.0
275.00	55144.0
276.00	7464.0
277.00	4375.0
277.90	1088.0
278.90	330.0
283.00	560.0
284.00	427.0
285.00	875.0
292.90	1227.0
296.00	14625.0
297.00	2216.0
301.90	332.0
303.00	1887.0
304.00	422.0
313.90	710.0
314.90	1610.0
316.00	1129.0
321.00	413.0
323.00	4383.0
324.00	930.0
326.90	1090.0
328.00	462.0
332.00	356.0
332.90	435.0
334.00	3201.0
335.00	797.0
341.00	594.0
345.90	1321.0
352.00	1478.0
353.00	1014.0
354.00	1382.0
355.00	326.0
365.00	7271.0
365.90	1008.0
370.90	407.0
372.00	2873.0
373.00	616.0
383.00	705.0
389.90	359.0
401.90	860.0
403.00	1744.0
403.90	519.0
420.90	1277.0
421.90	1122.0
423.00	8728.0
424.00	3135.0
425.00	540.0
441.00	26296.0
442.00	152000.0
443.00	27680.0
444.00	3095.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

m/z	Abundance
38.00	850.0
39.00	4203.0
39.90	987.0
40.90	324.0
44.00	1266.0
48.90	718.0
50.00	14458.0
51.00	57136.0
52.00	2996.0
55.00	486.0
56.00	1712.0
57.00	4427.0
61.00	727.0
62.00	900.0
63.00	2663.0
64.00	389.0
65.00	1286.0
67.90	1192.0
69.00	68024.0
70.00	559.0
73.00	1041.0
74.00	7631.0
75.00	11188.0
76.00	4058.0
77.00	83416.0
78.00	5595.0
79.00	5111.0
79.90	4400.0
80.90	5646.0
81.90	1425.0
83.00	1400.0
83.90	395.0
84.90	1257.0
85.90	1784.0
86.90	794.0
91.00	1234.0
92.00	1460.0
93.00	9784.0
93.90	752.0
96.00	638.0
97.90	7600.0
99.00	6619.0
99.90	486.0
101.00	3461.0
102.90	1006.0
103.90	1943.0
105.00	2087.0
105.90	634.0
107.00	26376.0
108.00	3484.0
108.90	599.0
109.90	42992.0
111.00	6947.0
111.90	945.0
115.90	1364.0
117.00	24816.0
117.90	1930.0
119.90	403.0
121.90	1635.0
122.90	2648.0
123.90	1163.0
124.90	1215.0
127.00	88616.0
128.00	6499.0
129.00	34656.0
129.90	3071.0
130.90	566.0
133.90	969.0
134.90	2831.0
135.90	1164.0
136.90	2007.0
137.90	394.0
139.90	369.0
140.90	4690.0
141.90	1504.0
142.90	1057.0
144.00	346.0
145.90	877.0
147.00	2525.0
147.90	6123.0
148.90	1466.0
150.00	378.0
151.10	572.0
151.80	307.0
152.90	1516.0
153.90	1236.0
155.00	2460.0
156.00	3905.0
157.00	638.0
157.90	800.0
158.90	640.0
159.90	1519.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

161.00	2434.0
161.90	662.0
163.90	311.0
164.90	1654.0
166.00	1377.0
166.90	10746.0
167.90	8221.0
169.00	1024.0
170.90	429.0
171.90	866.0
173.00	1262.0
174.00	2108.0
175.00	3922.0
176.00	1007.0
177.00	1682.0
177.90	580.0
178.90	6970.0
180.00	5223.0
180.90	2600.0
182.00	360.0
183.90	528.0
185.00	3607.0
186.00	27120.0
187.00	8026.0
187.90	885.0
189.00	1808.0
190.90	771.0
191.90	2422.0
193.00	2813.0
193.90	640.0
195.90	5240.0
196.80	885.0
197.90	176192.0
198.90	12481.0
199.90	888.0
201.30	622.0
203.00	1440.0
204.00	7229.0
205.00	12845.0
206.00	51744.0
207.00	7920.0
208.00	2016.0
209.00	772.0
210.00	751.0
211.00	2042.0
211.90	361.0
214.90	653.0
216.00	1206.0
216.90	14432.0
217.90	2043.0
221.00	7462.0
221.90	1961.0
223.00	3227.0
224.00	29480.0
225.00	7070.0
226.00	815.0
227.00	12715.0
227.90	1662.0
229.00	2448.0
229.90	359.0
231.00	1101.0
233.90	838.0
235.00	776.0
236.00	613.0
237.00	1355.0
238.90	464.0
239.90	534.0
240.90	735.0
242.00	1392.0
243.00	1708.0
244.00	24680.0
245.00	3137.0
245.90	4586.0
247.00	1032.0
249.00	729.0
252.90	494.0
253.90	649.0
255.00	118392.0
256.00	17984.0
256.90	1287.0
257.90	6964.0
259.00	1299.0
265.00	2799.0
266.00	518.0
273.00	3757.0
274.00	9055.0
275.00	54160.0
276.00	6746.0
277.00	4204.0
277.90	889.0
283.00	595.0
284.00	511.0
284.90	758.0
292.90	1109.0
294.00	342.0
295.00	306.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031

296.00	15894.0
297.00	2480.0
301.00	352.0
303.00	1704.0
304.00	496.0
314.00	617.0
315.00	1667.0
316.00	805.0
321.00	578.0
323.00	4969.0
324.00	967.0
327.00	973.0
327.90	468.0
332.00	406.0
333.00	473.0
334.00	3215.0
334.90	992.0
341.00	692.0
346.00	1439.0
352.00	1623.0
352.90	854.0
354.00	1699.0
354.90	303.0
364.90	7649.0
365.90	848.0
370.90	512.0
372.00	2490.0
372.90	594.0
382.90	802.0
390.00	361.0
402.00	1150.0
402.90	1351.0
404.00	575.0
421.00	1440.0
422.00	1556.0
423.00	10207.0
424.00	3668.0
424.90	337.0
441.00	31528.0
442.00	204672.0
443.00	38304.0
444.00	3730.0

Instrument :
BNA_N
ClientSampleId :
DFTPP031