

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041723\
 Data File : BN024951.D
 Acq On : 17 Apr 2023 12:19
 Operator : CG/JU
 Sample : 02340-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHE02

Quant Time: Apr 18 01:22:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 18 01:19:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	282003	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	1216226	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	783459	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	1721925	20.000	ng/u1	0.00
79) Chrysene-d12	21.521	240	1694278	20.000	ng/u1	0.00
88) Perylene-d12	23.957	264	1776269	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	29072	3.924	ng/uL	0.00
4) Pyridine-d5	3.740	84	133212	6.004	ng/u1	0.01
7) Phenol-d5	7.093	99	129070	4.695	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.257	67	444710	24.183	ng/u1	0.00
11) 2-Chlorophenol-d4	7.457	132	370438	18.334	ng/u1	0.00
15) 4-Methylphenol-d8	8.640	113	248598	11.390	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	243264	24.703	ng/u1	0.00
24) 2-Nitrophenol-d4	9.822	143	210005	19.688	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	403481	21.163	ng/u1	0.00
31) 4-Chloroaniline-d4	10.887	131	512131	17.701	ng/u1	0.00
46) Dimethylphthalate-d6	13.987	166	1614127	26.576	ng/u1	0.00
49) Acenaphthylene-d8	14.269	160	1754275	25.206	ng/u1	0.00
54) 4-Nitrophenol-d4	14.781	143	28315	2.274	ng/u1	0.00
60) Fluorene-d10	15.569	176	1447862	28.235	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	81649	7.286	ng/u1	0.00
73) Anthracene-d10	17.422	188	2611933	32.297	ng/u1	0.00
81) Pyrene-d10	19.722	212	3123270	31.746	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.804	264	2909620	31.328	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041723\
Data File : BN024951.D
Acq On : 17 Apr 2023 12:19
Operator : CG/JU
Sample : 02340-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BHE02

Quant Time: Apr 18 01:22:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 18 01:19:27 2023
Response via : Initial Calibration

