

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051225\
 Data File : BN036987.D
 Acq On : 12 May 2025 16:43
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: May 12 17:53:57 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 12 17:44:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.618	152	2281	0.400	ng	0.00
7) Naphthalene-d8	10.404	136	5775	0.400	ng	0.00
13) Acenaphthene-d10	14.267	164	3402	0.400	ng	0.00
19) Phenanthrene-d10	17.009	188	6663	0.400	ng	0.00
29) Chrysene-d12	21.207	240	6130	0.400	ng	# 0.00
35) Perylene-d12	23.412	264	5208	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.213	112	18486	3.144	ng	0.00
5) Phenol-d6	6.795	99	23568	3.306	ng	0.00
8) Nitrobenzene-d5	8.771	82	20517	3.371	ng	0.00
11) 2-Methylnaphthalene-d10	11.996	152	26844	3.310	ng	0.00
14) 2,4,6-Tribromophenol	15.767	330	5077	3.255	ng	0.00
15) 2-Fluorobiphenyl	12.889	172	49662	3.126	ng	0.00
27) Fluoranthene-d10	19.049	212	60220	3.327	ng	0.00
31) Terphenyl-d14	19.653	244	43834	3.159	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.141	88	8945	3.025	ng	98
3) n-Nitrosodimethylamine	3.451	42	18708	3.100	ng	98
6) bis(2-Chloroethyl)ether	7.055	93	21497	3.251	ng	99
9) Naphthalene	10.447	128	54123	3.226	ng	96
10) Hexachlorobutadiene	10.746	225	11147	3.097	ng	# 99
12) 2-Methylnaphthalene	12.072	142	36295	3.333	ng	98
16) Acenaphthylene	13.989	152	55484	3.337	ng	99
17) Acenaphthene	14.331	154	35245	3.228	ng	98
18) Fluorene	15.325	166	47676	3.257	ng	98
20) 4,6-Dinitro-2-methylph...	15.400	198	6294	3.694	ng	# 81
21) 4-Bromophenyl-phenylether	16.214	248	14539	3.360	ng	93
22) Hexachlorobenzene	16.326	284	14885	3.214	ng	99
23) Atrazine	16.487	200	13094	3.431	ng	92
24) Pentachlorophenol	16.674	266	9205	3.586	ng	97
25) Phenanthrene	17.058	178	72199	3.300	ng	99
26) Anthracene	17.145	178	68913	3.445	ng	100
28) Fluoranthene	19.082	202	87653	3.401	ng	99
30) Pyrene	19.444	202	88157	3.169	ng	100
32) Benzo(a)anthracene	21.189	228	75491	3.255	ng	99
33) Chrysene	21.242	228	78444	3.203	ng	98
34) Bis(2-ethylhexyl)phtha...	21.135	149	44488	3.115	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.620	276	68130	3.311	ng	96
37) Benzo(b)fluoranthene	22.749	252	72933	3.361	ng	# 89
38) Benzo(k)fluoranthene	22.793	252	72470	3.360	ng	# 90
39) Benzo(a)pyrene	23.313	252	60775	3.371	ng	# 83
40) Dibenzo(a,h)anthracene	25.634	278	54106	3.371	ng	# 90
41) Benzo(g,h,i)perylene	26.295	276	57030	3.211	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051225\
Data File : BN036987.D
Acq On : 12 May 2025 16:43
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Quant Time: May 12 17:53:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 12 17:44:09 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.618 min Scan# 61
 Delta R.T. -0.007 min
 Lab File: BN036987.D
 Acq: 12 May 2025 16:43

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2







Tgt Ion:152 Resp: 2281
 Ion Ratio Lower Upper
 152 100
 150 155.1 118.2 177.4
 115 70.1 52.5 78.7



#2
 1,4-Dioxane
 Concen: 3.025 ng
 RT: 3.141 min Scan# 36
 Delta R.T. -0.007 min
 Lab File: BN036987.D
 Acq: 12 May 2025 16:43

Tgt Ion: 88 Resp: 8945
 Ion Ratio Lower Upper
 88 100
 43 51.0 43.3 64.9
 58 87.8 70.7 106.1









#3
n-Nitrosodimethylamine
Concen: 3.100 ng
RT: 3.451 min Scan# 79
Delta R.T. -0.007 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 18708
Ion Ratio Lower Upper
42 100
74 70.7 55.2 82.8
44 5.6 4.7 7.1



#4
2-Fluorophenol
Concen: 3.144 ng
RT: 5.213 min Scan# 323
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion: 112 Resp: 18486
Ion Ratio Lower Upper
112 100
64 70.0 56.2 84.2
63 43.2 34.3 51.5









#5
Phenol-d6
Concen: 3.306 ng
RT: 6.795 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 23568
Ion Ratio Lower Upper
99 100
42 36.8 29.0 43.6
71 45.7 36.2 54.2



#6
bis(2-Chloroethyl)ether
Concen: 3.251 ng
RT: 7.055 min Scan# 578
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion: 93 Resp: 21497
Ion Ratio Lower Upper
93 100
63 88.3 69.6 104.4
95 31.9 25.1 37.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.404 min Scan# 948
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	5775
Ion	Ratio	Lower	Upper
136	100		
137	12.0	10.3	15.5
54	10.7	9.2	13.8
68	6.5	5.6	8.4



#8
Nitrobenzene-d5
Concen: 3.371 ng
RT: 8.771 min Scan# 795
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:	82	Resp:	20517
Ion	Ratio	Lower	Upper
82	100		
128	36.8	28.1	42.1
54	63.7	56.4	84.6









#9
Naphthalene
Concen: 3.226 ng
RT: 10.447 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 54123
Ion Ratio Lower Upper
128 100
129 10.9 10.0 15.0
127 13.7 12.1 18.1



#10
Hexachlorobutadiene
Concen: 3.097 ng
RT: 10.746 min Scan# 980
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:225 Resp: 11147
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.4 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 3.310 ng
RT: 11.996 min Scan# 1148
Delta R.T. -0.005 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 26844
Ion Ratio Lower Upper
152 100
151 21.4 17.4 26.2



#12
2-Methylnaphthalene
Concen: 3.333 ng
RT: 12.072 min Scan# 1163
Delta R.T. -0.005 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:142 Resp: 36295
Ion Ratio Lower Upper
142 100
141 91.6 71.8 107.8
115 47.3 38.6 58.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.267 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 3402
Ion Ratio Lower Upper
164 100
162 105.3 86.1 129.1
160 53.9 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 3.255 ng
RT: 15.767 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:330 Resp: 5077
Ion Ratio Lower Upper
330 100
332 95.7 75.6 113.4
141 63.9 47.4 71.2









#15
2-Fluorobiphenyl
Concen: 3.126 ng
RT: 12.889 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 49662
Ion Ratio Lower Upper
172 100
171 36.0 29.4 44.2
170 24.0 19.4 29.0



#16
Acenaphthylene
Concen: 3.337 ng
RT: 13.989 min Scan# 1439
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:152 Resp: 55484
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.4
153 12.9 10.9 16.3









#17
Acenaphthene
Concen: 3.228 ng
RT: 14.331 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 35245
Ion Ratio Lower Upper
154 100
153 116.3 94.6 142.0
152 59.8 49.4 74.2



#18
Fluorene
Concen: 3.257 ng
RT: 15.325 min Scan# 1564
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:166 Resp: 47676
Ion Ratio Lower Upper
166 100
165 99.4 81.0 121.4
167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.009 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 6663
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 16.5 13.4 20.0



#20
4,6-Dinitro-2-methylphenol
Concen: 3.694 ng
RT: 15.400 min Scan# 1571
Delta R.T. -0.011 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:198 Resp: 6294
Ion Ratio Lower Upper
198 100
51 79.5 79.7 119.5#
105 54.8 36.0 54.0#









#21
4-Bromophenyl-phenylether
Concen: 3.360 ng
RT: 16.214 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 14539
Ion Ratio Lower Upper
248 100
250 97.1 77.8 116.8
141 65.0 63.1 94.7



#22
Hexachlorobenzene
Concen: 3.214 ng
RT: 16.326 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:284 Resp: 14885
Ion Ratio Lower Upper
284 100
142 52.4 41.4 62.2
249 36.0 29.1 43.7









#23
 Atrazine
 Concen: 3.431 ng
 RT: 16.487 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: BN036987.D
 Acq: 12 May 2025 16:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.2	25.8	38.6
215	42.8	37.4	56.0



#24
 Pentachlorophenol
 Concen: 3.586 ng
 RT: 16.674 min Scan# 1675
 Delta R.T. 0.000 min
 Lab File: BN036987.D
 Acq: 12 May 2025 16:43

Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.0	52.8	79.2
268	63.6	50.0	75.0









#25
Phenanthrene
Concen: 3.300 ng
RT: 17.058 min Scan# 1706
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 72199
Ion Ratio Lower Upper
178 100
176 19.7 16.0 24.0
179 15.1 12.3 18.5



#26
Anthracene
Concen: 3.445 ng
RT: 17.145 min Scan# 1713
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:178 Resp: 68913
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 3.327 ng
RT: 19.049 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 60220
Ion Ratio Lower Upper
212 100
106 14.7 11.3 16.9
104 8.8 7.0 10.4



#28
Fluoranthene
Concen: 3.401 ng
RT: 19.082 min Scan# 1946
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:202 Resp: 87653
Ion Ratio Lower Upper
202 100
101 11.5 8.6 12.8
203 17.1 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.207 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 6130
Ion Ratio Lower Upper
240 100
120 13.0 14.5 21.7#
236 29.0 24.3 36.5



#30
Pyrene
Concen: 3.169 ng
RT: 19.444 min Scan# 2024
Delta R.T. 0.000 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Tgt Ion:202 Resp: 88157
Ion Ratio Lower Upper
202 100
200 21.6 17.3 25.9
203 17.7 14.4 21.6









#31
 Terphenyl-d14
 Concen: 3.159 ng
 RT: 19.653 min Scan# 2069
 Delta R.T. -0.005 min
 Lab File: BN036987.D
 Acq: 12 May 2025 16:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 43834
 Ion Ratio Lower Upper
 244 100
 212 10.8 9.5 14.3
 122 17.6 13.4 20.0



#32
 Benzo(a)anthracene
 Concen: 3.255 ng
 RT: 21.189 min Scan# 2277
 Delta R.T. -0.009 min
 Lab File: BN036987.D
 Acq: 12 May 2025 16:43

Tgt Ion:228 Resp: 75491
 Ion Ratio Lower Upper
 228 100
 226 28.4 22.4 33.6
 229 19.3 16.0 24.0





#33

Chrysene

Concen: 3.203 ng

RT: 21.242 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036987.D

Acq: 12 May 2025 16:43

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 78444

Ion Ratio Lower Upper

228 100

226 30.9 25.7 38.5

229 19.5 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 3.115 ng

RT: 21.135 min Scan# 2271

Delta R.T. 0.000 min

Lab File: BN036987.D

Acq: 12 May 2025 16:43




Tgt Ion:149 Resp: 44488


Ion Ratio Lower Upper

149 100

167 26.1 21.0 31.6

279 2.6 2.7 4.1#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.412 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036987.D

Acq: 12 May 2025 16:43

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 5208

Ion Ratio Lower Upper

264 100

260 26.4 22.3 33.5

265 65.2 55.4 83.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.311 ng

RT: 25.620 min Scan# 3512

Delta R.T. -0.006 min

Lab File: BN036987.D

Acq: 12 May 2025 16:43




Tgt Ion:276 Resp: 68130


Ion Ratio Lower Upper

276 100

138 30.7 21.8 32.8


277 25.1 20.2 30.4








#37
Benzo(b)fluoranthene
Concen: 3.361 ng
RT: 22.749 min Scan# 2530
Delta R.T. -0.006 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 72933
Ion Ratio Lower Upper
252 100
253 22.1 21.9 32.9
125 12.5 13.8 20.8#



#38
Benzo(k)fluoranthene
Concen: 3.360 ng
RT: 22.793 min Scan# 2545
Delta R.T. -0.003 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43


Tgt Ion:252 Resp: 72470
Ion Ratio Lower Upper
252 100
253 22.1 22.0 33.0
125 12.5 13.4 20.2#








#39
Benzo(a)pyrene
Concen: 3.371 ng
RT: 23.313 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 60775
Ion Ratio Lower Upper
252 100
253 22.2 24.8 37.2#
125 14.0 18.6 28.0#



#40
Dibenzo(a,h)anthracene
Concen: 3.371 ng
RT: 25.634 min Scan# 3517
Delta R.T. -0.003 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43


Tgt Ion:278 Resp: 54106
Ion Ratio Lower Upper
278 100
139 20.6 18.1 27.1
279 23.7 25.3 37.9#





#41
Benzo(g,h,i)perylene
Concen: 3.211 ng
RT: 26.295 min Scan# 31
Delta R.T. -0.003 min
Lab File: BN036987.D
Acq: 12 May 2025 16:43

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:	276	Resp:	57030
Ion	Ratio	Lower	Upper
276	100		
277	23.9	21.2	31.8
138	26.3	22.6	33.8

