

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014773.D
 Acq On : 26 May 2021 19:56
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: May 27 00:58:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 26 18:47:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	767	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	2601	0.40	ng	-0.01
13) Acenaphthene-d10	14.359	164	2102	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	5232	0.40	ng	0.00
29) Chrysene-d12	21.314	240	7952	0.40	ng	0.00
35) Perylene-d12	23.581	264	7334	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	6017	2.96	ng	0.00
5) Phenol-d6	6.895	99	8600	3.18	ng	0.00
8) Nitrobenzene-d5	8.857	82	7891	3.51	ng	-0.01
11) 2-Methylnaphthalene-d10	12.093	152	17120	2.70	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	4557	4.66	ng	-0.01
15) 2-Fluorobiphenyl	12.976	172	24792	3.22	ng	-0.01
27) Fluoranthene-d10	19.153	212	64519	2.76	ng	0.00
31) Terphenyl-d14	19.754	244	59322	3.16	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.271	88	2822	2.77	ng	# 91
3) n-Nitrosodimethylamine	3.609	42	1079	2.81	ng	# 8
6) bis(2-Chloroethyl)ether	7.144	93	7163	3.20	ng	98
9) Naphthalene	10.543	128	23961	3.49	ng	# 89
10) Hexachlorobutadiene	10.834	225	6142	3.71	ng	# 100
12) 2-Methylnaphthalene	12.168	142	18060	3.61	ng	97
16) Acenaphthylene	14.080	152	28870	3.73	ng	99
17) Acenaphthene	14.422	154	19962	3.49	ng	96
18) Fluorene	15.412	166	27758	3.60	ng	98
20) 4,6-Dinitro-2-methylph...	15.488	198	2804	4.69	ng	# 38
21) 4-Bromophenyl-phenylether	16.313	248	11852	3.70	ng	# 85
22) Hexachlorobenzene	16.422	284	12647	3.73	ng	99
23) Atrazine	16.581	200	8075	3.96	ng	# 88
24) Pentachlorophenol	16.775	266	3842	3.40	ng	88
25) Phenanthrene	17.153	178	49309	3.44	ng	99
26) Anthracene	17.250	178	45139	3.74	ng	97
28) Fluoranthene	19.183	202	69895	3.68	ng	99
30) Pyrene	19.546	202	70472	3.23	ng	99
32) Benzo(a)anthracene	21.292	228	79288	3.43	ng	99
33) Chrysene	21.345	228	83204	3.44	ng	98
34) Bis(2-ethylhexyl)phtha...	21.239	149	22310	3.47	ng	95
36) Indeno(1,2,3-cd)pyrene	25.861	276	107345	3.64	ng	100
37) Benzo(b)fluoranthene	22.897	252	89610	3.58	ng	# 92
38) Benzo(k)fluoranthene	22.941	252	93993	3.63	ng	# 89
39) Benzo(a)pyrene	23.479	252	81270	3.84	ng	# 85
40) Dibenzo(a,h)anthracene	25.875	278	90772	3.60	ng	# 92
41) Benzo(g,h,i)perylene	26.556	276	90832	3.78	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014773.D
 Acq On : 26 May 2021 19:56
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: May 27 00:58:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 26 18:47:42 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.716 min Scan# 613
 Delta R.T. 0.000 min
 Lab File: BN014773.D
 Acq: 26 May 2021 19:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC3.2


Tgt Ion:152 Resp: 767
 Ion Ratio Lower Upper
 152 100
 150 143.2 123.4 185.2
 115 78.8 73.2 109.8



#2
 1,4-Dioxane
 Concen: 2.77 ng
 RT: 3.271 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN014773.D
 Acq: 26 May 2021 19:56

Tgt Ion: 88 Resp: 2822
 Ion Ratio Lower Upper
 88 100
 58 53.0 44.7 67.1
 43 16.1 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 2.81 ng
RT: 3.609 min Scan# 103
Delta R.T. -0.032 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 1079
Ion Ratio Lower Upper
42 100
74 199.4 0.0 0.0#
44 7.6 80.0 120.0#



#4
2-Fluorophenol
Concen: 2.96 ng
RT: 5.316 min Scan# 315
Delta R.T. -0.008 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion: 112 Resp: 6017
Ion Ratio Lower Upper
112 100
64 55.0 43.6 65.4
63 29.1 25.0 37.6









#5
Phenol-d6
Concen: 3.18 ng
RT: 6.895 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	99	Resp:	8600
Ion	Ratio	Lower	Upper
99	100		
42	11.6	10.2	15.2
71	44.8	32.4	48.6



#6
bis(2-Chloroethyl)ether
Concen: 3.20 ng
RT: 7.144 min Scan# 542
Delta R.T. -0.008 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:	93	Resp:	7163
Ion	Ratio	Lower	Upper
93	100		
63	59.8	47.5	71.3
95	33.5	28.5	42.7









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.013 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	2601
Ion	Ratio	Lower	Upper
136	100		
137	14.2	13.2	19.8
54	11.2	9.0	13.4
68	9.4	8.8	13.2



#8
Nitrobenzene-d5
Concen: 3.51 ng
RT: 8.857 min Scan# 730
Delta R.T. -0.013 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:	82	Resp:	7891
Ion	Ratio	Lower	Upper
82	100		
128	27.5	34.2	51.2#
54	41.0	38.4	57.6









#9
Naphthalene
Concen: 3.49 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.013 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 23961
Ion Ratio Lower Upper
128 100
129 11.3 13.0 19.4#
127 14.3 15.0 22.6#



#10
Hexachlorobutadiene
Concen: 3.71 ng
RT: 10.834 min Scan# 886
Delta R.T. -0.013 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:225 Resp: 6142
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 50.8 76.2









#11
2-Methylnaphthalene-d10
Concen: 2.70 ng
RT: 12.093 min Scan# 1071
Delta R.T. -0.009 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 17120
Ion Ratio Lower Upper
152 100
151 20.9 16.8 25.2



#12
2-Methylnaphthalene
Concen: 3.61 ng
RT: 12.168 min Scan# 1088
Delta R.T. -0.005 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:142 Resp: 18060
Ion Ratio Lower Upper
142 100
141 88.4 72.9 109.3
115 49.1 41.8 62.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. -0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	164	Resp:	2102
Ion	Ratio	Lower	Upper
164	100		
162	103.4	83.2	124.8
160	47.7	41.8	62.6



#14
2,4,6-Tribromophenol
Concen: 4.66 ng
RT: 15.856 min Scan# 1428
Delta R.T. -0.013 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:	330	Resp:	4557
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	26.2	21.8	32.6









#15
2-Fluorobiphenyl
Concen: 3.22 ng
RT: 12.976 min Scan# 1201
Delta R.T. -0.013 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 24792
Ion Ratio Lower Upper
172 100
171 36.6 30.9 46.3
170 24.2 21.4 32.0



#16
Acenaphthylene
Concen: 3.73 ng
RT: 14.080 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:152 Resp: 28870
Ion Ratio Lower Upper
152 100
151 20.9 17.1 25.7
153 13.1 9.9 14.9









#17
Acenaphthene
Concen: 3.49 ng
RT: 14.422 min Scan# 1315
Delta R.T. -0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 19962
Ion Ratio Lower Upper
154 100
153 110.8 91.8 137.8
152 53.6 46.0 69.0



#18
Fluorene
Concen: 3.60 ng
RT: 15.412 min Scan# 1393
Delta R.T. -0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:166 Resp: 27758
Ion Ratio Lower Upper
166 100
165 98.3 79.9 119.9
167 13.7 10.5 15.7









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1531
Delta R.T. 0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 5232
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.4 8.2 12.4



#20
4,6-Dinitro-2-methylphenol
Concen: 4.69 ng
RT: 15.488 min Scan# 1399
Delta R.T. -0.026 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:198 Resp: 2804
Ion Ratio Lower Upper
198 100
51 35.6 96.4 144.6#
105 49.6 72.3 108.5#









#21
4-Bromophenyl-phenylether
Concen: 3.70 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 248 Resp: 11852
Ion Ratio Lower Upper
248 100
250 103.2 77.5 116.3
141 41.5 50.2 75.4#



#22
Hexachlorobenzene
Concen: 3.73 ng
RT: 16.422 min Scan# 1474
Delta R.T. 0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion: 284 Resp: 12647
Ion Ratio Lower Upper
284 100
142 26.6 20.3 30.5
249 32.8 26.4 39.6





#23

Atrazine

Concen: 3.96 ng

RT: 16.581 min Scan# 1487

Delta R.T. 0.000 min

Lab File: BN014773.D

Acq: 26 May 2021 19:56





Tgt Ion:	200	Resp:	8075
Ion Ratio	Lower	Upper	
200	100		
173	24.8	29.1	43.7#
215	51.5	45.1	67.7

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2



#24

Pentachlorophenol

Concen: 3.40 ng




RT: 16.775 min Scan# 1503


Delta R.T. 0.000 min

Lab File: BN014773.D

Acq: 26 May 2021 19:56

Tgt Ion:	266	Resp:	3842
Ion Ratio	Lower	Upper	
266	100		
264	63.1	44.2	66.2
268	65.1	44.2	66.2









#25
Phenanthrene
Concen: 3.44 ng
RT: 17.153 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 49309
Ion Ratio Lower Upper
178 100
176 19.4 16.0 24.0
179 15.4 12.6 19.0



#26
Anthracene
Concen: 3.74 ng
RT: 17.250 min Scan# 1542
Delta R.T. 0.000 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:178 Resp: 45139
Ion Ratio Lower Upper
178 100
176 19.0 14.2 21.2
179 15.3 13.7 20.5





#27

Fluoranthene-d10

Concen: 2.76 ng

RT: 19.153 min Scan# 1834

Delta R.T. -0.002 min

Lab File: BN014773.D

Acq: 26 May 2021 19:56





Tgt Ion:212	Resp:	64519
Ion Ratio	Lower	Upper
212	100	
106	9.4	7.4 11.2
104	5.5	4.1 6.1

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2



#28

Fluoranthene

Concen: 3.68 ng




RT: 19.183 min Scan# 1840


Delta R.T. -0.002 min

Lab File: BN014773.D

Acq: 26 May 2021 19:56

Tgt Ion:202	Resp:	69895
Ion Ratio	Lower	Upper
202	100	
101	8.9	7.0 10.6
203	17.4	13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.314 min Scan# 2129
Delta R.T. -0.008 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240	Resp:	7952
Ion Ratio	Lower	Upper
240	100	
120	6.3	5.2 7.8
236	27.0	21.6 32.4



#30
Pyrene
Concen: 3.23 ng
RT: 19.546 min Scan# 1913
Delta R.T. -0.007 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:202	Resp:	70472
Ion Ratio	Lower	Upper
202	100	
200	21.1	16.6 24.8
203	17.5	13.9 20.9









#31
 Terphenyl-d14
 Concen: 3.16 ng
 RT: 19.754 min Scan# 1955
 Delta R.T. -0.007 min
 Lab File: BN014773.D
 Acq: 26 May 2021 19:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244	Resp:	59322
Ion Ratio	Lower	Upper
244	100	
212	7.3	6.8 10.2
122	7.4	6.6 9.8



#32
 Benzo(a)anthracene
 Concen: 3.43 ng
 RT: 21.292 min Scan# 2127
 Delta R.T. -0.008 min
 Lab File: BN014773.D
 Acq: 26 May 2021 19:56

Tgt Ion:228	Resp:	79288
Ion Ratio	Lower	Upper
228	100	
226	28.2	22.9 34.3
229	19.2	16.2 24.2









#33
Chrysene
Concen: 3.44 ng
RT: 21.345 min Scan# 2132
Delta R.T. -0.008 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228	Resp:	83204
Ion Ratio	Lower	Upper
228	100	
226	30.9	24.3 36.5
229	19.4	16.4 24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.47 ng
RT: 21.239 min Scan# 2122
Delta R.T. 0.003 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:149	Resp:	22310
Ion Ratio	Lower	Upper
149	100	
167	38.4	28.3 42.5
279	7.3	5.5 8.3









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.581 min Scan# 2575
Delta R.T. -0.004 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:264 Resp: 7334




Ion	Ratio	Lower	Upper
264	100		
260	26.4	21.4	32.2
265	42.8	40.9	61.3




#36
Indeno(1,2,3-cd)pyrene
Concen: 3.64 ng
RT: 25.861 min Scan# 3241
Delta R.T. -0.008 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:276 Resp: 107345

Ion	Ratio	Lower	Upper
276	100		
138	12.5	10.2	15.4
277	25.6	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 3.58 ng
RT: 22.897 min Scan# 2375
Delta R.T. -0.004 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 89610
Ion Ratio Lower Upper
252 100
253 22.5 21.5 32.3
125 6.0 6.6 10.0#



#38
Benzo(k)fluoranthene
Concen: 3.63 ng
RT: 22.941 min Scan# 2388
Delta R.T. -0.008 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:252 Resp: 93993
Ion Ratio Lower Upper
252 100
253 22.1 22.7 34.1#
125 5.7 6.4 9.6#









#39
Benzo(a)pyrene
Concen: 3.84 ng
RT: 23.479 min Scan# 2545
Delta R.T. -0.008 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 81270
Ion Ratio Lower Upper
252 100
253 22.5 25.1 37.7#
125 7.2 9.4 14.0#



#40
Dibenzo(a,h)anthracene
Concen: 3.60 ng
RT: 25.875 min Scan# 3245
Delta R.T. -0.011 min
Lab File: BN014773.D
Acq: 26 May 2021 19:56

Tgt Ion:278 Resp: 90772
Ion Ratio Lower Upper
278 100
139 8.3 8.6 13.0#
279 23.7 22.8 34.2





#41
 Benzo(g,h,i)perylene
 Concen: 3.78 ng
 RT: 26.556 min Scan# 3444
 Delta R.T. -0.011 min
 Lab File: BN014773.D
 Acq: 26 May 2021 19:56

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	276	Resp:	90832
Ion Ratio	Lower	Upper	
276	100		
277	23.8	20.9	31.3
138	10.7	10.3	15.5

