

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014775.D
 Acq On : 26 May 2021 21:44
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN052621

Quant Time: May 27 01:53:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 01:39:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	755	0.40	ng	0.00
7) Naphthalene-d8	10.505	136	2674	0.40	ng	# 0.00
13) Acenaphthene-d10	14.359	164	2073	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	5462	0.40	ng	0.00
29) Chrysene-d12	21.314	240	7650	0.40	ng	0.00
35) Perylene-d12	23.581	264	7471	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	951	0.47	ng	0.00
5) Phenol-d6	6.895	99	1216	0.44	ng	0.00
8) Nitrobenzene-d5	8.870	82	900	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.097	152	2141	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	454	0.38	ng	-0.01
15) 2-Fluorobiphenyl	12.976	172	3261	0.43	ng	-0.01
27) Fluoranthene-d10	19.153	212	8375	0.39	ng	0.00
31) Terphenyl-d14	19.759	244	7756	0.43	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.271	88	381	0.43	ng	# 94
3) n-Nitrosodimethylamine	3.649	42	60	0.36	ng	# 13
6) bis(2-Chloroethyl)ether	7.153	93	905	0.39	ng	99
9) Naphthalene	10.543	128	3171	0.41	ng	94
10) Hexachlorobutadiene	10.834	225	847	0.42	ng	# 97
12) 2-Methylnaphthalene	12.173	142	2275	0.40	ng	99
16) Acenaphthylene	14.080	152	3263	0.39	ng	98
17) Acenaphthene	14.422	154	2463	0.41	ng	99
18) Fluorene	15.412	166	3571	0.42	ng	97
20) 4,6-Dinitro-2-methylph...	15.501	198	154	0.34	ng	# 80
21) 4-Bromophenyl-phenylether	16.313	248	1490	0.40	ng	91
22) Hexachlorobenzene	16.422	284	1698	0.41	ng	99
23) Atrazine	16.581	200	871	0.36	ng	98
24) Pentachlorophenol	16.775	266	264	0.30	ng	# 90
25) Phenanthrene	17.153	178	6555	0.40	ng	99
26) Anthracene	17.250	178	5326	0.39	ng	97
28) Fluoranthene	19.183	202	8930	0.39	ng	100
30) Pyrene	19.546	202	9129	0.42	ng	99
32) Benzo(a)anthracene	21.303	228	9009	0.39	ng	98
33) Chrysene	21.345	228	11198	0.44	ng	98
34) Bis(2-ethylhexyl)phtha...	21.229	149	2601	0.39	ng	98
36) Indeno(1,2,3-cd)pyrene	25.865	276	11450	0.37	ng	98
37) Benzo(b)fluoranthene	22.941	252	11989	0.44	ng	97
38) Benzo(k)fluoranthene	22.941	252	11989	0.42	ng	95
39) Benzo(a)pyrene	23.482	252	10095	0.41	ng	# 93
40) Dibenzo(a,h)anthracene	25.885	278	9225	0.36	ng	97
41) Benzo(g,h,i)perylene	26.559	276	10252	0.37	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014775.D
 Acq On : 26 May 2021 21:44
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN052621

Quant Time: May 27 01:53:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 01:39:07 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.716 min Scan# 613
 Delta R.T. 0.000 min
 Lab File: BN014775.D
 Acq: 26 May 2021 21:44




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN052621


Tgt Ion:152 Resp: 755
 Ion Ratio Lower Upper
 152 100
 150 146.2 123.4 185.2
 115 83.7 73.2 109.8



#2
 1,4-Dioxane
 Concen: 0.43 ng
 RT: 3.271 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN014775.D
 Acq: 26 May 2021 21:44

Tgt Ion: 88 Resp: 381
 Ion Ratio Lower Upper
 88 100
 58 58.0 44.7 67.1
 43 20.5 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 0.36 ng
RT: 3.649 min Scan# 108
Delta R.T. 0.008 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion: 42 Resp: 60
Ion Ratio Lower Upper
42 100
74 105.0 0.0 0.0#
44 13.3 80.0 120.0#



#4
2-Fluorophenol
Concen: 0.47 ng
RT: 5.316 min Scan# 315
Delta R.T. -0.008 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion: 112 Resp: 951
Ion Ratio Lower Upper
112 100
64 56.7 43.6 65.4
63 29.4 25.0 37.6









#5
Phenol-d6
Concen: 0.44 ng
RT: 6.895 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:	99	Resp:	1216
Ion	Ratio	Lower	Upper
99	100		
42	12.3	10.2	15.2
71	46.6	32.4	48.6



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.153 min Scan# 543
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:	93	Resp:	905
Ion	Ratio	Lower	Upper
93	100		
63	59.9	47.5	71.3
95	36.4	28.5	42.7









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.505 min Scan# 860
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:	136	Resp:	2674
Ion	Ratio	Lower	Upper
136	100		
137	13.2	13.2	19.8
54	7.2	9.0	13.4#
68	7.3	8.8	13.2#



#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.870 min Scan# 731
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:	82	Resp:	900
Ion	Ratio	Lower	Upper
82	100		
128	40.6	34.2	51.2
54	40.4	38.4	57.6









#9
Naphthalene
Concen: 0.41 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.013 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:	128	Resp:	3171
Ion Ratio	Lower	Upper	
128	100		
129	13.7	13.0	19.4
127	16.4	15.0	22.6



#10
Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.834 min Scan# 886
Delta R.T. -0.013 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:	225	Resp:	847
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	65.5	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.097 min Scan# 1072
Delta R.T. -0.005 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:152 Resp: 2141
Ion Ratio Lower Upper
152 100
151 21.3 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.173 min Scan# 1089
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:142 Resp: 2275
Ion Ratio Lower Upper
142 100
141 89.5 72.9 109.3
115 51.5 41.8 62.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:	164	Resp:	2073
Ion	Ratio	Lower	Upper
164	100		
162	107.3	83.2	124.8
160	49.2	41.8	62.6



#14
2,4,6-Tribromophenol
Concen: 0.38 ng
RT: 15.856 min Scan# 1428
Delta R.T. -0.013 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:	330	Resp:	454
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	27.5	21.8	32.6









#15
2-Fluorobiphenyl
Concen: 0.43 ng
RT: 12.976 min Scan# 1201
Delta R.T. -0.013 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:	172	Resp:	3261
Ion	Ratio	Lower	Upper
172	100		
171	37.3	30.9	46.3
170	26.5	21.4	32.0



#16
Acenaphthylene
Concen: 0.39 ng
RT: 14.080 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:	152	Resp:	3263
Ion	Ratio	Lower	Upper
152	100		
151	20.2	17.1	25.7
153	13.0	9.9	14.9









#17
Acenaphthene
Concen: 0.41 ng
RT: 14.422 min Scan# 1315
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:154 Resp: 2463
Ion Ratio Lower Upper
154 100
153 116.2 91.8 137.8
152 56.6 46.0 69.0



#18
Fluorene
Concen: 0.42 ng
RT: 15.412 min Scan# 1393
Delta R.T. -0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:166 Resp: 3571
Ion Ratio Lower Upper
166 100
165 97.0 79.9 119.9
167 13.4 10.5 15.7









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1531
Delta R.T. 0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:188 Resp: 5462
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.7 8.2 12.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.34 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.012 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:198 Resp: 154
Ion Ratio Lower Upper
198 100
51 98.1 96.4 144.6
105 71.2 72.3 108.5#









#21
4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.001 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:	248	Resp:	1490
Ion	Ratio	Lower	Upper
248	100		
250	104.4	77.5	116.3
141	53.6	50.2	75.4



#22
Hexachlorobenzene
Concen: 0.41 ng
RT: 16.422 min Scan# 1474
Delta R.T. 0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:	284	Resp:	1698
Ion	Ratio	Lower	Upper
284	100		
142	26.7	20.3	30.5
249	32.9	26.4	39.6









#23
Atrazine
Concen: 0.36 ng
RT: 16.581 min Scan# 1487
Delta R.T. 0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:	200	Resp:	871
Ion Ratio	Lower	Upper	
200	100		
173	33.6	29.1	43.7
215	56.1	45.1	67.7



#24
Pentachlorophenol
Concen: 0.30 ng
RT: 16.775 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:	266	Resp:	264
Ion Ratio	Lower	Upper	
266	100		
264	58.7	44.2	66.2
268	66.3	44.2	66.2#









#25
Phenanthrene
Concen: 0.40 ng
RT: 17.153 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:178 Resp: 6555
Ion Ratio Lower Upper
178 100
176 19.2 16.0 24.0
179 16.2 12.6 19.0



#26
Anthracene
Concen: 0.39 ng
RT: 17.250 min Scan# 1542
Delta R.T. 0.000 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:178 Resp: 5326
Ion Ratio Lower Upper
178 100
176 19.0 14.2 21.2
179 15.8 13.7 20.5





#27

Fluoranthene-d10

Concen: 0.39 ng





RT: 19.153 min Scan# 1834

Delta R.T. -0.002 min

Lab File: BN014775.D

Acq: 26 May 2021 21:44

Tgt Ion: 212	Resp: 8375
Ion Ratio	Lower Upper
212 100	
106 9.6	7.4 11.2
104 5.3	4.1 6.1



#28

Fluoranthene

Concen: 0.39 ng




RT: 19.183 min Scan# 1840


Delta R.T. -0.002 min

Lab File: BN014775.D

Acq: 26 May 2021 21:44

Tgt Ion: 202	Resp: 8930
Ion Ratio	Lower Upper
202 100	
101 8.8	7.0 10.6
203 17.2	13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.314 min Scan# 2129
Delta R.T. -0.008 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:240 Resp: 7650
Ion Ratio Lower Upper
240 100
120 7.4 5.2 7.8
236 27.7 21.6 32.4



#30
Pyrene
Concen: 0.42 ng
RT: 19.546 min Scan# 1913
Delta R.T. -0.007 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:202 Resp: 9129
Ion Ratio Lower Upper
202 100
200 21.2 16.6 24.8
203 17.8 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.43 ng
 RT: 19.759 min Scan# 1956
 Delta R.T. -0.002 min
 Lab File: BN014775.D
 Acq: 26 May 2021 21:44




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN052621


Tgt Ion:244	Resp:	7756
Ion Ratio	Lower	Upper
244	100	
212	8.0	6.8 10.2
122	8.2	6.6 9.8



#32
 Benzo(a)anthracene
 Concen: 0.39 ng
 RT: 21.303 min Scan# 2128
 Delta R.T. 0.003 min
 Lab File: BN014775.D
 Acq: 26 May 2021 21:44

Tgt Ion:228	Resp:	9009
Ion Ratio	Lower	Upper
228	100	
226	27.5	22.9 34.3
229	21.4	16.2 24.2









#33
Chrysene
Concen: 0.44 ng
RT: 21.345 min Scan# 2132
Delta R.T. -0.008 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:	228	Resp:	11198
Ion Ratio	Lower	Upper	
228	100		
226	31.6	24.3	36.5
229	19.7	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.39 ng
RT: 21.229 min Scan# 2121
Delta R.T. -0.008 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:	149	Resp:	2601
Ion Ratio	Lower	Upper	
149	100		
167	36.9	28.3	42.5
279	6.8	5.5	8.3









#35
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.581 min Scan# 2575
 Delta R.T. -0.004 min
 Lab File: BN014775.D
 Acq: 26 May 2021 21:44




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN052621


Tgt Ion:264	Resp:	7471
Ion Ratio	Lower	Upper
264	100	
260	26.2	21.4 32.2
265	45.1	40.9 61.3



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.37 ng
 RT: 25.865 min Scan# 3242
 Delta R.T. -0.004 min
 Lab File: BN014775.D
 Acq: 26 May 2021 21:44

Tgt Ion:276	Resp:	11450
Ion Ratio	Lower	Upper
276	100	
138	11.1	10.2 15.4
277	25.3	20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.44 ng
RT: 22.941 min Scan# 2388
Delta R.T. 0.040 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:252 Resp: 11989
Ion Ratio Lower Upper
252 100
253 25.3 21.5 32.3
125 7.3 6.6 10.0



#38
Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 22.941 min Scan# 2388
Delta R.T. -0.008 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:252 Resp: 11989
Ion Ratio Lower Upper
252 100
253 25.3 22.7 34.1
125 7.3 6.4 9.6









#39
Benzo(a)pyrene
Concen: 0.41 ng
RT: 23.482 min Scan# 2546
Delta R.T. -0.004 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44




Instrument :
BNA_N
ClientSampleId :
ICVBN052621


Tgt Ion:252 Resp: 10095
Ion Ratio Lower Upper
252 100
253 29.0 25.1 37.7
125 17.4 9.4 14.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.885 min Scan# 3248
Delta R.T. -0.001 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Tgt Ion:278 Resp: 9225
Ion Ratio Lower Upper
278 100
139 10.0 8.6 13.0
279 26.9 22.8 34.2





#41
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 26.559 min Scan# 3445
Delta R.T. -0.008 min
Lab File: BN014775.D
Acq: 26 May 2021 21:44

Instrument :
BNA_N
ClientSampleId :
ICVBN052621

Tgt Ion:	276	Resp:	10252
Ion Ratio	Lower	Upper	
276	100		
277	23.9	20.9	31.3
138	12.9	10.3	15.5

