

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014778.D
 Acq On : 26 May 2021 23:32
 Operator : CG/JU
 Sample : M2411-06
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW8-GW-798-800

Quant Time: May 27 06:55:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 01:39:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	926	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	3311	0.40	ng	-0.01
13) Acenaphthene-d10	14.359	164	2541	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	6606	0.40	ng	# 0.00
29) Chrysene-d12	21.311	240	9330	0.40	ng	-0.01
35) Perylene-d12	23.576	264	8539	0.40	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.325	112	577	0.23	ng	0.00
5) Phenol-d6	6.895	99	591	0.17	ng	0.00
8) Nitrobenzene-d5	8.870	82	929	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.098	152	1872	0.28	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	424	0.29	ng	-0.01
15) 2-Fluorobiphenyl	12.976	172	3178	0.34	ng	-0.01
27) Fluoranthene-d10	19.151	212	7962	0.30	ng	0.00
31) Terphenyl-d14	19.757	244	9540	0.43	ng	0.00
Target Compounds						
25) Phenanthrene	17.152	178	459	0.02	ng	# 86
34) Bis(2-ethylhexyl)phtha...	21.237	149	4802	0.60	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
Data File : BN014778.D
Acq On : 26 May 2021 23:32
Operator : CG/JU
Sample : M2411-06
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW8-GW-798-800

Quant Time: May 27 06:55:46 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 01:39:07 2021
Response via : Initial Calibration

