

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014775.D
 Acq On : 26 May 2021 21:44
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN052621

Quant Time: May 27 01:56:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 01:39:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	194	0.00
2	1,4-Dioxane	0.400	0.433	-8.2	228	0.00
3	n-Nitrosodimethylamine	0.400	0.364	9.0	207	0.00
4 S	2-Fluorophenol	0.400	0.471	-17.7	201	0.00
5 S	Phenol-d6	0.400	0.440	-10.0	186	0.00
6	bis(2-Chloroethyl)ether	0.400	0.389	2.8	181	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	180	0.00
8 S	Nitrobenzene-d5	0.400	0.370	7.5	165	0.00
9	Naphthalene	0.400	0.412	-3.0	179	-0.01
10	Hexachlorobutadiene	0.400	0.420	-5.0	183	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.396	1.0	174	0.00
12	2-Methylnaphthalene	0.400	0.403	-0.8	179	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	174	0.00
14 S	2,4,6-Tribromophenol	0.400	0.379	5.3	179	-0.01
15 S	2-Fluorobiphenyl	0.400	0.428	-7.0	182	-0.01
16	Acenaphthylene	0.400	0.393	1.8	175	0.00
17	Acenaphthene	0.400	0.406	-1.5	172	0.00
18	Fluorene	0.400	0.416	-4.0	172	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	165	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.340	15.0	169	-0.01
21	4-Bromophenyl-phenylether	0.400	0.396	1.0	169	0.00
22	Hexachlorobenzene	0.400	0.414	-3.5	167	0.00
23	Atrazine	0.400	0.359	10.3	162	0.00
24	Pentachlorophenol	0.400	0.299	25.3#	171	0.00
25	Phenanthrene	0.400	0.399	0.3	166	0.00
26	Anthracene	0.400	0.385	3.8	164	0.00
27 SURR	Fluoranthene-d10	0.400	0.387	3.3	152	0.00
28	Fluoranthene	0.400	0.390	2.5	150	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	141	0.00
30	Pyrene	0.400	0.416	-4.0	149	0.00
31 S	Terphenyl-d14	0.400	0.428	-7.0	149	0.00
32	Benzo(a)anthracene	0.400	0.387	3.3	138	0.00
33	Chrysene	0.400	0.437	-9.2	148	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.393	1.8	136	0.00
35 I	Perylene-d12	0.400	0.400	0.0	137	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	123	0.00
37	Benzo(b)fluoranthene	0.400	0.422	-5.5	141	0.00
38	Benzo(k)fluoranthene	0.400	0.416	-4.0	136	0.00
39 C	Benzo(a)pyrene	0.400	0.414	-3.5	140	0.00
40	Dibenzo(a,h)anthracene	0.400	0.360	10.0	120	0.00
41	Benzo(g,h,i)perylene	0.400	0.372	7.0	117	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN052621