

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014783.D
 Acq On : 27 May 2021 03:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 27 07:14:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 01:39:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	158	0.00
2	1,4-Dioxane	0.400	0.506	-26.5#	218	0.00
3	n-Nitrosodimethylamine	0.400	0.587	-46.7#	400	0.02
4 S	2-Fluorophenol	0.400	0.523	-30.8#	182	0.00
5 S	Phenol-d6	0.400	0.515	-28.7#	178	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	153	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	158	-0.01
8 S	Nitrobenzene-d5	0.400	0.403	-0.8	157	0.00
9	Naphthalene	0.400	0.418	-4.5	160	-0.01
10	Hexachlorobutadiene	0.400	0.413	-3.2	158	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.402	-0.5	155	0.00
12	2-Methylnaphthalene	0.400	0.416	-4.0	162	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	156	0.00
14 S	2,4,6-Tribromophenol	0.400	0.495	-23.7	210	-0.01
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	157	-0.01
16	Acenaphthylene	0.400	0.413	-3.2	165	0.00
17	Acenaphthene	0.400	0.408	-2.0	155	0.00
18	Fluorene	0.400	0.413	-3.2	154	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	150	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.406	-1.5	189	-0.01
21	4-Bromophenyl-phenylether	0.400	0.390	2.5	151	0.00
22	Hexachlorobenzene	0.400	0.420	-5.0	154	0.00
23	Atrazine	0.400	0.390	2.5	160	0.00
24	Pentachlorophenol	0.400	0.469	-17.2	244	0.00
25	Phenanthrene	0.400	0.399	0.3	151	0.00
26	Anthracene	0.400	0.402	-0.5	155	0.00
27 SURR	Fluoranthene-d10	0.400	0.396	1.0	141	0.00
28	Fluoranthene	0.400	0.401	-0.3	140	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	133	0.00
30	Pyrene	0.400	0.415	-3.7	140	0.00
31 S	Terphenyl-d14	0.400	0.414	-3.5	136	0.00
32	Benzo(a)anthracene	0.400	0.425	-6.2	143	0.00
33	Chrysene	0.400	0.422	-5.5	135	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.443	-10.7	144	0.00
35 I	Perylene-d12	0.400	0.400	0.0	136	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.374	6.5	123	-0.01
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	135	0.00
38	Benzo(k)fluoranthene	0.400	0.401	-0.3	130	-0.01
39 C	Benzo(a)pyrene	0.400	0.423	-5.7	142	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.363	9.3	120	-0.01
41	Benzo(g,h,i)perylene	0.400	0.373	6.8	116	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4