

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014768.D
 Acq On : 26 May 2021 16:56
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: May 26 18:50:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 26 18:47:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	923	0.40	ng	0.00
7) Naphthalene-d8	10.505	136	3338	0.40	ng	# 0.00
13) Acenaphthene-d10	14.359	164	2626	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	6860	0.40	ng	0.00
29) Chrysene-d12	21.324	240	9743	0.40	ng	0.00
35) Perylene-d12	23.589	264	9754	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.317	112	280	0.11	ng	0.00
5) Phenol-d6	6.895	99	377	0.12	ng	0.00
8) Nitrobenzene-d5	8.870	82	335	0.12	ng	0.00
11) 2-Methylnaphthalene-d10	12.098	152	695	0.09	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	154	0.13	ng	-0.01
15) 2-Fluorobiphenyl	12.976	172	989	0.10	ng	-0.01
27) Fluoranthene-d10	19.154	212	2847	0.09	ng	0.00
31) Terphenyl-d14	19.759	244	2373	0.10	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.279	88	120	0.10	ng	# 88
6) bis(2-Chloroethyl)ether	7.145	93	317	0.12	ng	97
9) Naphthalene	10.543	128	976	0.11	ng	92
10) Hexachlorobutadiene	10.835	225	264	0.12	ng	# 98
12) 2-Methylnaphthalene	12.173	142	725	0.11	ng	99
16) Acenaphthylene	14.080	152	1024	0.11	ng	97
17) Acenaphthene	14.422	154	785	0.11	ng	99
18) Fluorene	15.412	166	1116	0.12	ng	# 96
20) 4,6-Dinitro-2-methylph...	15.501	198	73	0.09	ng	86
21) 4-Bromophenyl-phenylether	16.313	248	489	0.12	ng	97
22) Hexachlorobenzene	16.423	284	518	0.12	ng	93
23) Atrazine	16.581	200	318	0.12	ng	91
25) Phenanthrene	17.153	178	2273	0.12	ng	99
26) Anthracene	17.250	178	1759	0.11	ng	97
28) Fluoranthene	19.183	202	2877	0.12	ng	100
30) Pyrene	19.551	202	3138	0.12	ng	99
32) Benzo(a)anthracene	21.303	228	3060	0.11	ng	96
33) Chrysene	21.356	228	3234	0.11	ng	99
36) Indeno(1,2,3-cd)pyrene	25.878	276	3385	0.09	ng	98
37) Benzo(b)fluoranthene	22.907	252	3421	0.10	ng	# 74
38) Benzo(k)fluoranthene	22.952	252	3580	0.10	ng	# 78
39) Benzo(a)pyrene	23.493	252	3003	0.11	ng	# 65
40) Dibenzo(a,h)anthracene	25.899	278	2715	0.08	ng	# 84
41) Benzo(g,h,i)perylene	26.577	276	3198	0.10	ng	# 93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
Data File : BN014768.D
Acq On : 26 May 2021 16:56
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Quant Time: May 26 18:50:27 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 26 18:47:42 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.717 min Scan# 613
Delta R.T. 0.001 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 923
Ion Ratio Lower Upper
152 100
150 147.4 123.4 185.2
115 82.8 73.2 109.8



#2
1,4-Dioxane
Concen: 0.10 ng
RT: 3.279 min Scan# 62
Delta R.T. 0.008 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion: 88 Resp: 120
Ion Ratio Lower Upper
88 100
58 56.7 44.7 67.1
43 45.0 21.0 31.6#









#4
2-Fluorophenol
Concen: 0.11 ng
RT: 5.317 min Scan# 315
Delta R.T. -0.008 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	112	Resp:	280
Ion	Ratio	Lower	Upper
112	100		
64	58.6	43.6	65.4
63	30.7	25.0	37.6



#5
Phenol-d6
Concen: 0.12 ng
RT: 6.895 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:	99	Resp:	377
Ion	Ratio	Lower	Upper
99	100		
42	19.4	10.2	15.2#
71	44.8	32.4	48.6









#6
bis(2-Chloroethyl)ether
Concen: 0.12 ng
RT: 7.145 min Scan# 542
Delta R.T. -0.008 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampled :
SSTDICC0.1


Tgt Ion: 93 Resp: 317
Ion Ratio Lower Upper
93 100
63 57.1 47.5 71.3
95 36.6 28.5 42.7



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion: 136 Resp: 3338
Ion Ratio Lower Upper
136 100
137 13.1 13.2 19.8#
54 6.9 9.0 13.4#
68 6.9 8.8 13.2#









#8
Nitrobenzene-d5
Concen: 0.12 ng
RT: 8.870 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 82 Resp: 335
Ion Ratio Lower Upper
82 100
128 49.8 34.2 51.2
54 50.7 38.4 57.6



#9
Naphthalene
Concen: 0.11 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.013 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion: 128 Resp: 976
Ion Ratio Lower Upper
128 100
129 19.4 13.0 19.4
127 22.2 15.0 22.6









#10
Hexachlorobutadiene
Concen: 0.12 ng
RT: 10.835 min Scan# 886
Delta R.T. -0.013 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:225 Resp: 264
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.1 50.8 76.2



#11
2-Methylnaphthalene-d10
Concen: 0.09 ng
RT: 12.098 min Scan# 1072
Delta R.T. -0.004 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:152 Resp: 695
Ion Ratio Lower Upper
152 100
151 20.1 16.8 25.2









#12
2-Methylnaphthalene
Concen: 0.11 ng
RT: 12.173 min Scan# 1089
Delta R.T. 0.000 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	142	Resp:	725
Ion	Ratio	Lower	Upper
142	100		
141	89.9	72.9	109.3
115	52.3	41.8	62.6



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:	164	Resp:	2626
Ion	Ratio	Lower	Upper
164	100		
162	104.6	83.2	124.8
160	48.3	41.8	62.6









#14
2,4,6-Tribromophenol
Concen: 0.13 ng
RT: 15.856 min Scan# 1428
Delta R.T. -0.013 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:330 Resp: 154
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 31.8 21.8 32.6



#15
2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.976 min Scan# 1201
Delta R.T. -0.013 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:172 Resp: 989
Ion Ratio Lower Upper
172 100
171 42.9 30.9 46.3
170 29.1 21.4 32.0









#16
Acenaphthylene
Concen: 0.11 ng
RT: 14.080 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	152	Resp:	1024
Ion	Ratio	Lower	Upper
152	100		
151	20.4	17.1	25.7
153	14.4	9.9	14.9



#17
Acenaphthene
Concen: 0.11 ng
RT: 14.422 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:	154	Resp:	785
Ion	Ratio	Lower	Upper
154	100		
153	113.0	91.8	137.8
152	58.0	46.0	69.0









#18
Fluorene
Concen: 0.12 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:166 Resp: 1116
Ion Ratio Lower Upper
166 100
165 97.0 79.9 119.9
167 17.5 10.5 15.7#



#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1531
Delta R.T. 0.001 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:188 Resp: 6860
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.4 8.2 12.4









#20
4,6-Dinitro-2-methylphenol
Concen: 0.09 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.013 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:198 Resp: 73
Ion Ratio Lower Upper
198 100
51 141.2 96.4 144.6
105 98.8 72.3 108.5



#21
4-Bromophenyl-phenylether
Concen: 0.12 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.001 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:248 Resp: 489
Ion Ratio Lower Upper
248 100
250 98.6 77.5 116.3
141 58.7 50.2 75.4









#22
Hexachlorobenzene
Concen: 0.12 ng
RT: 16.423 min Scan# 1474
Delta R.T. 0.001 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampled :
SSTDIC0.1


Tgt Ion:284 Resp: 518
Ion Ratio Lower Upper
284 100
142 29.3 20.3 30.5
249 36.9 26.4 39.6



#23
Atrazine
Concen: 0.12 ng
RT: 16.581 min Scan# 1487
Delta R.T. 0.001 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:200 Resp: 318
Ion Ratio Lower Upper
200 100
173 41.4 29.1 43.7
215 62.8 45.1 67.7









#25
Phenanthrene
Concen: 0.12 ng
RT: 17.153 min Scan# 1534
Delta R.T. 0.001 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 2273
Ion Ratio Lower Upper
178 100
176 20.0 16.0 24.0
179 17.0 12.6 19.0



#26
Anthracene
Concen: 0.11 ng
RT: 17.250 min Scan# 1542
Delta R.T. 0.001 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:178 Resp: 1759
Ion Ratio Lower Upper
178 100
176 18.5 14.2 21.2
179 15.5 13.7 20.5









#27
Fluoranthene-d10
Concen: 0.09 ng
RT: 19.154 min Scan# 1834
Delta R.T. -0.002 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 2847
Ion Ratio Lower Upper
212 100
106 8.7 7.4 11.2
104 5.8 4.1 6.1



#28
Fluoranthene
Concen: 0.12 ng
RT: 19.183 min Scan# 1840
Delta R.T. -0.002 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:202 Resp: 2877
Ion Ratio Lower Upper
202 100
101 9.1 7.0 10.6
203 16.9 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.324 min Scan# 2130
Delta R.T. 0.003 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:240	Resp:	9743
Ion Ratio	Lower	Upper
240	100	
120	5.7	5.2 7.8
236	27.2	21.6 32.4



#30
Pyrene
Concen: 0.12 ng
RT: 19.551 min Scan# 1914
Delta R.T. -0.002 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:202	Resp:	3138
Ion Ratio	Lower	Upper
202	100	
200	20.9	16.6 24.8
203	17.0	13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.10 ng
 RT: 19.759 min Scan# 1956
 Delta R.T. -0.002 min
 Lab File: BN014768.D
 Acq: 26 May 2021 16:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 2373
 Ion Ratio Lower Upper
 244 100
 212 9.6 6.8 10.2
 122 10.6 6.6 9.8#



#32
 Benzo(a)anthracene
 Concen: 0.11 ng
 RT: 21.303 min Scan# 2128
 Delta R.T. 0.003 min
 Lab File: BN014768.D
 Acq: 26 May 2021 16:56

Tgt Ion:228 Resp: 3060
 Ion Ratio Lower Upper
 228 100
 226 30.3 22.9 34.3
 229 22.2 16.2 24.2









#33
Chrysene
Concen: 0.11 ng
RT: 21.356 min Scan# 2133
Delta R.T. 0.003 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	228	Resp:	3234
Ion Ratio	Lower	Upper	
228	100		
226	30.8	24.3	36.5
229	21.6	16.4	24.6



#35
Perylene-d12
Concen: 0.40 ng
RT: 23.589 min Scan# 2577
Delta R.T. 0.003 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:	264	Resp:	9754
Ion Ratio	Lower	Upper	
264	100		
260	26.6	21.4	32.2
265	37.6	40.9	61.3









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.09 ng
RT: 25.878 min Scan# 3246
Delta R.T. 0.010 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:276 Resp: 3385
Ion Ratio Lower Upper
276 100
138 11.1 10.2 15.4
277 25.3 20.4 30.6



#37
Benzo(b)fluoranthene
Concen: 0.10 ng
RT: 22.907 min Scan# 2378
Delta R.T. 0.006 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:252 Resp: 3421
Ion Ratio Lower Upper
252 100
253 42.4 21.5 32.3#
125 13.0 6.6 10.0#









#38
Benzo(k)fluoranthene
Concen: 0.10 ng
RT: 22.952 min Scan# 2391
Delta R.T. 0.003 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 3580
Ion Ratio Lower Upper
252 100
253 42.6 22.7 34.1#
125 10.4 6.4 9.6#



#39
Benzo(a)pyrene
Concen: 0.11 ng
RT: 23.493 min Scan# 2549
Delta R.T. 0.006 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:252 Resp: 3003
Ion Ratio Lower Upper
252 100
253 53.0 25.1 37.7#
125 21.0 9.4 14.0#









#40
Dibenzo(a,h)anthracene
Concen: 0.08 ng
RT: 25.899 min Scan# 3252
Delta R.T. 0.013 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:278 Resp: 2715
Ion Ratio Lower Upper
278 100
139 14.7 8.6 13.0#
279 38.3 22.8 34.2#



#41
Benzo(g,h,i)perylene
Concen: 0.10 ng
RT: 26.577 min Scan# 3450
Delta R.T. 0.010 min
Lab File: BN014768.D
Acq: 26 May 2021 16:56

Tgt Ion:276 Resp: 3198
Ion Ratio Lower Upper
276 100
277 29.0 20.9 31.3
138 17.2 10.3 15.5#

