

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014771.D
 Acq On : 26 May 2021 18:44
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: May 26 19:16:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 26 18:47:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	617	0.40	ng	0.00
7) Naphthalene-d8	10.505	136	2326	0.40	ng	# 0.00
13) Acenaphthene-d10	14.359	164	1822	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	4614	0.40	ng	0.00
29) Chrysene-d12	21.314	240	6937	0.40	ng	0.00
35) Perylene-d12	23.582	264	6868	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.317	112	1286	0.79	ng	0.00
5) Phenol-d6	6.895	99	1691	0.78	ng	0.00
8) Nitrobenzene-d5	8.870	82	1561	0.78	ng	0.00
11) 2-Methylnaphthalene-d10	12.098	152	3680	0.65	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	758	0.89	ng	-0.01
15) 2-Fluorobiphenyl	12.976	172	5399	0.81	ng	-0.01
27) Fluoranthene-d10	19.154	212	14039	0.68	ng	0.00
31) Terphenyl-d14	19.759	244	13102	0.80	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.271	88	576	0.70	ng	# 92
3) n-Nitrosodimethylamine	3.666	42	190	0.61	ng	# 1
6) bis(2-Chloroethyl)ether	7.153	93	1532	0.85	ng	99
9) Naphthalene	10.543	128	5326	0.87	ng	# 93
10) Hexachlorobutadiene	10.847	225	1384	0.94	ng	# 100
12) 2-Methylnaphthalene	12.169	142	3844	0.86	ng	99
16) Acenaphthylene	14.080	152	5602	0.84	ng	98
17) Acenaphthene	14.422	154	4182	0.84	ng	99
18) Fluorene	15.412	166	5927	0.89	ng	100
20) 4,6-Dinitro-2-methylph...	15.501	198	383	0.73	ng	# 47
21) 4-Bromophenyl-phenylether	16.313	248	2538	0.90	ng	91
22) Hexachlorobenzene	16.423	284	2817	0.94	ng	98
23) Atrazine	16.581	200	1510	0.84	ng	# 93
24) Pentachlorophenol	16.776	266	565	0.57	ng	# 85
25) Phenanthrene	17.153	178	10906	0.86	ng	100
26) Anthracene	17.250	178	8986	0.85	ng	97
28) Fluoranthene	19.183	202	15481	0.92	ng	99
30) Pyrene	19.546	202	15417	0.81	ng	100
32) Benzo(a)anthracene	21.303	228	16210	0.80	ng	98
33) Chrysene	21.346	228	19198	0.91	ng	99
34) Bis(2-ethylhexyl)phtha...	21.239	149	4339	0.77	ng	99
36) Indeno(1,2,3-cd)pyrene	25.865	276	23174	0.84	ng	99
37) Benzo(b)fluoranthene	22.897	252	20349	0.87	ng	95
38) Benzo(k)fluoranthene	22.945	252	21520	0.89	ng	92
39) Benzo(a)pyrene	23.482	252	18432	0.93	ng	93
40) Dibenzo(a,h)anthracene	25.878	278	19352	0.82	ng	94
41) Benzo(g,h,i)perylene	26.560	276	20359	0.90	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
Data File : BN014771.D
Acq On : 26 May 2021 18:44
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: May 26 19:16:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 26 18:47:42 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.716 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 617
Ion Ratio Lower Upper
152 100
150 154.2 123.4 185.2
115 85.5 73.2 109.8



#2
1,4-Dioxane
Concen: 0.70 ng
RT: 3.271 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion: 88 Resp: 576
Ion Ratio Lower Upper
88 100
58 60.6 44.7 67.1
43 20.0 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 0.61 ng
RT: 3.666 min Scan# 110
Delta R.T. 0.024 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion: 42 Resp: 190
Ion Ratio Lower Upper
42 100
74 0.0 0.0 0.0
44 0.0 80.0 120.0#



#4
2-Fluorophenol
Concen: 0.79 ng
RT: 5.317 min Scan# 315
Delta R.T. -0.008 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion: 112 Resp: 1286
Ion Ratio Lower Upper
112 100
64 54.4 43.6 65.4
63 29.1 25.0 37.6









#5
Phenol-d6
Concen: 0.78 ng
RT: 6.895 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion: 99 Resp: 1691




Ion	Ratio	Lower	Upper
99	100		
42	11.9	10.2	15.2
71	45.5	32.4	48.6




#6
bis(2-Chloroethyl)ether
Concen: 0.85 ng
RT: 7.153 min Scan# 543
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion: 93 Resp: 1532

Ion	Ratio	Lower	Upper
93	100		
63	59.8	47.5	71.3
95	34.6	28.5	42.7









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	2326
Ion	Ratio	Lower	Upper
136	100		
137	14.8	13.2	19.8
54	8.1	9.0	13.4#
68	8.2	8.8	13.2#



#8
Nitrobenzene-d5
Concen: 0.78 ng
RT: 8.870 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:	82	Resp:	1561
Ion	Ratio	Lower	Upper
82	100		
128	38.4	34.2	51.2
54	37.4	38.4	57.6#









#9
Naphthalene
Concen: 0.87 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.013 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	128	Resp:	5326
Ion Ratio	Lower	Upper	
128	100		
129	12.8	13.0	19.4#
127	16.0	15.0	22.6



#10
Hexachlorobutadiene
Concen: 0.94 ng
RT: 10.847 min Scan# 887
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:	225	Resp:	1384
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.7	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.65 ng
RT: 12.098 min Scan# 1072
Delta R.T. -0.004 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 3680
Ion Ratio Lower Upper
152 100
151 21.2 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.86 ng
RT: 12.169 min Scan# 1088
Delta R.T. -0.004 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:142 Resp: 3844
Ion Ratio Lower Upper
142 100
141 90.4 72.9 109.3
115 50.8 41.8 62.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	164	Resp:	1822
Ion	Ratio	Lower	Upper
164	100		
162	103.5	83.2	124.8
160	49.5	41.8	62.6



#14
2,4,6-Tribromophenol
Concen: 0.89 ng
RT: 15.856 min Scan# 1428
Delta R.T. -0.013 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:	330	Resp:	758
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	27.6	21.8	32.6









#15
2-Fluorobiphenyl
Concen: 0.81 ng
RT: 12.976 min Scan# 1201
Delta R.T. -0.013 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	172	Resp:	5399
Ion Ratio	Lower	Upper	
172	100		
171	38.3	30.9	46.3
170	25.9	21.4	32.0



#16
Acenaphthylene
Concen: 0.84 ng
RT: 14.080 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:	152	Resp:	5602
Ion Ratio	Lower	Upper	
152	100		
151	20.6	17.1	25.7
153	13.3	9.9	14.9









#17
Acenaphthene
Concen: 0.84 ng
RT: 14.422 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 4182
Ion Ratio Lower Upper
154 100
153 114.2 91.8 137.8
152 56.8 46.0 69.0



#18
Fluorene
Concen: 0.89 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:166 Resp: 5927
Ion Ratio Lower Upper
166 100
165 99.6 79.9 119.9
167 13.8 10.5 15.7









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1531
Delta R.T. 0.001 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:188 Resp: 4614
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.9 8.2 12.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.73 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.013 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:198 Resp: 383
Ion Ratio Lower Upper
198 100
51 54.0 96.4 144.6#
105 49.8 72.3 108.5#









#21
4-Bromophenyl-phenylether
Concen: 0.90 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.001 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampled :
SSTDICC0.8


Tgt Ion:	248	Resp:	2538
Ion	Ratio	Lower	Upper
248	100		
250	101.2	77.5	116.3
141	50.8	50.2	75.4



#22
Hexachlorobenzene
Concen: 0.94 ng
RT: 16.423 min Scan# 1474
Delta R.T. 0.001 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:	284	Resp:	2817
Ion	Ratio	Lower	Upper
284	100		
142	27.4	20.3	30.5
249	32.7	26.4	39.6









#23
Atrazine
Concen: 0.84 ng
RT: 16.581 min Scan# 1487
Delta R.T. 0.001 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 1510
Ion Ratio Lower Upper
200 100
173 28.4 29.1 43.7#
215 54.2 45.1 67.7



#24
Pentachlorophenol
Concen: 0.57 ng
RT: 16.776 min Scan# 1503
Delta R.T. 0.001 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:266 Resp: 565
Ion Ratio Lower Upper
266 100
264 66.0 44.2 66.2
268 66.4 44.2 66.2#









#25
Phenanthrene
Concen: 0.86 ng
RT: 17.153 min Scan# 1534
Delta R.T. 0.001 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 10906
Ion Ratio Lower Upper
178 100
176 19.7 16.0 24.0
179 15.8 12.6 19.0



#26
Anthracene
Concen: 0.85 ng
RT: 17.250 min Scan# 1542
Delta R.T. 0.001 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:178 Resp: 8986
Ion Ratio Lower Upper
178 100
176 18.6 14.2 21.2
179 15.6 13.7 20.5









#27
Fluoranthene-d10
Concen: 0.68 ng
RT: 19.154 min Scan# 1834
Delta R.T. -0.002 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 14039
Ion Ratio Lower Upper
212 100
106 9.4 7.4 11.2
104 5.6 4.1 6.1



#28
Fluoranthene
Concen: 0.92 ng
RT: 19.183 min Scan# 1840
Delta R.T. -0.002 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:202 Resp: 15481
Ion Ratio Lower Upper
202 100
101 8.6 7.0 10.6
203 17.1 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.314 min Scan# 2129
Delta R.T. -0.008 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:240 Resp: 6937




Ion	Ratio	Lower	Upper
240	100		
120	7.4	5.2	7.8
236	28.1	21.6	32.4




#30
Pyrene
Concen: 0.81 ng
RT: 19.546 min Scan# 1913
Delta R.T. -0.007 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:202 Resp: 15417

Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.6	24.8
203	17.3	13.9	20.9









#31
Terphenyl-d14
Concen: 0.80 ng
RT: 19.759 min Scan# 1956
Delta R.T. -0.002 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:244 Resp: 13102
Ion Ratio Lower Upper
244 100
212 7.6 6.8 10.2
122 7.4 6.6 9.8



#32
Benzo(a)anthracene
Concen: 0.80 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.003 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:228 Resp: 16210
Ion Ratio Lower Upper
228 100
226 26.8 22.9 34.3
229 20.5 16.2 24.2









#33
Chrysene
Concen: 0.91 ng
RT: 21.346 min Scan# 2132
Delta R.T. -0.008 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	228	Resp:	19198
Ion	Ratio	Lower	Upper
228	100		
226	30.9	24.3	36.5
229	19.9	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.77 ng
RT: 21.239 min Scan# 2122
Delta R.T. 0.003 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:	149	Resp:	4339
Ion	Ratio	Lower	Upper
149	100		
167	36.2	28.3	42.5
279	7.4	5.5	8.3









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.582 min Scan# 2575
Delta R.T. -0.004 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:264 Resp: 6868




Ion	Ratio	Lower	Upper
264	100		
260	25.7	21.4	32.2
265	44.6	40.9	61.3




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.84 ng
RT: 25.865 min Scan# 3242
Delta R.T. -0.004 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:276 Resp: 23174

Ion	Ratio	Lower	Upper
276	100		
138	11.9	10.2	15.4
277	25.7	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 0.87 ng
RT: 22.897 min Scan# 2375
Delta R.T. -0.004 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 20349
Ion Ratio Lower Upper
252 100
253 23.7 21.5 32.3
125 7.7 6.6 10.0



#38
Benzo(k)fluoranthene
Concen: 0.89 ng
RT: 22.945 min Scan# 2389
Delta R.T. -0.004 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:252 Resp: 21520
Ion Ratio Lower Upper
252 100
253 23.8 22.7 34.1
125 6.4 6.4 9.6









#39
Benzo(a)pyrene
Concen: 0.93 ng
RT: 23.482 min Scan# 2546
Delta R.T. -0.004 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	252	Resp:	18432
Ion Ratio	Lower	Upper	
252	100		
253	26.5	25.1	37.7
125	13.1	9.4	14.0



#40
Dibenzo(a,h)anthracene
Concen: 0.82 ng
RT: 25.878 min Scan# 3246
Delta R.T. -0.007 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44


Tgt Ion:	278	Resp:	19352
Ion Ratio	Lower	Upper	
278	100		
139	9.5	8.6	13.0
279	25.1	22.8	34.2





#41
Benzo(g,h,i)perylene
Concen: 0.90 ng
RT: 26.560 min Scan# 3445
Delta R.T. -0.007 min
Lab File: BN014771.D
Acq: 26 May 2021 18:44

Tgt Ion:	276	Resp:	20359
Ion	Ratio	Lower	Upper
276	100		
277	25.1	20.9	31.3
138	11.9	10.3	15.5



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8